

Air



Continuous Rural Ozone Data Collection in the Northeast United States

Continuous Rural Ozone Data Collection in the Northeast United States

by

Environmental Research and Technology, Inc.
696 Virginia Road
Concord, Massachusetts 01742

Contract No. 68-02-3522

EPA Project Officer: Edward J. Hanks, Jr.

U.S. ENVIRONMENTAL PROTECTION AGENCY
Office of Air, Noise, and Radiation
Office of Air Quality Planning and Standards
Research Triangle Park, North Carolina 27711

November 1980

TECHNICAL REPORT DATA

(Please read Instructions on the reverse before completing)

1. REPORT NO. EPA-450/4-80-035		2.		3. RECIPIENT'S ACCESSION NO.	
4. TITLE AND SUBTITLE Continuous Rural Ozone Data Collection in the Northeast United States				5. REPORT DATE November 1980	
				6. PERFORMING ORGANIZATION CODE	
7. AUTHOR(S)				8. PERFORMING ORGANIZATION REPORT NO.	
9. PERFORMING ORGANIZATION NAME AND ADDRESS Environmental Research & Technology, Inc. 696 Virginia Road Concord, Massachusetts 01742				10. PROGRAM ELEMENT NO.	
				11. CONTRACT/GRANT NO. 68-02-3522	
12. SPONSORING AGENCY NAME AND ADDRESS United States Environmental Protection Agency Office of Air Quality Planning and Standards Research Triangle Park, North Carolina 27711				13. TYPE OF REPORT AND PERIOD COVERED	
				14. SPONSORING AGENCY CODE	
15. SUPPLEMENTARY NOTES EPA Project Officer: Edward J. Hanks, Jr. Contractor Project Manager: Richard D. Andrews					
16. ABSTRACT This report presents rural ozone data collected in the Northeast United States during the summer period of July 1 to September 12, 1980. The monitoring was performed at four former Electric Power Research Institute-Sulfate Regional Experiment (EPRI-SURE) monitoring stations at Montague, MA; Scranton, PA; Indian River, DE; and Lewisburg, WV.					
17. KEY WORDS AND DOCUMENT ANALYSIS					
a. DESCRIPTORS		b. IDENTIFIERS/OPEN ENDED TERMS		c. COSATI Field/Group	
Air Quality Monitoring Ozone Rural Ozone EPRI-SURE Northeast United States NECRMP Data Collection					
18. DISTRIBUTION STATEMENT		19. SECURITY CLASS (This Report)		21. NO. OF PAGES 41	
		20. SECURITY CLASS (This page)		22. PRICE	

This document is issued by the Environmental Protection Agency to report technical data of interest to a limited number of readers. Copies are available free of charge to Federal employees, current EPA contractors and grantees, and nonprofit organizations - in limited quantities - from the Library Services Office (MD-35), U.S. Environmental Protection Agency, Research Triangle Park, North Carolina 27711; or, for a fee, from the National Technical Information Service, 5285 Port Royal Road, Springfield, Virginia 22161.

This report was furnished to the Environmental Protection Agency by Environmental Research and Technology, Inc., 696 Virginia Road, Concord, Massachusetts 01742, in fulfillment of Contract No. 68-02-3522. The contents of this report are reproduced herein as received from Environmental Research and Technology, Inc. The opinions, findings and conclusions expressed are those of the author and not necessarily those of the Environmental Protection Agency.

Publication No. EPA-450/4-80-035

TABLE OF CONTENTS

	Page
DISCLAIMER	ii
ABSTRACT	1
INTRODUCTION	2
CONFIGURATION	3
OPERATIONS	11
DATA VALIDATION	15
APPENDIX A - DATA REPORT	21
APPENDIX B - CALIBRATION SUMMARIES	34
APPENDIX C - ERT STANDARD OPERATING PROCEDURES	40

LIST OF FIGURES

Figure	Page
Figure 1 Sample and Data Flow for Continuous Monitoring	6
Figure 2 System for Generating Test Atmospheres	8
Figure 3 Level I Data Validation Flow	16
Figure 4 Level II Data Validation Flow	18

LIST OF TABLES

Table		Page
1	Monitoring Station Locations and Identification	4
2	External Sampling Location Criteria for Sites	5

ABSTRACT

This report presents rural ozone data collected in the Northeast United States during the summer period of July 1 to September 12, 1980. The monitoring was performed at four former Electric Power Research Institute-Sulfate Regional Experiment (EPRI-SURE) monitoring stations at Montague, MA; Scranton, PA; Indian River, DE; and Lewisburg, WV.

INTRODUCTION

The network of ozone monitoring stations was defined by EPA to be the four former EPRI-SURE monitoring stations located near Lewisburg, West Virginia; Indian River, Delaware; Scranton, Pennsylvania; and Montague, Massachusetts. Analyzers were supplied as Government Furnished Equipment. Specifically, Monitor Labs Model ML 8410E ozone analyzers and strip chart recorders were used in the program.

The monitoring program was conducted during the period of July 1 through September 12, 1980. Scheduled checks and calibrations were performed at times other than during the daily time period of 9:00 AM and 6:00 PM LDT, thus allowing a continuous record of daytime ozone concentrations.

CONFIGURATION

Four existing EPRI-SURE monitoring station locations were selected by the EPA for the conduct of this program. Station locations and their identifications are listed in Table 1.

When the stations were originally located in 1977, it was desirable that the locations meet EPRI-SURE program specified criteria. The external siting criteria is presented in Table 2. It was impossible for all sites to comply exactly with all of the external criteria. Confirmation of or deviations to the external criteria follow:

- Montague - Met all external criteria except for the proximity of tall trees.
- Scranton - Met all external criteria except for its close proximity to a dirt road and corn field.
- Indian River - Met all external criteria and the Prevention of Significant Deterioration (PSD) Guideline siting criteria.
- Lewisburg - Met all external criteria and the PSD Guideline siting criteria.

Environmentally controlled shelters housed the monitoring equipment. Each shelter was equipped with a Monitor Labs Model 8410E ozone analyzer, Esterline Angus MS 412C strip chart recorders, signal distribution system, and an air inlet manifold system. An equipment block diagram is presented in Figure 1.

Ambient air entered an inverted sampling probe located approximately one meter above the shelter roof (between 4 and 7 meters above grade elevation) and was moved into the sampling manifold at approximately 80 l/min using a continuously operating pump at the opposite end of the manifold. Thus, a constant supply of ambient air was available for sampling.

The Monitor Lab 8410E chemiluminescent O_3 analyzer operating on the 0.5 ppm scale was used to monitor O_3 . This instrument was chosen because of its availability on loan from EPA. In this system the fixed air flow is mixed with a fixed ethylene flow. Any O_3

TABLE 1
MONITORING STATION LOCATIONS AND IDENTIFICATION

<u>Site Name/No.</u>	<u>SURE ID No.</u>		<u>SAROAD ID No.</u>			<u>UTM Coordinates</u>	<u>UTM Zone</u>
Montague - U1	1	22	1392	001	A05	00702880 E 4715550 N	18
Scranton - U2	2	39	8040	100	A05	00410640 E 4604800 N	18
Indian River - U3	3	08	0240	001	A05	00476160 E 4270480 N	18
Lewisburg - U4	9	50	0580	002	A05	00558720 E 4180800 N	17



TABLE 2
EXTERNAL SAMPLING LOCATION CRITERIA FOR SITES
EPRI-SURE PROGRAM

These criteria involve local environment requirements of the sampling sites.

- station location generally to the west of significant point sources or urban areas. This is upwind under prevailing wind conditions;
- in the case of river-valley terrain, station location above and away from valley flow regimes;
- minimum flow interference by tall trees and obstacles. A horizontal clearance of ten times the obstacle height is preferred when the available alternatives permit this. However, an angular clearance of 30° or less from the horizontal was acceptable;
- avoidance of dirt roads and plowed fields in the vicinity of the station. A minimum acceptable distance of 30 meters was sought, but notable exceptions could not be obviated when placing stations into rural areas; and
- a distance of 16 km from any significant sources of sulfur such as polluted waterways and ponds, and waste-water treatment ponds.




Figure 1. Sample and Data Flow for Continuous Monitoring

present reacts with the ethylene to produce a blue light, the intensity of which is proportional to the amount of O_3 present (measured with a photomultiplier). No significant interferences are present in normal, ambient air. The minimum detectable concentration of this instrument is 2 ppb according to the manufacturer, and this was attained under field conditions.

The complementary component to the continuous gas measurement process was the in-station test atmosphere generation system. This system produced standardized concentrations of O_3 against which to compare instrument responses; it also provided zero air against which instrument baselines were established.

Figure 2 shows the test atmosphere generation system in block diagram form. The principle of this system is to dilute a standard gas concentration with known volumes of air from which that gas has been scrubbed. This dilution process allows a number of standard concentrations to be generated so that instrument response in a number of concentration ranges could be tested.

Zero air was produced by feeding ambient air with an oilless compressor through a water trap and heatless dryer which consisted of two tightly-packed columns each containing molecular sieve and activated charcoal. These columns were switched with a control valve every 30 seconds to allow recovery and to prevent saturation of the column. Nearly all of the NO_2 , and O_3 were scrubbed by the charcoal. Following the dryer, O_3 was introduced into the airstream to convert any remaining NO to NO_2 which was subsequently scrubbed by the next activated charcoal column along with the excess O_3 . A filter following the scrubber removed any particulate matter entrained in the airstream.

The zero airstream was divided into three lines to serve as zero or dilution air. The dilution air flow was varied manually with a Brooks mass flow controller. Changes in this flow rate over its 2 to 15 l/min range allowed different gas concentrations to be generated. For this program, the uses for the zero air supply were as follows:




Figure 2. The System for Generating Test Atmospheres was used for Continual Tracking of Gas Monitoring Performance

- To supply the ozonator, 50 cc/min of the zero air was passed over the ultraviolet lamp. The amount of O_3 produced was controlled by positioning an adjustable cylindrical shield at several fixed lengths of lamp exposure to the passing air, or by adjusting the zero air flow rate.
- To dilute or to mix with established O_3 concentrations, a standard concentration of NO in N_2 from an NBS traceable standardized compressed cylinder. The NO was introduced into the system at 40 cc/min.

To produce standard concentrations, the O_3 sweep gases and the NO standard gas were routed from their normal paths through the exhaust manifold into the 150cc mixing chamber where they mixed with the zero dilution air under turbulent air conditions.

NO reacted with O_3 in this chamber to create NO_2 . With NO in excess, the difference between the readings of the NO and NO_x channels of a calibrated instrument is equal to the O_3 concentration at the given rate of dilution for a setting of the ozonator UV lamp. Thus the NO source, together with the ozonator and NO source flow controllers, served as the in-station standard for O_3 .

The gas mixture was presented to the monitors through a calibration manifold, distinct from the sampling manifold, which was accessible by the monitors when a set of solenoid valves switched the analyzer input to it.

The test-atmospheres could be presented to the monitoring instruments by either automatically or manually switching the valves. The automatic feature was used to check baseline and span values for each instrument daily. These checks took place in the early morning hours of each day.

The manual switching was used to check the automatic spans when it showed instrument performance tracking outside of specifications and on a weekly basis to present a range of test gas concentrations to the instruments.

All monitoring equipment was field calibrated prior to the start of data collection. The ML 8410E analyzer was calibrated with a Dasibi 1003 (Federal Register, Vol. 44, No. 28, Thursday, February 8, 1979) ultraviolet absorption ozone monitor with its internal calibration source. The Dasibi was also used as a transfer standard to verify the in-station calibration system. A multipoint calibration check of the ML 8410E was then performed using the in-station calibration system.

OPERATIONS

Station operations consisted of keeping all instruments functioning in such a way that precise and accurate measurements were made. ERT standard operating procedures were used to govern these activities. These procedures were supplemented by the manufacturers' manuals for each instrument and a series of Field Operations Bulletins which modified standard operating procedures in light of new experience. A listing of ERT proprietary standard operating procedures is presented in the Appendix.

A part-time technician was assigned to each monitoring station; these were veterans of the SURE. Each technician received initial and follow-up training in the standard procedures for each station and instrument. The objectives of the field technicians' routine visits were to:

- assess the status of all instruments;
- perform maintenance tasks as scheduled or required;
- document all significant information for later use in the analysis of data validity;
- report the station status and any problems to the network operations supervisor in Concord, MA; and
- deliver supplies to the station.

A routine visit checklist was filled out each time the sites were visited. Any negative response to an item was cause for corrective action. In most cases, this was a regular maintenance item, such as changing charcoal, which was handled by the field technician without further consultation. If the problem could not be corrected by the field technician, he would immediately call the network operations supervisor to advise him of it. The network operations supervisor would pinpoint the problem and devise the most efficient method of solving it. The options ranged from a temporary modification of the standard operating procedure to airshipping replacement equipment or sending a repair technician. All negative checklist items and the

corrective actions taken were noted in the station logbook. The logbook entry was made on a form in three copies one of which remained in the shelter for future reference; the other two were sent to Concord, MA on a weekly basis where they were reviewed by the network operations supervisor and the data validation technicians. This review resulted in further operations follow-up and was used for evaluating the data. Since the logbook and checklist entries were meant to be of use in the data validation process, the field technicians were instructed to provide as much useful information as possible which might explain anomalies in the data. The logbook entries also included:

- changes and calibrations of all equipment and serial numbers of each piece of equipment;
- a description of abnormal conditions in the vicinity of the site such as prevailing weather, or construction which might affect or bias the data;
- a classification of data as valid, invalid or suspect based on an evaluation of the monitoring instrument status;
- activities related to data collection such as the performance of routine and weekly checks, mailing checklists, logs and data.

Once per week, the routine visit was extended to perform a more lengthy evaluation of station performance. The key element of this check was in-station tracking of all measured quantities. This involved generating test atmospheres at two concentrations daily (zero and one concentration level), and at several concentrations weekly (zero and four concentration levels).

The daily checks were diagnostic; if the instrument response to the test atmosphere differed from the previous days response by more than $\pm 10\%$, the span was re-checked and if still out of tolerance, a full-scale instrument checkout was initiated.

Additional calibrations were made following repairs or replacements of components which failed. The performance of each of the in-station calibration devices was traceable to a primary standard.

Maintenance items such as changing exhaust filters, etc., were also attended to on a weekly basis. After these checks, the strip charts for the week were cut and mailed to the network operations supervisor in Concord, MA together with the routine visit checklists, weekly visit checklists, logbook entries and multipoint tracking results.

Aside from numerous power outages during the program, a few equipment problems caused losses of data. The following is a data capture (%) summary for the contract period.

Site	July	August	September
Montague - U1	98.52	98.92	99.31
Scranton - U2	97.31	95.70	98.26
Indian River - U3	81.59	98.66	98.61
Lewisburg - U4	98.66	99.33	98.26

Copies of the data reports are enclosed in the Appendix. A magnetic tape of all valid data was submitted to the EPA project officer as a part of this report.

Also enclosed in the appendix are summaries of the daily automatic span delta percents, and summaries of the multipoint calibration check delta percents for each station. The multipoint summary also presents the results of the reference Dasibi checks and the RTI performed audits.

Noted equipment problems during the program were as follows:

Montague - U1

Balston auto drain broke on 7/23 causing air leak. Suspect water leak caused erroneous span value on 7/30. Power failure interrupted span check on 8/11.

Scranton - U2

Multiple power failures disrupted the timer, resulting in missed span checks.

Indian River - U3

The ML8410E was down on 7/1. A broken wire on the fuse holder was replaced on 7/2. A power failure on 7/17 caused the ML 8410E to go full-scale. The chopper motor belt was replaced on 7/21, restoring the unit to service. On 8/20, the shelter air conditioner failed; the unit was repaired on 8/29. During a part of this period, the in-station calibration system was shut down due to the air conditioner problems. On 9/12 the primary chart recorder jammed. Data was recovered from the backup recorder.

Lewisburg - U4

Failure of the zero air system on 7/10 and the NO dilution air system on 7/22 resulted in losses of daily span values. Three span values were lost in August due to electrical storms and power failures.

On 9/3, the mass flow meter failed causing full-scale span checks from 9/3 to 9/5. A needle valve was installed in the dilution air line, restoring the calibration system to partial service.

DATA VALIDATION

The hourly averages determined from the continuous gaseous measurements at the sites were subjected routinely to two levels of validation. These levels have been defined as follows:

- Level I - Data which have been compared with station records but have been only partially edited; this level of data was used to support initial analyses and interpretations.
- Level II - Data which have been validated through instrument performance checks, calibration, digitization and machine reading.

The process for obtaining Level I validation is illustrated in Figure 3. The flow of data began with the instrumentation at each station providing the station technician with various information about instrument status, strip chart traces, zero and span values, and analyzer responses to various tests (multipoint tracking). With this information, the station technician classified data for specified periods into three classes.

- INVALID over periods when the monitor was not operating at all.
- SUSPECT when the analyzer was operating, but zero/span and multipoint tracking checks were out of specification, or when some monitor or system failure which did not stop data acquisition cast doubt on the data quality.
- VALID when all instruments were functioning properly, and zero/span and multipoint tracking criteria were met.

These designations were recorded in the logbook, and INVALID and SUSPECT periods were communicated to the network supervisor for immediate corrective action. The network supervisor also recorded high and low percent deviations of weekly multipoint tracking to observe trends which might indicate preventive maintenance. The strip




Figure 3. Level I Data Validation Flow for Measurements at Sites.

charts, logs and tracking data were then transferred to the data analyst who verified the field technicians' INVALID designations and confirmed that all such data over the specified period was actually deleted from the data base.

Strip chart traces were machine digitized. Periods of VALID data designed by the technician were checked to assure that the data had been entered into the data base.

SUSPECT data were underlined on the monthly data summaries (CAPTUR) for each site and parameter. The reasons for the SUSPECT designation were transferred from the station logsheets onto the CAPTUR together with daily span and weekly multipoint performance test data.

During this process, the designations of the station technician were double-checked to assure that all VALID and INVALID data had been properly classified. A group leader supervising a team of data analysts reviewed all monthly data summaries to assure that the above procedures had been carried out and returned them to the data analyst, who entered the completed data set into the computerized data base and archived the annotated CAPTUR. At this point, the data for the particular parameter, site and month under consideration was designated as being at Level I validation.

To achieve Level II validation, data from the station calibrations, daily and weekly performance tests, and performance evaluations were reviewed together with the tracking graphs, and, if necessary, the logs and strip charts. This part of the process is illustrated schematically in Figure 4. All decisions concerning Level II data validation were documented by the data validation coordinator, approved by the Technical Director, and implemented by the data analyst. Typical actions required in the course of this process included:

- examination of the tracking graphs. These were plots of the changes in percent deviation from designated values of the lowest and highest test atmosphere concentrations observed weekly;




Figure 4. Level II Data Validation Flow for Measurements at the Sites

- assignment of uncertainties to periods of erratic analyzer responses; and
- reclassification of SUSPECT into VALID or INVALID data.

During periods in which deviations of the O₃ analyzer responses were less than 25%, no changes were made in VALID classification or in the data. When the responses to the test atmospheres exceeded these deviations, three possibilities were considered:

- the test atmosphere was generated within tolerances and the analyzer response had drifted;
- the test atmosphere composition had changed, and the air analyzer was operating within tolerance; and
- both the test atmosphere concentrations and the analyzer performance had changed.

To determine which possibility applied, an evaluation was done of the station operating logs in which failures of specific components and the repair or replacement were recorded, and of the performance test and recalibration data provided by the visits to each station with the reference Dasibi.

If the Dasibi performance test showed the first possibility, that in-station test atmosphere was in tolerance, but the analyzer response was not, then the analyzer was recalibrated.

When malfunction of the test atmosphere generation was indicated, and the analyzer response remained correct, the SUSPECT data was reclassified as VALID. Tracking drifts caused by the test atmosphere generator were primarily due to changes in the gas concentrations introduced to the dilution air system, rather than due to inaccurate dilution air flowrates.

Verification of this with the Dasibi performance test allowed the standard concentrations to be redesignated and their reliability preserved by the quality control test function.

When the Dasibi performance tests showed both the analyzer and the in-station test atmosphere standard to be outside of specification, a data qualification with a higher uncertainty would have been required. This condition did not occur.


Level II of these continuous analyzer data validation activities included some redundancy of the procedures carried out at the previous level. This served incidentally as a continual verification of the data validation procedures.

APPENDIX A

U.S. EPA


STATION		OZONE										(PPM)										DATA FOR JULY				1980
												HOURS (LST)														
NO.	NAME	01	02	03	04	05	06	07	08	09	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	AVG
1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27
1	00	00	00	00	00	00	00	009	026	039	045	044	046	051	058	057	058	059	057	046	059	059	060	055	035	
2	005	053	050	050	052	049	046	039	041	042	043	040	053	059	072	068	065	063	057	037	017	012	006	001	045	
3	00	00	00	004	005	006	015	023	026	033	049	061	068	070	077	088	088	096	093	075	052	037	032	016	042	
4	000	00	00	00	001	00	023	050	046	047	048	048	048	049	054	056	047	045	034	019	006	002	00	00	026	
5	0001	00	0002	00	00	00	010	031	045	056	071	087	076	087	090	104	104	089	091	096	083	064	058	039	053	
6	0032	029	025	020	019	029	026	025	032	034	032	028	026	023	028	037	037	034	031	030	025	018	015	008	027	
7	0005	004	003	00	00	00	006	022	022	028	031	033	034	034	037	038	039	042	039	022	014	014	017	011	021	
8	0007	00	00	00	00	00	001	007	016	013	023	027	043	042	030	099	099	034	029	014	005	00	00	00	013	
9	00	00	00	00	00	00	00	004	019	037	043	042	043	039	036	040	041	031	013	006	003	001	00	00	017	
10	0001	00	00	00	00	00	006	019	035	040	042	042	048	055	059	063	073	083	099	035	019	040	057	057	034	
11	0001	041	041	040	036	036	040	046	053	060	068	075	071	079	083	056	044	028	042	053	046	046	029	021	049	
12	0018	015	009	004	014	015	020	023	024	025	026	027	029	029	026	027	028	028	026	016	010	010	007	003	019	
13	00	00	00	00	00	00	007	020	026	027	028	029	030	028	031	033	032	027	016	008	006	003	001	00	015	
14	00	00	00	00	00	00	002	015	025	032	035	036	034	035	036	038	042	044	027	017	015	007	002	003	017	
15	006	017	030	038	026	021	033	044	061	072	085	087	084	091	090	078	077	099	110	091	075	056	047	042	057	
16	000	009	034	031	031	026	029	032	048	069	082	085	074	070	077	078	076	072	043	058	047	038	035	029	052	
17	000	019	015	003	004	001	012	036	037	035	042	053	056	060	064	056	054	099	099	099	099	032	029	022	032	
18	000	009	008	002	00	001	016	023	025	027	026	027	021	023	030	026	025	025	024	013	003	00	00	00	010	
19	000	00	00	00	00	00	006	026	040	049	062	062	079	096	090	084	071	068	064	059	059	056	055	043	043	
20	000	060	055	045	039	033	035	037	043	063	071	090	091	085	081	060	063	057	054	043	037	025	018	010	052	
21	0007	001	001	002	00	00	001	002	013	040	056	071	073	067	070	070	066	056	055	050	051	036	035	039	036	
22	0009	010	007	007	011	013	013	022	022	028	036	051	051	050	052	053	060	056	054	050	048	039	024	029	034	
23	0002	013	016	018	022	022	020	018	019	022	032	038	041	043	047	044	039	043	099	027	017	014	015	010	026	
24	0004	00	00	00	00	001	012	021	023	027	029	026	028	028	028	032	033	032	023	019	011	003	00	00	016	
25	00	00	00	00	00	00	004	005	023	033	036	036	035	033	039	047	057	060	046	030	017	015	010	00	022	
26	00	00	00	00	00	00	006	019	032	050	064	068	070	070	067	064	065	061	043	025	018	014	009	008	032	
27	003	00	001	00	00	00	008	031	048	060	074	086	095	095	094	101	111	106	087	076	064	059	051	050	054	
28	0007	021	028	019	009	013	020	035	036	045	050	056	058	057	060	065	063	053	058	056	054	042	029	021	042	
29	0016	003	007	002	002	00	006	015	018	016	015	018	019	019	015	014	013	018	012	014	012	009	011	011	017	
30	0006	003	002	00	00	002	002	007	021	034	044	053	051	059	059	051	055	099	099	021	014	003	002	00	022	
31	00	00	001	00	00	00	00	013	033	037	037	036	039	041	047	050	051	051	035	010	006	003	002	002	021	
32	0013	011	011	009	009	009	014	023	031	039	046	050	052	053	056	056	056	052	047	037	030	025	021	017	032	
33	00	01	01	01	01	01	01	01	01	01	01	01	01	01	01	01	01	01	01	01	01	01	01	01	01	

TOTAL HOURS = 744 NUMBER OF GOOD HOURS = 733 NUMBER OF MISSING HOURS = 11 DATA CAPTURE = 98.52 (PERCENT)
 ABOVE THREE ROWS ARE TOTAL HOURLY AVERAGES, TOTAL OBSERVATIONS/HOUR AND DATA CAPTURE STATISTICS
 TOTAL AVERAGE = 0.032 MAXIMUM HOURLY VALUE = 0.111 STANDARD DEVIATION = 0.0259



SCRANTON					OZONE								(PPM)								DATA FOR JULY								1980				
					HOURS (LST)																												
HR-SEC	00	01	02	03	04	05	06	07	08	09	10	11	12	13	14	15	16	17	18	19	20	21	22	23			AVG						
HA-LEG	01	02	03	04	05	06	07	08	09	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24									
DAY																																	
1	.030	.032	.032	.029	.023	.013	.019	.021	.027	.034	.035	.040	.041	.043	.044	.040	.045	.050	.042	.048	.052	.049	.040	.028	.036			.036					
2	.012	.019	.016	.019	.009	.021	.027	.028	.030	.034	.031	.035	.037	.045	.056	.056	.049	.045	.035	.039	.033	.023	.018	.012	.030			.030					
3	.011	.014	.019	.019	.012	.023	.025	.031	.030	.029	.029	.036	.039	.038	.032	.032	.030	.026	.025	.023	.022	.013	.013	.011	.024			.024					
4	.012	.016	.019	.014	.015	.013	.015	.020	.038	.055	.059	.059	.059	.057	.056	.057	.057	.056	.062	.051	.043	.044	.029	.024	.039			.039					
5	.016	.016	.011	.008	.037	.007	.008	.009	.021	.041	.073	.074	.071	.066	.066	.070	.062	.047	.050	.050	.043	.039	.041	.041	.039			.039					
6	.037	.035	.041	.038	.032	.033	.033	.031	.032	.034	.035	.033	.030	.031	.037	.042	.041	.043	.046	.043	.041	.041	.035	.031	.036			.036					
7	.025	.023	.018	.014	.011	.009	.004	.012	.027	.033	.034	.035	.035	.037	.034	.034	.034	.099	.099	.022	.021	.026	.023	.017	.024			.024					
8	.013	.011	.009	.008	.014	.018	.026	.025	.023	.029	.038	.045	.044	.042	.044	.049	.033	.037	.045	.048	.055	.053	.056	.054	.034			.034					
9	.043	.040	.038	.039	.027	.018	.014	.017	.031	.038	.039	.041	.044	.046	.047	.048	.049	.043	.043	.039	.032	.023	.024	.025	.035			.035					
10	.020	.021	.013	.013	.012	.015	.016	.019	.028	.039	.052	.062	.065	.067	.067	.062	.055	.058	.058	.050	.046	.040	.043	.041	.040			.040					
11	.019	.029	.022	.016	.013	.013	.018	.022	.013	.015	.018	.035	.044	.052	.056	.059	.060	.052	.045	.042	.043	.038	.036	.035	.034			.034					
12	.031	.031	.031	.027	.027	.025	.024	.022	.024	.028	.035	.040	.043	.042	.042	.042	.044	.047	.044	.038	.038	.038	.030	.031	.034			.034					
13	.026	.023	.019	.009	.012	.012	.016	.026	.034	.037	.036	.039	.045	.046	.046	.044	.043	.041	.044	.099	.099	.024	.029	.029	.031			.031					
14	.021	.020	.016	.010	.008	.009	.011	.017	.022	.015	.034	.048	.049	.048	.053	.054	.050	.046	.046	.047	.051	.060	.051	.051	.034			.034					
15	.025	.018	.020	.021	.018	.014	.008	.016	.033	.043	.054	.053	.058	.071	.076	.076	.060	.088	.087	.089	.086	.083	.066	.069	.052			.052					
16	.024	.052	.046	.036	.025	.022	.030	.034	.030	.050	.049	.051	.055	.056	.055	.050	.046	.042	.039	.035	.034	.039	.042	.032	.042			.042					
17	.021	.017	.019	.037	.019	.016	.036	.038	.999	.033	.044	.048	.047	.044	.045	.048	.040	.039	.040	.038	.042	.046	.047	.039	.037			.037					
18	.036	.032	.023	.020	.024	.026	.022	.030	.032	.036	.999	.039	.042	.049	.045	.042	.045	.045	.045	.043	.036	.028	.030	.017	.034			.034					
19	.014	.015	.009	.002	.001	.003	.008	.001	.005	.013	.031	.054	.055	.050	.057	.058	.059	.066	.078	.082	.061	.077	.082	.062	.041			.041					
20	.070	.063	.031	.045	.030	.025	.012	.018	.021	.043	.050	.050	.057	.065	.076	.090	.106	.110	.097	.097	.107	.096	.094	.065	.065			.065					
21	.060	.059	.035	.051	.049	.027	.012	.020	.047	.050	.052	.054	.056	.059	.057	.060	.066	.073	.999	.999	.999	.999	.999	.999	.031			.031					
22	.999	.999	.999	.999	.010	.011	.015	.011	.012	.015	.018	.029	.035	.040	.045	.045	.050	.035	.999	.999	.999	.999	.030	.027	.020	.999			.999				
23	.013	.015	.021	.025	.022	.021	.021	.023	.032	.037	.039	.038	.040	.044	.045	.043	.034	.032	.029	.030	.024	.017	.021	.023	.029			.029					
24	.019	.015	.013	.011	.009	.013	.020	.021	.026	.030	.031	.035	.044	.044	.043	.042	.046	.046	.048	.040	.034	.033	.031	.025	.030			.030					
25	.020	.013	.011	.012	.010	.005	.008	.012	.006	.020	.041	.053	.061	.060	.063	.064	.071	.077	.077	.068	.065	.067	.060	.051	.041			.041					
26	.044	.032	.025	.020	.013	.006	.003	.001	.006	.033	.069	.068	.090	.086	.078	.079	.062	.087	.101	.107	.100	.085	.060	.060	.057			.057					
27	.052	.032	.026	.020	.008	.007	.008	.016	.020	.042	.062	.069	.078	.082	.080	.078	.087	.081	.083	.086	.086	.067	.056	.047	.053			.053					
28	.047	.018	.045	.043	.030	.021	.033	.036	.040	.042	.040	.043	.060	.062	.061	.057	.054	.052	.041	.999	.033	.038	.036	.031	.043			.043					
29	.019	.026	.021	.017	.009	.008	.009	.016	.013	.017	.015	.025	.035	.029	.038	.041	.041	.039	.030	.034	.038	.026	.023	.021	.025			.025					
30	.030	.021	.046	.047	.038	.034	.024	.025	.026	.030	.046	.048	.049	.053	.061	.060	.051	.044	.044	.049	.045	.042	.027	.032	.040			.040					
31	.026	.017	.010	.010	.010	.007	.012	.012	.018	.015	.032	.047	.049	.048	.046	.040	.052	.048	.056	.051	.038	.029	.020	.010	.029			.029					
AVG	.031	.027	.025	.022	.018	.016	.017	.020	.025	.033	.041	.047	.050	.052	.053	.054	.054	.053	.053	.051	.049	.044	.040	.035	.038			.038					
DAYS	30	30	30	30	31	31	31	31	30	31	30	31	31	31	31	31	31	30	28	27	28	30	30	30	724								


TOTAL HOURS = 744 NUMBER OF GOOD HOURS = 724 NUMBER OF MISSING HOURS = 20 DATA CAPTURE = 97.31 (PERCENT)
 ABOVE THREE ROWS ARE TOTAL HOURLY AVERAGES, TOTAL OBSERVATIONS/HOUR AND DATA CAPTURE STATISTICS
 TOTAL AVERAGE = 0.038 MAXIMUM HOURLY VALUE = 0.110 STANDARD DEVIATION = 0.0204



U.S. EPA


INDIAN RIVER						OZONE						(PPM)						DATA FOR JULY								1980
						HOURS (LST)																				
HR-BEG	00	01	02	03	04	05	06	07	08	09	10	11	12	13	14	15	16	17	18	19	20	21	22	23		
HR-END	01	02	03	04	05	06	07	08	09	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24		
DAY																									AVG	
1	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
2	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.041	.050	.043	.037	.036	.035	.036	.037	.041	.999
3	.041	.043	.042	.036	.027	.025	.023	.027	.029	.034	.039	.048	.051	.050	.051	.044	.038	.041	.051	.047	.043	.041	.045	.039	.040	
4	.030	.032	.026	.020	.014	.009	.010	.017	.027	.047	.057	.074	.089	.094	.108	.112	.110	.106	.101	.100	.081	.069	.061	.045	.060	
5	.040	.039	.030	.013	.004	.005	.020	.045	.060	.079	.080	.063	.060	.049	.051	.050	.054	.056	.055	.044	.038	.036	.036	.031	.043	
6	.023	.025	.024	.019	.019	.017	.019	.033	.047	.043	.047	.044	.048	.044	.049	.051	.050	.051	.048	.043	.034	.034	.032	.022	.036	
7	.022	.024	.999	.999	.999	.999	.999	.999	.020	.030	.041	.044	.046	.042	.031	.029	.048	.049	.049	.045	.042	.035	.042	.042	.038	
8	.037	.029	.031	.033	.028	.033	.039	.035	.032	.020	.018	.025	.036	.039	.043	.048	.040	.037	.026	.019	.016	.018	.018	.025	.030	
9	.022	.020	.021	.016	.012	.012	.010	.012	.027	.040	.041	.057	.067	.079	.077	.078	.080	.073	.075	.067	.060	.052	.031	.044	.045	
10	.054	.060	.056	.040	.025	.016	.023	.021	.021	.023	.027	.037	.046	.056	.063	.059	.055	.054	.049	.052	.043	.040	.037	.038	.041	
11	.040	.034	.044	.055	.054	.053	.045	.049	.042	.049	.052	.073	.081	.069	.070	.070	.067	.071	.073	.105	.112	.092	.076	.072	.064	
12	.060	.058	.056	.052	.043	.037	.037	.034	.041	.056	.056	.059	.067	.072	.071	.086	.087	.087	.086	.079	.064	.059	.044	.035	.059	
13	.035	.042	.038	.031	.032	.029	.030	.034	.035	.040	.043	.042	.046	.063	.063	.067	.072	.080	.083	.072	.058	.045	.042	.038	.048	
14	.132	.026	.023	.015	.010	.006	.999	.999	.028	.049	.071	.075	.074	.081	.084	.086	.086	.080	.074	.068	.063	.055	.048	.044	.054	
15	.039	.039	.032	.022	.022	.020	.019	.027	.036	.047	.059	.071	.079	.079	.078	.078	.074	.075	.068	.058	.049	.041	.036	.036	.049	
16	.036	.036	.036	.031	.026	.024	.022	.025	.033	.044	.056	.067	.072	.069	.080	.085	.080	.077	.999	.999	.042	.052	.038	.034	.048	
17	.033	.034	.031	.027	.025	.030	.033	.033	.034	.039	.051	.060	.068	.065	.069	.081	.088	.075	.049	.999	.999	.999	.999	.999	.999	
18	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
19	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
20	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
21	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
22	.042	.041	.037	.037	.032	.024	.017	.013	.019	.029	.041	.045	.053	.048	.043	.057	.047	.037	.027	.016	.034	.030	.028	.024	.034	
23	.013	.016	.014	.019	.019	.018	.019	.021	.021	.020	.027	.026	.023	.025	.026	.028	.030	.029	.021	.023	.023	.022	.016	.010	.021	
24	.005	.006	.004	.005	.002	.005	.001	.003	.006	.010	.018	.033	.052	.054	.047	.050	.054	.047	.044	.038	.025	.028	.008	.001	.023	
25	.011	.013	.012	.011	.016	.007	.003	.002	.024	.046	.058	.058	.057	.062	.071	.079	.078	.071	.076	.077	.061	.056	.070	.069	.045	
26	.067	.068	.060	.057	.052	.038	.025	.026	.035	.045	.056	.066	.068	.070	.070	.070	.067	.065	.067	.065	.061	.056	.057	.059	.057	
27	.057	.034	.020	.006	.020	.017	.005	.020	.046	.053	.053	.052	.054	.053	.051	.047	.041	.042	.042	.044	.039	.038	.032	.027	.037	
28	.027	.030	.029	.030	.031	.033	.034	.999	.999	.040	.037	.034	.037	.035	.036	.034	.030	.029	.029	.027	.025	.027	.029	.028	.031	
29	.027	.025	.027	.020	.021	.019	.021	.019	.020	.017	.016	.027	.031	.034	.034	.041	.053	.070	.072	.065	.049	.052	.048	.044	.035	
30	.035	.020	.023	.027	.023	.019	.017	.018	.030	.041	.054	.064	.062	.065	.070	.074	.074	.100	.095	.075	.056	.043	.043	.049	.049	
31	.037	.041	.042	.036	.029	.028	.013	.014	.033	.039	.039	.068	.055	.052	.071	.084	.086	.089	.089	.081	.068	.063	.051	.044	.052	
AVG	.030	.033	.032	.027	.024	.022	.021	.024	.031	.039	.045	.053	.057	.058	.060	.063	.063	.063	.059	.056	.048	.044	.040	.038	.044	
DATA	25	25	24	24	24	24	23	22	24	26	26	26	26	26	26	27	27	27	26	25	26	26	26	26	607	

TOTAL HOURS = 744 NUMBER OF GOOD HOURS = 607 NUMBER OF MISSING HOURS = 137 DATA CAPTURE = 81.59 (PERCENT)
 ABOVE THESE ROWS ARE TOTAL HOURLY AVERAGES, TOTAL OBSERVATIONS/HOUR AND DATA CAPTURE STATISTICS
 TOTAL AVERAGE = 0.044 MAXIMUM HOURLY VALUE = 0.112 STANDARD DEVIATION = 0.0216



U.S. EPA

LEWISBURG										OZONE				(PPM)				DATA FOR JULY								1980
										HOURS (LST)																
HR-BEG	00	01	02	03	04	05	06	07	08	09	10	11	12	13	14	15	16	17	18	19	20	21	22	23		
HR-END	01	02	03	04	05	06	07	08	09	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24		
DAY																								AVG		
1	.028	.019	.015	.016	.011	.010	.011	.019	.031	.040	.043	.045	.045	.045	.045	.046	.051	.059	.055	.044	.051	.039	.036	.032	.034	
2	.030	.023	.021	.017	.014	.013	.015	.027	.043	.059	.062	.065	.064	.070	.072	.072	.068	.073	.066	.058	.050	.038	.040	.046	.046	
3	.039	.054	.052	.047	.043	.040	.041	.039	.037	.041	.045	.039	.040	.042	.042	.040	.041	.041	.040	.040	.038	.038	.034	.028	.041	
4	.025	.024	.027	.028	.027	.023	.023	.032	.038	.043	.045	.047	.046	.043	.043	.039	.042	.043	.042	.036	.026	.024	.019	.022	.034	
5	.021	.021	.026	.023	.029	.033	.037	.038	.028	.032	.045	.054	.056	.999	.040	.038	.042	.047	.050	.053	.054	.052	.050	.043	.039	
6	.041	.050	.050	.048	.048	.049	.047	.048	.045	.046	.046	.045	.044	.047	.047	.049	.048	.050	.048	.042	.036	.032	.028	.028	.044	
7	.022	.018	.018	.017	.018	.019	.020	.026	.034	.039	.048	.053	.051	.053	.057	.053	.063	.056	.050	.044	.036	.037	.028	.027	.037	
8	.024	.023	.023	.021	.019	.021	.016	.023	.037	.049	.056	.057	.058	.060	.060	.059	.058	.049	.044	.042	.042	.999	.034	.034	.040	
9	.030	.032	.036	.039	.043	.031	.031	.038	.041	.039	.038	.043	.055	.058	.053	.053	.056	.050	.047	.047	.040	.035	.036	.033	.042	
10	.035	.035	.038	.039	.039	.040	.040	.040	.041	.041	.047	.050	.050	.053	.047	.046	.047	.048	.046	.046	.047	.046	.037	.035	.043	
11	.036	.029	.018	.021	.014	.015	.023	.999	.040	.043	.050	.054	.054	.052	.052	.052	.055	.058	.064	.072	.066	.065	.059	.050	.045	
12	.049	.045	.039	.031	.027	.027	.026	.031	.056	.063	.075	.080	.079	.074	.065	.060	.062	.059	.056	.048	.039	.037	.036	.038	.050	
13	.042	.043	.039	.029	.027	.027	.031	.035	.039	.036	.037	.041	.044	.045	.050	.048	.047	.041	.040	.041	.035	.031	.027	.026	.038	
14	.023	.022	.018	.010	.013	.016	.018	.025	.030	.047	.050	.051	.051	.049	.050	.047	.051	.055	.048	.037	.029	.030	.026	.024	.034	
15	.019	.021	.016	.010	.012	.013	.016	.019	.021	.049	.054	.056	.052	.060	.051	.052	.054	.056	.054	.999	.046	.038	.031	.023	.036	
16	.020	.016	.014	.012	.011	.009	.013	.022	.031	.050	.055	.069	.066	.060	.063	.061	.064	.066	.059	.044	.041	.036	.025	.024	.039	
17	.025	.022	.014	.014	.012	.008	.011	.025	.046	.057	.065	.069	.075	.070	.060	.053	.055	.078	.077	.075	.060	.038	.034	.031	.045	
18	.029	.025	.022	.017	.014	.017	.022	.021	.028	.044	.055	.058	.061	.063	.060	.054	.050	.052	.047	.036	.029	.028	.019	.029	.037	
19	.026	.022	.017	.014	.016	.014	.019	.015	.021	.047	.058	.062	.064	.063	.063	.061	.065	.063	.057	.039	.033	.031	.027	.023	.033	
20	.021	.016	.014	.015	.013	.016	.015	.020	.039	.046	.047	.050	.054	.053	.054	.056	.056	.057	.044	.032	.023	.021	.019	.018	.033	
21	.016	.011	.011	.013	.008	.008	.010	.018	.031	.037	.038	.036	.037	.037	.038	.041	.040	.034	.028	.027	.021	.017	.014	.016	.024	
22	.014	.010	.007	.007	.008	.005	.011	.015	.024	.033	.043	.048	.048	.037	.035	.032	.032	.037	.999	.032	.027	.027	.032	.027	.026	
23	.026	.028	.024	.025	.025	.023	.024	.022	.024	.028	.039	.047	.052	.049	.056	.063	.060	.999	.999	.999	.999	.042	.043	.036	.037	
24	.029	.025	.024	.024	.020	.021	.028	.026	.035	.045	.049	.066	.064	.065	.067	.067	.068	.063	.054	.048	.045	.035	.027	.026	.043	
25	.023	.016	.022	.023	.023	.020	.025	.026	.028	.042	.065	.076	.080	.075	.069	.073	.079	.077	.066	.050	.048	.037	.034	.029	.046	
26	.029	.030	.029	.034	.021	.028	.035	.048	.048	.062	.073	.077	.078	.078	.073	.074	.077	.066	.060	.060	.059	.055	.060	.067	.055	
27	.045	.045	.043	.038	.037	.033	.034	.039	.051	.064	.069	.070	.077	.076	.080	.073	.065	.064	.061	.055	.048	.049	.047	.037	.055	
28	.035	.032	.033	.028	.026	.025	.016	.016	.026	.027	.030	.040	.040	.039	.037	.038	.034	.028	.029	.036	.050	.062	.042	.035	.034	
29	.021	.024	.022	.022	.022	.016	.021	.033	.051	.051	.057	.060	.062	.074	.060	.074	.082	.035	.085	.077	.999	.050	.045	.037	.050	
30	.030	.032	.024	.022	.019	.022	.033	.032	.033	.040	.055	.061	.060	.060	.064	.084	.090	.087	.078	.071	.053	.049	.039	.040	.049	
31	.027	.025	.022	.021	.020	.017	.014	.015	.025	.056	.069	.069	.066	.068	.068	.064	.067	.067	.066	.060	.058	.045	.038	.030	.045	
AVG	.029	.027	.025	.023	.022	.021	.023	.028	.036	.045	.052	.056	.057	.057	.056	.056	.057	.057	.054	.048	.043	.039	.034	.032	.041	
DTYS	31	31	31	31	31	31	31	30	31	31	31	31	31	30	31	31	31	30	29	29	29	30	31	31	734	
TOTAL HOURS = 744 NUMBER OF GOOD HOURS = 734 NUMBER OF MISSING HOURS = 10 DATA CAPTURE = 98.66 (PERCENT)																										
ABOVE THREE ROWS ARE TOTAL HOURLY AVERAGES, TOTAL OBSERVATIONS/HOUR AND DATA CAPTURE STATISTICS																										
TOTAL AVERAGE = 0.041 MAXIMUM HOURLY VALUE = 0.090 STANDARD DEVIATION = 0.0176																										



MONTAGUE

OZONE

(PPM)


DATA FOR AUGUST

1980

HOURS (LST)

HR-BEGIN	00	01	02	03	04	05	06	07	08	09	10	11	12	13	14	15	16	17	18	19	20	21	22	23	AVG
HR-END	01	02	03	04	05	06	07	08	09	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	
DAY																									
1	.001	.0	.0	.0	.0	.0	.0	.007	.013	.039	.073	.090	.101	.115	.113	.102	.100	.081	.060	.028	.015	.003	.002	.0	.039
2	.001	.004	.009	.018	.014	.013	.017	.023	.043	.053	.057	.062	.067	.068	.064	.061	.055	.048	.036	.029	.027	.032	.022	.016	.035
3	.014	.020	.023	.024	.027	.022	.018	.022	.023	.028	.039	.052	.050	.045	.046	.047	.047	.048	.033	.034	.016	.011	.009	.003	.029
4	.002	.0	.0	.0	.001	.005	.006	.014	.025	.042	.050	.049	.047	.051	.053	.057	.057	.053	.030	.017	.008	.005	.001	.002	.024
5	.002	.0	.0	.0	.0	.0	.006	.016	.029	.034	.041	.056	.078	.107	.114	.098	.101	.082	.070	.052	.026	.014	.009	.004	.039
6	.002	.0	.0	.0	.0	.0	.002	.012	.033	.047	.060	.061	.059	.067	.065	.047	.047	.033	.022	.010	.003	.007	.0	.001	.024
7	.0	.0	.0	.0	.0	.0	.0	.021	.040	.050	.050	.050	.051	.050	.049	.052	.061	.999	.032	.021	.015	.005	.002	.0	.024
8	.0	.0	.0	.0	.0	.0	.001	.011	.030	.059	.070	.065	.060	.063	.065	.065	.061	.055	.040	.029	.031	.040	.047	.048	.035
9	.047	.029	.025	.031	.033	.036	.035	.040	.043	.042	.040	.045	.046	.045	.042	.046	.038	.035	.030	.024	.018	.016	.009	.007	.033
10	.004	.012	.021	.015	.010	.010	.010	.013	.018	.025	.031	.033	.037	.035	.036	.034	.036	.031	.025	.016	.010	.014	.016	.018	.021
11	.019	.014	.014	.004	.012	.032	.015	.004	.007	.007	.009	.011	.012	.011	.009	.007	.010	.005	.007	.016	.021	.017	.009	.008	.012
12	.006	.003	.004	.007	.006	.004	.003	.005	.012	.014	.030	.046	.045	.043	.047	.050	.045	.050	.038	.023	.018	.007	.004	.0	.021
13	.0	.0	.0	.0	.0	.0	.003	.006	.021	.034	.043	.042	.038	.034	.036	.034	.031	.025	.008	.005	.003	.0	.0	.0	.015
14	.0	.0	.0	.0	.0	.0	.0	.002	.008	.018	.018	.024	.023	.031	.039	.037	.042	.999	.999	.028	.027	.011	.007	.003	.014
15	.001	.003	.003	.002	.0	.0	.0	.0	.005	.018	.033	.041	.046	.048	.045	.040	.039	.035	.031	.023	.017	.008	.006	.005	.019
16	.005	.004	.002	.0	.0	.0	.003	.013	.016	.014	.015	.013	.014	.020	.022	.022	.022	.019	.020	.018	.014	.010	.006	.002	.011
17	.003	.001	.002	.001	.0	.0	.0	.009	.018	.021	.027	.029	.030	.029	.029	.032	.034	.028	.017	.012	.005	.0	.0	.0	.014
18	.0	.0	.0	.0	.0	.0	.0	.004	.013	.021	.026	.038	.051	.054	.076	.085	.093	.081	.074	.061	.052	.04			

TOTAL HOURS = 744 NUMBER OF GOOD HOURS = 736 NUMBER OF MISSING HOURS = 8 DATA CAPTURE = 98.92 (PERCENT)
 ABOVE THREE ROWS ARE TOTAL HOURLY AVERAGES, TOTAL OBSERVATIONS/HOUR AND DATA CAPTURE STATISTICS
 TOTAL AVERAGE = 0.023 MAXIMUM HOURLY VALUE = 0.115 STANDARD DEVIATION = 0.0222



SCRANTON						OZONE										(PPM)						DATA FOR AUGUST						1980
						HOURS (LST)																						
HR-BEG	00	01	02	03	04	05	06	07	08	09	10	11	12	13	14	15	16	17	18	19	20	21	22	23				
HR-END	01	02	03	04	05	06	07	08	09	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	AVG			
DAY																												
1	.004	.034	.028	.039	.051	.059	.051	.041	.050	.057	.057	.062	.070	.077	.077	.071	.067	.070	.063	.066	.064	.057	.058	.058	.056			
2	.050	.046	.044	.036	.024	.011	.004	.017	.014	.051	.052	.056	.059	.060	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999			
3	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.033	.049	.056	.059	.053	.055	.048	.042	.043	.999			
4	.035	.029	.024	.012	.009	.012	.999	.015	.033	.040	.049	.056	.060	.057	.054	.053	.056	.048	.044	.047	.034	.030	.024	.020	.037			
5	.016	.010	.013	.005	.005	.005	.001	.003	.004	.020	.057	.067	.070	.070	.067	.067	.059	.052	.051	.049	.053	.053	.048	.055	.037			
6	.054	.051	.046	.034	.026	.023	.018	.018	.034	.030	.030	.039	.042	.044	.051	.049	.048	.043	.049	.050	.040	.037	.038	.030	.036			
7	.025	.022	.021	.017	.017	.006	.004	.004	.008	.025	.059	.067	.080	.083	.081	.068	.063	.073	.076	.074	.071	.067	.073	.077	.048			
8	.073	.071	.066	.062	.054	.051	.048	.041	.040	.048	.054	.054	.057	.058	.065	.058	.062	.061	.059	.063	.067	.065	.059	.056	.058			
9	.055	.057	.055	.051	.049	.049	.051	.046	.043	.046	.050	.052	.053	.054	.057	.059	.058	.059	.062	.061	.057	.047	.043	.036	.052			
10	.032	.025	.025	.015	.012	.014	.013	.009	.008	.013	.027	.040	.046	.050	.056	.050	.058	.058	.049	.050	.045	.036	.025	.021	.032			
11	.032	.030	.026	.019	.024	.022	.012	.015	.021	.031	.042	.046	.053	.062	.064	.064	.068	.999	.063	.058	.067	.065	.064	.055	.044			
12	.051	.057	.057	.050	.048	.047	.043	.042	.040	.037	.039	.047	.058	.058	.056	.055	.054	.056	.058	.047	.049	.046	.039	.037	.049			
13	.037	.034	.035	.031	.029	.030	.022	.015	.020	.022	.029	.033	.035	.039	.047	.053	.050	.041	.037	.031	.027	.021	.020	.018	.031			
14	.023	.018	.017	.013	.011	.014	.012	.009	.009	.013	.012	.027	.059	.067	.070	.071	.082	.083	.074	.076	.076	.072	.076	.064	.044			
15	.064	.067	.065	.058	.056	.051	.047	.041	.036	.036	.045	.051	.053	.051	.055	.053	.054	.053	.054	.048	.047	.045	.044	.039	.051			
16	.036	.033	.032	.029	.027	.023	.023	.024	.022	.025	.026	.024	.026	.026	.031	.033	.033	.031	.030	.026	.025	.025	.024	.026	.027			
17	.023	.019	.021	.015	.008	.007	.006	.007	.012	.029	.033	.038	.039	.041	.042	.042	.041	.043	.042	.031	.023	.018	.018	.008	.025			
18	.009	.013	.016	.016	.006	.004	.009	.031	.042	.046	.043	.044	.046	.047	.048	.048	.044	.041	.999	.036	.030	.034	.030	.027	.031			
19	.029	.029	.024	.024	.023	.021	.019	.020	.020	.030	.040	.051	.061	.067	.066	.064	.065	.066	.059	.057	.062	.059	.059	.047	.044			
20	.044	.039	.040	.033	.027	.027	.027	.031	.036	.044	.054	.055	.059	.059	.057	.054	.054	.057	.052	.048	.048	.036	.031	.036	.044			
21	.039	.038	.035	.026	.035	.030	.030	.033	.036	.037	.042	.043	.046	.044	.043	.044	.045	.040	.038	.038	.034	.029	.025	.034	.037			
22	.034	.030	.031	.027	.031	.029	.022	.025	.030	.031	.036	.036	.040	.046	.042	.046	.044	.044	.043	.039	.039	.033	.033	.027	.035			
23	.023	.025	.023	.023	.018	.018	.019	.027	.029	.999	.999	.999	.047	.048	.048	.045	.047	.045	.047	.054	.051	.050	.050	.044	.037			
24	.045	.042	.044	.044	.037	.032	.032	.032	.041	.046	.054	.055	.057	.059	.057	.057	.056	.058	.057	.051	.051	.048	.043	.044	.048			
25	.045	.032	.030	.023	.021	.020	.015	.018	.029	.038	.043	.052	.056	.053	.053	.056	.061	.999	.043	.047	.043	.038	.040	.036	.039			
26	.036	.036	.033	.025	.022	.020	.013	.015	.024	.044	.057	.068	.070	.074	.074	.087	.097	.104	.103	.098	.093	.089	.082	.078	.060			
27	.070	.068	.063	.047	.046	.044	.034	.036	.027	.029	.073	.098	.095	.092	.091	.094	.096	.096	.096	.099	.102	.095	.084	.077	.073			
28	.076	.067	.066	.065	.056	.042	.038	.034	.064	.071	.079	.087	.089	.094	.095	.099	.092	.087	.081	.073	.072	.067	.064	.060	.072			
29	.057	.049	.045	.038	.028	.032	.023	.030	.025	.023	.040	.064	.082	.075	.065	.062	.060	.057	.047	.050	.045	.053	.055	.045	.048			
30	.061	.052	.066	.073	.063	.062	.045	.052	.065	.066	.078	.089	.091	.094	.095	.094	.088	.088	.082	.076	.065	.054	.050	.051	.071			
31	.050	.033	.028	.030	.026	.039	.044	.034	.032	.035	.045	.063	.069	.073	.077	.086	.085	.086	.073	.068	.057	.057	.057	.055	.054			
AVG	.041	.039	.037	.033	.030	.028	.025	.025	.030	.037	.046	.054	.059	.061	.062	.060	.061	.061	.058	.055	.053	.049	.047	.043	.046			
DAYS	30	30	30	30	30	30	29	30	30	29	29	29	30	30	29	30	30	28	29	30	30	30	30	30	712			

TOTAL HOURS = 744 NUMBER OF GOOD HOURS = 712 NUMBER OF MISSING HOURS = 32 DATA CAPTURE = 95.70 (PERCENT)
 ABOVE THREE ROWS ARE TOTAL HOURLY AVERAGES, TOTAL OBSERVATIONS/HOUR AND DATA CAPTURE STATISTICS
 TOTAL AVERAGE = 0.046 MAXIMUM HOURLY VALUE = 0.104 STANDARD DEVIATION = 0.0208



U.S. EPA

INDIAN RIVER

OZONE

(PPM)

DATA FOR AUGUST

1980


HOURS (LST)

HR-BEG	00	01	02	03	04	05	06	07	08	09	10	11	12	13	14	15	16	17	18	19	20	21	22	23	
HR-END	01	02	03	04	05	06	07	08	09	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	
DAY																									AVG
1	.038	.031	.028	.023	.036	.044	.040	.039	.041	.049	.067	.073	.085	.090	.097	.097	.050	.029	.016	.034	.032	.032	.035	.031	.047
2	.030	.025	.020	.015	.016	.020	.016	.022	.033	.051	.061	.069	.075	.078	.080	.081	.083	.085	.084	.068	.059	.078	.064	.058	.053
3	.055	.049	.046	.045	.045	.042	.040	.042	.043	.053	.061	.067	.071	.070	.074	.066	.066	.066	.067	.060	.057	.055	.054	.051	.056
4	.051	.042	.033	.029	.045	.033	.027	.999	.010	.026	.044	.054	.061	.067	.070	.067	.063	.060	.055	.048	.037	.029	.023	.025	.043
5	.025	.016	.010	.002	.008	.006	.011	.009	.020	.032	.041	.049	.054	.054	.057	.056	.059	.065	.072	.054	.053	.051	.045	.035	.037
6	.030	.042	.042	.039	.035	.036	.030	.030	.031	.039	.050	.049	.053	.058	.062	.059	.053	.058	.064	.058	.048	.040	.048	.047	.046
7	.036	.036	.026	.017	.015	.013	.011	.016	.030	.044	.049	.049	.061	.063	.066	.069	.069	.073	.081	.077	.074	.067	.061	.052	.048
8	.037	.035	.036	.040	.037	.015	.015	.012	.033	.046	.053	.066	.074	.075	.071	.069	.068	.074	.070	.055	.049	.049	.052	.045	.049
9	.045	.044	.040	.027	.018	.016	.011	.015	.027	.045	.055	.062	.067	.071	.073	.069	.073	.082	.091	.084	.071	.055	.045	.041	.051
10	.025	.031	.033	.030	.029	.021	.024	.014	.025	.035	.041	.044	.043	.043	.043	.046	.046	.043	.043	.038	.029	.028	.021	.018	.033
11	.021	.043	.039	.028	.013	.018	.016	.999	.999	.015	.032	.050	.059	.052	.055	.071	.076	.073	.074	.070	.040	.037	.035	.032	.043
12	.027	.023	.021	.017	.017	.013	.005	.007	.019	.022	.030	.039	.035	.035	.040	.055	.065	.077	.080	.076	.065	.048	.032	.021	.036
13	.018	.027	.028	.023	.019	.017	.010	.017	.023	.027	.040	.044	.048	.048	.053	.055	.059	.077	.083	.081	.051	.032	.026	.008	.038
14	.004	.002	.008	.016	.011	.015	.022	.013	.030	.044	.049	.052	.053	.042	.024	.023	.019	.016	.020	.019	.015	.018	.020	.022	.023
15	.027	.029	.032	.035	.030	.029	.028	.027	.029	.032	.033	.031	.034	.050	.048	.033	.025	.041	.041	.031	.024	.024	.027	.026	.032
16	.022	.026	.024	.021	.018	.015	.013	.010	.019	.025	.027	.027	.031	.038	.040	.040	.044	.046	.040	.033	.026	.022	.023	.019	.027
17	.012	.012	.012	.011	.007	.007	.004	.004	.011	.012	.015	.025	.025	.035	.040	.045	.035	.023	.025	.017	.008	.007	.005	.006	.017
18	.016	.021	.018	.015	.017	.016	.017	.003	.007	.022	.018	.021	.027	.022	.021	.017	.024	.027	.021	.022	.019	.016	.017	.017	.018
19	.017	.012	.999	.002	.008	.008	.004	.005	.002	.003	.002	.003	.008	.014	.020	.026	.034	.037	.032	.032	.028	.021	.020	.015	
20	.011	.013	.013	.011	.009	.012	.012	.011	.012	.022	.053	.070	.082	.093	.104	.112	.110	.035	.017	.030	.032	.031	.023	.026	.039
21	.045	.039	.035	.031	.026	.019	.015	.019	.017	.020	.018	.020	.023	.026	.030	.025	.029	.034	.999	.999	.018	.023	.026	.023	.025
22	.023	.021	.019	.019	.018	.023	.019	.017	.021	.021	.025	.020	.022	.022	.023	.017	.012	.018	.012	.004	.011	.033	.027	.021	.019
23	.026	.024	.018	.015	.013	.009	.009	.013	.020	.029	.031	.043	.070	.081	.088	.095	.096	.099	.119	.126	.104	.080	.073	.044	.055
24	.029	.027	.030	.021	.029	.019	.021	.026	.043	.049	.999	.999	.050	.060	.084	.070	.054	.059	.079	.077	.065	.037	.039	.036	.046
25	.010	.012	.004	.0	.002	.0	.0	.001	.021	.038	.057	.079	.093	.094	.084	.058	.056	.055	.052	.041	.036	.041	.041	.030	.038
26	.029	.033	.034	.030	.025	.025	.018	.008	.025	.048	.062	.086	.102	.110	.106	.106	.106	.077	.055	.059	.056	.048	.035	.028	.055
27	.021	.016	.010	.015	.022	.025	.032	.038	.045	.054	.068	.081	.090	.094	.101	.109	.115	.116	.108	.091	.077	.070	.062	.052	.063
28	.060	.068	.058	.043	.033	.031	.028	.027	.039	.052	.072	.091	.103	.102	.123	.136	.130	.115	.099	.097	.099	.092	.076	.047	.076
29	.040	.028	.022	.022	.026	.009	.015	.011	.031	.057	.065	.095	.106	.113	.113	.106	.091	.082	.075	.999	.999	.029	.027	.018	.054
30	.011	.008	.010	.010	.009	.009	.008	.005	.010	.014	.019	.023	.025	.025	.026	.025	.025	.023	.016	.014	.012	.011	.010	.010	.015
31	.010	.010	.010	.010	.010	.008	.006	.006	.008	.012	.019	.022	.024	.026	.027	.027	.026	.022	.018	.014	.011	.010	.010	.010	.015
AVG	.027	.027	.025	.021	.021	.018	.017	.016	.024	.033	.042	.050	.056	.060	.062	.062	.060	.058	.057	.052	.044	.039	.036	.030	.039
DAYS	31	31	30	31	31	31	31	29	30	31	30	30	31	31	31	31	31	31	30	29	30	31	31	31	734

TOTAL HOURS = 744 NUMBER OF GOOD HOURS = 734 NUMBER OF MISSING HOURS = 10 DATA CAPTURE = 98.66 (PERCENT)

ABOVE THREE ROWS ARE TOTAL HOURLY AVERAGES, TOTAL OBSERVATIONS/HOUR AND DATA CAPTURE STATISTICS

TOTAL AVERAGE = 0.039 MAXIMUM HOURLY VALUE = 0.136 STANDARD DEVIATION = 0.0266



U.S. EPA

LEWISBURG

OZONE


(PPM)

DATA FOR AUGUST

1980

HH-BEG	00	01	02	03	04	05	06	07	08	09	HOURS (LST)												15	16	17	18	19	20	21	22	23	AVG
											10	11	12	13	14	15	16	17	18	19	20	21										
HH-END	01	02	03	04	05	06	07	08	09	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24								
DAY																																
1	.030	.032	.031	.026	.037	.040	.046	.056	.059	.068	.063	.057	.062	.067	.078	.078	.075	.076	.071	.065	.049	.034	.033	.032	.053							
2	.033	.030	.025	.025	.013	.017	.012	.019	.038	.062	.067	.074	.076	.074	.071	.070	.072	.074	.082	.067	.057	.056	.051	.060	.051							
3	.050	.050	.050	.054	.053	.046	.041	.052	.062	.063	.068	.064	.062	.063	.064	.065	.069	.063	.056	.042	.049	.044	.045	.034	.055							
4	.034	.025	.027	.030	.024	.024	.021	.020	.028	.052	.063	.062	.069	.066	.056	.062	.055	.045	.040	.049	.051	.047	.043	.026	.042							
5	.027	.025	.020	.020	.016	.016	.017	.017	.028	.035	.045	.056	.056	.054	.064	.057	.056	.055	.047	.045	.999	.042	.037	.031	.038							
6	.031	.026	.024	.016	.010	.006	.010	.015	.029	.036	.039	.047	.043	.052	.053	.049	.050	.053	.042	.036	.022	.026	.021	.020	.031							
7	.022	.010	.013	.022	.021	.018	.018	.022	.026	.032	.042	.053	.065	.065	.060	.055	.050	.047	.046	.041	.030	.017	.023	.028	.034							
8	.038	.039	.039	.038	.036	.029	.029	.029	.031	.034	.042	.051	.064	.064	.055	.066	.067	.068	.067	.056	.054	.040	.028	.022	.045							
9	.018	.021	.023	.021	.015	.019	.021	.022	.036	.049	.067	.082	.079	.079	.097	.091	.075	.063	.063	.062	.057	.049	.035	.033	.049							
10	.032	.025	.023	.022	.023	.018	.016	.021	.039	.054	.064	.070	.077	.080	.074	.077	.072	.079	.072	.055	.047	.037	.025	.030	.047							
11	.037	.034	.027	.028	.029	.021	.018	.023	.033	.055	.071	.067	.073	.074	.072	.072	.072	.071	.069	.067	.065	.065	.051	.049	.052							
12	.051	.060	.060	.051	.049	.048	.045	.044	.038	.042	.048	.054	.058	.063	.070	.071	.069	.074	.075	.999	.051	.046	.039	.034	.054							
13	.025	.020	.020	.025	.022	.019	.019	.024	.025	.034	.048	.069	.067	.065	.067	.068	.072	.068	.063	.045	.039	.035	.024	.017	.041							
14	.019	.023	.028	.021	.018	.017	.019	.021	.029	.041	.067	.071	.072	.075	.071	.070	.067	.065	.061	.060	.046	.037	.030	.026	.044							
15	.021	.022	.020	.020	.018	.014	.012	.016	.034	.053	.060	.067	.070	.071	.066	.065	.061	.060	.050	.039	.048	.053	.056	.064	.044							
16	.054	.052	.050	.049	.045	.038	.035	.038	.040	.037	.048	.051	.051	.049	.047	.042	.040	.044	.044	.040	.031	.024	.023	.019	.041							
17	.022	.034	.037	.037	.034	.033	.034	.031	.032	.034	.036	.036	.037	.038	.034	.041	.040	.035	.033	.029	.028	.025	.022	.017	.032							
18	.018	.016	.018	.020	.016	.014	.009	.012	.014	.013	.018	.021	.018	.025	.035	.030	.031	.027	.031	.038	.036	.038	.035	.031	.023							
19	.031	.030	.026	.023	.025	.027	.026	.023	.023	.027	.037	.046	.043	.046	.050	.045	.046	.048	.047	.037	.023	.019	.021	.018	.033							
20	.020	.021	.025	.031	.032	.035	.031	.999	.032	.037	.040	.044	.053	.060	.063	.065	.059	.051	.049	.043	.041	.028	.018	.020	.039							
21	.025	.022	.023	.027	.029	.023	.018	.026	.036	.037	.044	.045	.055	.064	.072	.069	.058	.051	.047	.045	.031	.021	.046	.048	.040							
22	.046	.041	.039	.037	.035	.032	.025	.036	.040	.041	.041	.046	.050	.054	.049	.044	.050	.052	.048	.043	.042	.039	.042	.042	.042							
23	.034	.034	.032	.032	.034	.031	.033	.043	.049	.052	.054	.055	.056	.052	.053	.049	.045	.045	.035	.033	.029	.028	.022	.019	.040							
24	.019	.021	.019	.020	.021	.025	.027	.032	.034	.035	.033	.038	.048	.053	.051	.050	.052	.046	.040	.029	.027	.017	.017	.015	.032							
25	.013	.013	.014	.016	.017	.023	.029	.026	.029	.033	.047	.056	.058	.060	.059	.059	.055	.049	.043	.031	.027	.021	.019	.019	.034							
26	.015	.009	.008	.010	.009	.013	.015	.017	.022	.037	.058	.070	.069	.071	.072	.065	.063	.057	.048	.034	.028	.024	.023	.023	.036							
27	.019	.016	.011	.012	.014	.018	.014	.016	.021	.031	.056	.069	.073	.074	.074	.073	.067	.067	.055	.035	.030	.031	.025	.017	.038							
28	.018	.014	.015	.013	.014	.013	.013	.015	.018	.029	.055	.066	.079	.079	.084	.085	.080	.070	.049	.999	.999	.025	.020	.020	.040							
29	.020	.019	.014	.021	.018	.021	.017	.023	.028	.037	.053	.075	.079	.078	.076	.082	.090	.085	.076	.057	.048	.037	.032	.029	.046							
30	.021	.025	.021	.024	.019	.020	.021	.027	.032	.045	.063	.069	.062	.057	.049	.048	.048	.047	.041	.037	.028	.022	.029	.030	.037							
31	.026	.028	.025	.025	.025	.023	.022	.020	.025	.030	.038	.050	.056	.056	.060	.060	.059	.049	.047	.042	.041	.039	.029	.021	.037							
AVG	.028	.027	.026	.026	.025	.024	.023	.026	.033	.041	.051	.057	.061	.062	.063	.062	.060	.058	.053	.045	.040	.034	.031	.029	.041							
DAYS	31	31	31	31	31	31	31	30	31	31	31	31	31	31	31	31	31	31	31	29	29	31	31	31	739							

TOTAL HOURS = 744 NUMBER OF GOOD HOURS = 739 NUMBER OF MISSING HOURS = 5 DATA CAPTURE = 99.33 (PERCENT)
 ABOVE THREE ROWS ARE TOTAL HOURLY AVERAGES, TOTAL OBSERVATIONS/HOUR AND DATA CAPTURE STATISTICS
 TOTAL AVERAGE = 0.041 MAXIMUM HOURLY VALUE = 0.097 STANDARD DEVIATION = 0.0190



DATA FOR SEPTEMBER 1980

30


TOTAL HOURS = 288 NUMBER OF GOOD HOURS = 286 NUMBER OF MISSING HOURS = 2 DATA CAPTURE = 99.31 (PERCENT)
 ABOVE THESE FIGURES ARE TOTAL HOURLY AVERAGES, TOTAL OBSERVATIONS/HOUR AND DATA CAPTURE STATISTICS
 TOTAL AVERAGE = 0.027* MAXIMUM HOURLY VALUE = 0.119 STANDARD DEVIATION = 0.0210



U.S. EPA

STATION		OZONE											(PPM)					DATA FOR SEPTEMBER 1980									
													HOURS (LST)														
HH-BEG	MM	00	01	02	03	04	05	06	07	08	09	10	11	12	13	14	15	16	17	18	19	20	21	22	23		
HH-END	MM	01	02	03	04	05	06	07	08	09	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24		
DAY																											AVG
1		.049	.044	.035	.028	.029	.031	.022	.029	.038	.060	.067	.070	.073	.074	.080	.087	.089	.079	.075	.070	.067	.065	.065	.053	.057	
2		.045	.048	.046	.047	.043	.041	.036	.020	.030	.048	.059	.062	.067	.078	.081	.076	.071	.999	.999	.035	.037	.031	.026	.026	.048	
3		.024	.035	.036	.040	.042	.038	.020	.027	.038	.042	.046	.049	.053	.053	.056	.058	.057	.056	.051	.049	.044	.042	.035	.031	.043	
4		.024	.025	.021	.013	.010	.007	.008	.012	.014	.016	.038	.058	.070	.073	.074	.078	.081	.087	.087	.095	.096	.093	.090	.095	.053	
5		.045	.076	.073	.062	.052	.052	.045	.038	.040	.047	.052	.059	.064	.066	.071	.071	.071	.069	.061	.054	.056	.060	.051	.059		
6		.044	.043	.036	.026	.019	.016	.013	.015	.020	.018	.038	.053	.056	.046	.058	.060	.061	.069	.072	.073	.073	.064	.058	.046		
7		.035	.030	.046	.036	.037	.035	.028	.025	.026	.031	.035	.037	.041	.045	.043	.045	.043	.042	.044	.046	.044	.035	.035	.032	.039	
8		.033	.027	.024	.020	.015	.013	.011	.011	.019	.037	.040	.042	.042	.042	.041	.041	.043	.041	.999	.029	.025	.027	.027	.027	.029	
9		.025	.017	.015	.016	.011	.010	.005	.010	.012	.030	.054	.064	.072	.077	.082	.080	.083	.080	.081	.070	.076	.067	.042	.046		
10		.029	.020	.026	.021	.018	.018	.015	.019	.024	.025	.030	.034	.036	.037	.041	.043	.042	.045	.044	.040	.039	.034	.029	.030		
11		.028	.023	.017	.013	.013	.013	.006	.005	.007	.020	.052	.057	.058	.060	.063	.065	.057	.058	.056	.054	.057	.056	.056	.055	.040	
12		.035	.030	.044	.036	.037	.029	.025	.021	.025	.028	.040	.054	.054	.062	.069	.071	.068	.999	.999	.062	.053	.038	.039	.041	.046	
13		.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
14		.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
15		.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
16		.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
17		.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
18		.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
19		.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
20		.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
21		.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
22		.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
23		.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
24		.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
25		.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
26		.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
27		.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
28		.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
29		.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
30		.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
TOT		.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.045	
NO		12	12	12	12	12	12	12	12	12	12	12	12	12	12	12	12	10	9	12	12	12	12	12	12	253	

TOTAL HOURS = 288 NUMBER OF GOOD HOURS = 283 NUMBER OF MISSING HOURS = 5 DATA CAPTURE = 98.26 (PERCENT)
 ABOVE THREE ROWS ARE TOTAL HOURLY AVERAGES, TOTAL OBSERVATIONS/HOUR AND DATA CAPTURE STATISTICS
 TOTAL AVERAGE = 0.045* MAXIMUM HOURLY VALUE = 0.096 STANDARD DEVIATION = 0.0212



U.S. EPA

INDIAN RIVER


OZONE

(PPM)

DATA FOR SEPTEMBER 1980

DAY	HOURS (LST)																								AVG
	00	01	02	03	04	05	06	07	08	09	10	11	12	13	14	15	16	17	18	19	20	21	22	23	
1	.014	.013	.012	.010	.009	.008	.008	.008	.009	.018	.032	.046	.054	.062	.060	.062	.050	.047	.036	.026	.022	.021	.016	.017	.027
2	.010	.015	.013	.008	.007	.007	.006	.006	.013	.022	.034	.045	.056	.063	.058	.056	.064	.073	.065	.051	.035	.029	.022	.022	.033
3	.021	.024	.025	.024	.025	.022	.019	.018	.020	.024	.031	.027	.036	.047	.058	.065	.060	.057	.949	.999	.002	.001	.001	.002	.028
4	.002	.002	.003	.001	.001	.0	.0	.0	.002	.012	.023	.032	.038	.044	.043	.045	.044	.047	.045	.045	.039	.029	.018	.017	.022
5	.020	.020	.024	.010	.004	.011	.007	.011	.015	.025	.032	.020	.023	.025	.018	.022	.031	.032	.034	.023	.012	.013	.003	.0	.019
6	.002	.001	.001	.002	.001	.002	.0	.001	.005	.010	.018	.024	.040	.046	.045	.052	.063	.062	.053	.046	.031	.016	.013	.010	.015
7	.010	.010	.014	.010	.009	.042	.036	.026	.024	.034	.040	.046	.056	.055	.051	.045	.042	.047	.045	.035	.032	.027	.024	.021	.033
8	.010	.003	.0	.001	.001	.003	.006	.005	.007	.015	.018	.021	.029	.036	.034	.043	.043	.048	.046	.042	.032	.019	.013	.017	.021
9	.023	.017	.022	.014	.005	.010	.002	.0	.014	.036	.045	.058	.063	.059	.061	.059	.068	.071	.068	.055	.048	.046	.044	.044	.039
10	.039	.033	.030	.029	.029	.025	.025	.010	.015	.016	.024	.020	.017	.020	.018	.027	.034	.038	.033	.017	.009	.001	.002	.0	.021
11	.002	.003	.009	.009	.009	.003	.004	.005	.008	.016	.020	.024	.048	.054	.061	.066	.075	.059	.054	.046	.029	.016	.021	.023	.028
12	.020	.030	.023	.024	.015	.010	.025	.015	.023	.035	.047	.065	.071	.081	.091	.080	.077	.074	.075	.999	.999	.042	.058	.056	.043
13	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999
14	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999
15	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999
16	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999
17	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999
18	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999
19	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999
20	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999
21	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999
22	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999
23	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999
24	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999
25	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999
26	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999
27	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999
28	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999
29	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999
30	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999
AVG	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.026
DAYS	12	12	12	12	12	12	12	12	12	12	12	12	12	12	12	12	12	12	11	10	11	12	12	12	234


TOTAL HOURS = 288 NUMBER OF GOOD HOURS = 284 NUMBER OF MISSING HOURS = 4 DATA CAPTURE = 98.61 (PERCENT)
 ABOVE TOTAL HOURS ARE TOTAL HOURLY AVERAGES, TOTAL OBSERVATIONS/HOUR AND DATA CAPTURE STATISTICS
 TOTAL AVERAGE = 0.026 MAXIMUM HOURLY VALUE = 0.091 STANDARD DEVIATION = 0.0206



U.S. EPA

LAISEGAL						OZONE						(PPM)						DATA FOR SEPTEMBER								1980
		HOURS (LST)																								
HR-BEG	00	01	02	03	04	05	06	07	08	09	10	11	12	13	14	15	16	17	18	19	20	21	22	23		
HR-END	01	02	03	04	05	06	07	08	09	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24		
DAY																								AVG		
1	.021	.016	.020	.010	.010	.016	.018	.024	.025	.038	.047	.047	.051	.049	.046	.051	.048	.049	.041	.031	.023	.022	.014	.010	.031	
2	.015	.011	.011	.013	.010	.010	.008	.010	.019	.037	.047	.050	.049	.052	.051	.054	.059	.063	.049	.045	.045	.047	.044	.035	.035	
3	.030	.034	.033	.030	.021	.016	.013	.024	.030	.034	.038	.037	.041	.045	.049	.051	.050	.052	.038	.034	.023	.022	.022	.013	.033	
4	.025	.010	.010	.021	.010	.019	.020	.023	.030	.053	.060	.061	.065	.065	.065	.066	.064	.059	.054	.054	.039	.026	.023	.023	.040	
5	.023	.025	.025	.024	.025	.031	.020	.030	.027	.035	.044	.050	.055	.054	.055	.057	.063	.068	.070	.999	.055	.038	.038	.023	.041	
6	.025	.020	.024	.030	.031	.034	.045	.044	.053	.047	.048	.052	.062	.064	.067	.073	.080	.080	.078	.066	.050	.046	.045	.036	.050	
7	.031	.028	.027	.023	.024	.020	.022	.032	.039	.040	.061	.065	.061	.060	.060	.061	.062	.058	.051	.042	.034	.030	.023	.019	.041	
8	.020	.023	.020	.010	.010	.018	.013	.020	.024	.034	.050	.061	.069	.071	.072	.070	.069	.067	.047	.037	.027	.023	.021	.020	.030	
9	.019	.022	.019	.017	.017	.015	.013	.021	.026	.042	.055	.071	.072	.070	.070	.067	.062	.056	.043	.037	.034	.031	.030	.029	.039	
10	.040	.045	.047	.044	.043	.045	.057	.050	.057	.041	.043	.044	.047	.048	.046	.051	.044	.999	.999	.999	.999	.033	.021	.020	.051	
11	.017	.015	.015	.012	.014	.011	.010	.013	.017	.025	.042	.052	.058	.060	.059	.056	.056	.052	.044	.033	.026	.005	.021	.019	.030	
12	.017	.017	.017	.015	.014	.013	.010	.014	.020	.035	.040	.062	.074	.070	.060	.063	.062	.060	.040	.039	.034	.029	.020	.023	.036	
13	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
14	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
15	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
16	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
17	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
18	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
19	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
20	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
21	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
22	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
23	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
24	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
25	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
26	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
27	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
28	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
29	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
30	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	
206	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.999	.033	
DAYS	12	12	12	12	12	12	12	12	12	12	12	12	12	12	12	12	12	11	11	10	11	12	12	12	283	

TOTAL HOURS = 288 NUMBER OF GOOD HOURS = 263 NUMBER OF MISSING HOURS = 5 DATA CAPTURE = 98.26 (PERCENT)
 ABOVE TABLE ROWS ARE TOTAL HOURLY AVERAGES, TOTAL OBSERVATIONS/HOUR AND DATA CAPTURE STATISTICS
 TOTAL AVERAGE = 0.0308 MAXIMUM HOURLY VALUE = 0.080 STANDARD DEVIATION = 0.0183



APPENDIX B

Daily span delta percents were calculated using the following equation:

$$\text{Delta \%} = \frac{Y - X}{X} \times 100$$

where:

Y = Analyzer's indicated concentration

X = Known concentration of generated test atmosphere

DAILY SPAN (DELTA%) SUMMARY FOR U.S. EPA OZONE JULY 1980

	MONTAGUE	SCRANTON	INDIAN RIVER	LEWISBURG
	+6.5	-7.0	NO SPAN	-17.94
	+3.2	-5.0	NO SPAN	-17.94
	0	-7.8	-6.7	-19.85
	-3.2	-7.8	-5.5	-19.85
	0	-10.9	-8.7	-19.85
	-3.2	-8.9	-9.4	-17.94
	-1.6	NO SPAN	-7.5	-19.85
	+1.6	-7.8	-5.5	-19.85
	NO SPAN	-10.9	-7.1	NO SPAN
	-3.2	-11.6	-8.7	NO SPAN
	-3.2	-11.6	-7.9	-23.66
	-1.6	-7.7	-8.7	-19.85
	0	-5.5	-14.6	-21.76
	-1.6	-7.7	-12.2	-18.60
	-1.6	-6.6	-14.6	-18.60
	-3.2	-8.8	-14.6	-18.60
	-3.2	-12.1	-14.6	-18.60
	-3.2	NO SPAN	NO SPAN	-18.60
	-1.6	-12.1	NO SPAN	-24.42
	-3.2	-17.6	NO SPAN	-24.42
	-6.4	NO SPAN	NO SPAN	-18.60
	-6.4	NO SPAN	-12.2	NO SPAN
	-22.6	-2.2	-14.6	NO SPAN
	-1.6	-3.3	-15.9	-1.16
	+1.6	-5.5	-14.6	-1.16
	+1.6	-6.6	-14.6	-1.16
	-3.2	-6.6	-14.6	-1.16
	-3.2	-6.6	-14.6	-1.16
	-3.2	-2.2	-15.9	-1.16
	-12.9	-3.3	-15.9	-1.16
	0	NO SPAN	-15.9	-1.16

DAILY SPAN (DELTA%) SUMMARY FOR U.S. EPA OZONE AUGUST 1980

Y	MONTAGUE	SCRANTON	INDIAN RIVER	LEWISBURG
	0	-2.2	-17.1	-1.16
1	-1.6	-5.5	-23.2	-1.16
1	-3.0	NO SPAN	-23.2	NO SPAN
1	-4.8	NO SPAN	NO SPAN	-1.16
3	-3.0	-3.3	-20.7	-1.16
3	-3.0	-4.4	-20.7	-1.16
7	-3.0	-7.7	-20.7	-1.16
8	-1.6	-11.0	-20.7	-1.16
9	-4.8	NO SPAN	-20.7	-1.16
9	-4.8	-8.8	-20.7	-1.16
1	NO SPAN	NO SPAN	-20.7	NO SPAN
2	1.6	-7.7	-23.2	-1.16
3	1.6	-3.3	-20.7	-1.16
4	3.0	-3.3	-23.2	-1.16
5	3.0	-4.4	-23.2	-1.16
6	1.6	-3.3	-20.7	-1.16
7	3.0	NO SPAN	-18.3	-1.16
8	3.0	-5.5	-19.5	-1.16
9	-3.0	-4.4	NO SPAN	-1.16
9	-3.0	-3.3	-18.3	-1.16
1	-3.0	-3.3	NO SPAN	-1.16
2	-1.6	-3.3	NO SPAN	-1.16
3	-1.6	-3.3	NO SPAN	NO SPAN
4	0	-5.5	NO SPAN	-1.16
5	0	-3.3	NO SPAN	-1.16
5	-1.6	-6.6	-2.4	-1.16
7	-3.0	-5.5	NO SPAN	-1.16
8	-3.0	-7.7	NO SPAN	-1.16
9	0	-6.6	NO SPAN	-1.16
0	4.8	-13.2	-26.8	-1.16
1	-1.6	-14.3	-25.8	-1.16

DAILY SPAN SUMMARY (DELTA %) FOR U.S. EPA OZONE SEPTEMBER 1980

AY	MONTAGUE	SCRANTON	INDIAN RIVER	LEWISBURG
1	-4.8	-13.2	-26.8	0
2	-3.0	-15.4	-26.8	-1.16
3	-1.6	-9.9	-26.8	FULL SCALE
4	1.6	-5.5	-32.9	FULL SCALE
5	-3.0	-9.9	-26.8	FULL SCALE
6	0	-6.6	NO SPAN	1.16
7	-3.0	-9.9	-26.8	0
8	1.6	-7.7	-26.8	-1.16
9	1.6	-12.1	-26.8	0
10	-3.0	-12.1	-26.8	1.16
11	0	-7.7	-26.8	-4.65
12	4.8	-8.8	NO SPAN	-3.49

MULTIPOINT SUMMARY SHEET

NETWORK: USEPA O3 STUDY

TEAM/ANALYST: TEAM B/ ARTHUR C.

SITE	PAR.	DATE	REASON	S/N	%	COMMENTS/DATA ACTION
*****	*****	*****	*****	*****	*****	*****
1	03	7-17-80	TRACKING	75	-2.2%	
1	03	7-23-80	TRACKING	75	-4.4%	
1	03	7-30-80	TRACKING	75	-6.6%	
2	03	7-07-80	TRACKING	70	-12.3%	
2	03	7-13-80	TRACKING	70	-13.7%	
2	03	7-22-80	DASIBI-CAL	70	-16.2%	
2	03	7-28-80	TRACKING	70	-6.2%	
3	03	7-07-80	TRACKING	78	-0.7%	
3	03	7-14-80	TRACKING	78	-7.4%	
3	03	7-21-80	TRACKING	78	-8.0%	
3	03	7-28-80	TRACKING	78	-9.9%	
4	03	6-30-80	TRACKING	63	-18.5%	
4	03	7-08-80	TRACKING	63	-19.3%	
4	03	7-11-80	AUDIT	63	-23.0%	
4	03	7-15-80	TRACKING	63	-20.3%	
4	03	7-22-80	TRACKING	63	-25.4%	
4	03	7-23-80	DASIBI-CAL	63	-19.9%	
4	03	7-29-80	TRACKING	63	-3.6%	
*****	*****	*****	*****	*****	*****	*****

AUGUST MP'S

1	03	8-07-80	TRACKING	75	-5.6%
1	03	8-14-80	TRACKING	75	-0.05%
1	03	8-21-80	TRACKING	75	-4.6%
1	03	8-24-80	AUDIT	75	-13.8%
1	03	8-28-80	TRACKING	75	-2.8%
2	03	8-04-80	TRACKING	70	-10.4%
2	03	8-11-80	TRACKING	70	-8.1%
2	03	8-18-80	TRACKING	70	-7.7%
2	03	8-23-80	AUDIT	70	+4.9%
2	03	8-25-80	TRACKING	70	-9.9%
3	03	8-04-80	TRACKING	78	-11.4%
3	03	8-11-80	TRACKING	78	-15.8%
3	03	8-19-80	TRACKING	78	-16.6%
3	03	8-21-80	DASIBI-CAL	78	-5.8%
3	03	8-24-80	AUDIT	78	+12.5%
3	03	8-29-80	TRACKING	78	-18.9%
4	03	8-05-80	TRACKING	63	-2.7%
4	03	8-12-80	TRACKING	63	-4.7%
4	03	8-20-80	TRACKING	63	-5.0%
4	03	8-28-80	TRACKING	63	-2.3%

SEPTEMBER MPS

1	03	9-04-80	TRACKING	75	-0.7%	
1	03	9-11-80	TRACKING	75	+4.2%	
1	03	9-25-80	DASIBI-CAL	75	-1.27%	SHUTDOWN
2	03	9-02-80	TRACKING	70	-13.1%	
2	03	9-08-80	TRACKING	70	-13.1%	
2	03	9-12-80	DASIBI-CAL	70	+1.13%	SHUTDOWN
3	03	9-03-80	TRACKING	78	-23.4%	
3	03	9-12-80	TRACKING	78	-24.5%	
3	03	9-17-80	DASIBI-CAL	78	-9.23%	SHUTDOWN
4	03	9-16-80	DASIBI-CAL	63	-4.3%	SHUTDOWN

APPENDIX C

ERT STANDARD OPERATING PROCEDURES

<u>Number</u>	<u>Rev.</u>	<u>Title</u>
2630	1	Routine Field Operations and Maintenance
2600	1	Field Calibration of Continuous Air Quality Analyzers
2901-001	0	Documentation of Field Calibration of Continuous Air Quality Analyzers
2630-005	0	Calibration Control Logs and Charts
2000-261	A	Multipoint Span Check/Field Calibration: Monitor Labs 8410 Ozone Analyzer
2600-850	0	Strip Chart Recorder Calibration
2600-100	0	Calibration Check/Audit Using a Dasibi 1003 PC or 1003 RS
2750	0	Test Equipment Calibration and Repair
2600-001	0	Precision Checks and Zero and Span Checks of Continuous Analyzers
2400	0	Traceability of Standards
2580-001	0	Flowmeter Calibration
4000	0	Data Validation