

EPA-600/4-76-032
June 1976

Environmental Monitoring Series

REGIONAL AIR POLLUTION STUDY: QUALITY ASSURANCE AUDITS



**Environmental Sciences Research Laboratory
Office of Research and Development
U.S. Environmental Protection Agency
Research Triangle Park, North Carolina 27711**

RESEARCH REPORTING SERIES

Research reports of the Office of Research and Development, U.S. Environmental Protection Agency, have been grouped into five series. These five broad categories were established to facilitate further development and application of environmental technology. Elimination of traditional grouping was consciously planned to foster technology transfer and a maximum interface in related fields. The five series are:

1. Environmental Health Effects Research
2. Environmental Protection Technology
3. Ecological Research
4. Environmental Monitoring
5. Socioeconomic Environmental Studies

This report has been assigned to the ENVIRONMENTAL MONITORING series. This series describes research conducted to develop new or improved methods and instrumentation for the identification and quantification of environmental pollutants at the lowest conceivably significant concentrations. It also includes studies to determine the ambient concentrations of pollutants in the environment and/or the variance of pollutants as a function of time or meteorological factors.

This document is available to the public through the National Technical Information Service, Springfield, Virginia 22161.

REGIONAL AIR POLLUTION STUDY:

QUALITY ASSURANCE AUDITS

By

John R. Hibar
Air Monitoring Center
Rockwell International
Creve Coeur, MO 63141

Contract No. 68-02-1081
Task Order 58

Project Officer

Francis A. Schiermeier
Regional Air Pollution Study
Environmental Sciences Research Laboratory
11640 Administration Drive
Creve Coeur, MO 63141

U.S. ENVIRONMENTAL PROTECTION AGENCY
OFFICE OF RESEARCH AND DEVELOPMENT
ENVIRONMENTAL SCIENCES RESEARCH LABORATORY
RESEARCH TRIANGLE PARK, N.C. 27711

DISCLAIMER

This report has been reviewed by the Environmental Sciences Research Laboratory, U.S. Environmental Protection Agency, and approved for publication. Approval does not signify that the contents necessarily reflect the views and policies of the U.S. Environmental Protection Agency, nor does mention of trade names or commercial products constitute endorsement or recommendation for use.

ABSTRACT

Aerometric data is being collected by the Regional Air Pollution Study (RAPS) for use in developing and evaluating air quality simulation models. In addition to the Regional Air Monitoring System (RAMS), data is also collected by research teams in periodic expeditions to the St. Louis study area. Data from all sources are made available for integration through the RAPS Data Bank. A quality assurance audit of instrument systems employed in the Summer 1975 RAPS Expeditionary Research Program was conducted to check the various systems for accuracy. Additionally such checks provide a basis for determining the extent to which data from different instrument systems may be integrated. This report describes the audit equipment and standards used and problems encountered. Quantitative audit results from individual instrument systems are presented. The audits included analyzers for NO , NO_x , O_3 , SO_2 , CO , CH_4 and THC measurements.

CONTENTS

ABSTRACT	iii
FIGURES	v
TABLES	vi
1.0 INTRODUCTION	1
2.0 TASK ORDER REQUIREMENTS	2
2.1 TECHNICAL REQUIREMENTS	2
2.2 EQUIPMENT	2
2.3 PERIOD OF PERFORMANCE	2
2.4 REPORTS	2
3.0 WORK PERFORMED	3
3.1 VAN PREPARATION	3
3.1.1 Physical Support System	3
3.1.2 Instrumentation	3
3.2 CALIBRATION METHODS	3
3.2.1 NO-NO _x Source Calibration	5
3.2.2 SO ₂ Source Calibration	5
3.2.3 CO and CH ₄ Source Calibration	5
3.2.4 O ₃ Source Calibration	5
3.3 INSTRUMENT CALIBRATION METHOD	6
3.4 PREPARATION OF CALIBRATION CURVES	6
3.5 AUDIT SITES	8
3.5.1 RAPS Helicopters	8
3.5.2 Meteorology Research Incorporated Plane (MRI)	8
3.5.3 Aerosol Trailer	9
3.5.4 Research Triangle Institute Van	9
3.5.5 Environmental Measurements Incorporated Van (EMI)	9
3.5.6 RAMS Stations	9
3.5.7 Lou Chaney's O ₃ and CO Monitors	10
3.5.8 Environmental Quality Research (EQR) Theta Sensor	10
3.5.9 Battelle Northwest DC-3	10
3.5.10 Gas Chromatography Laboratory	11
3.6 RAW DATA	11
4.0 PROBLEMS AND RECOMMENDATIONS	12

FIGURES

NUMBER

PAGE

- | | |
|---|----------------------------|
| 1 | POWER CIRCUIT MODIFICATION |
| 2 | DYNAMIC CALIBRATION SYSTEM |

4

7

TABLES

<u>NUMBER</u>		<u>PAGE</u>
1	RAPS HELICOPTER #1	13
2	RAPS HELICOPTER #3	14
3	RAPS HELICOPTER #3	15
4	MRI PLANE	16
5	MRI PLANE	17
6	MRI PLANE	18
7	AEROSOL TRAILER	19
8	AEROSOL TRAILER	20
9	AEROSOL TRAILER	21
10	RTI TRAILER	22
11	EMI VAN	23
12	EMI CALIBRATOR	24
13	STATION 120	25
14	STATION 120	26
15	STATION 120	27
16	STATION 113	28
17	STATION 113	29
18	STATION 120	30
19	STATION 122	31
20	LOU CHANEY O ₃ MONITORS	32
21	CO ANALYSIS FOR LOU CHANEY	33
22	O ₃ ANALYSIS FOR LOU CHANEY	34
23	CALIBRATION DIFFERENCE SUMMARY	35

1.0 INTRODUCTION

This report summarizes the preparation and use of a mobile calibration system for the RAPS 1975 Summer Study Quality Assurance Audits. The purpose of these audits was to check the analyzers of the RAPS field investigations and RAMS stations for accuracy. A mobile laboratory van was used to carry a controlled environment to the field. A brief description of the methods used to acquire the data and a listing of the pertinent data is included.

2.0 TASK ORDER REQUIREMENTS

2.1 TECHNICAL REQUIREMENTS

Under this task order, the contractor was to:

- A) Prepare the EPA mobile laboratory for quality assurance audits.
- B) Drive the van to various sites and perform calibration checks of various air monitoring systems including the following:
 - 1) Operate and maintain chemiluminescent oxone analyzers, chemiluminescent NO-NO_x analyzers, total sulfur analyzers, sulfur chromatographs, Beckman Model 6800 gas chromatographs, integrating nephelometers, and strip-chart recorders.
 - 2) Prepare and operate the gas dilution systems, gas sampling systems, and air purification systems.
 - 3) Compute the concentrations of standard gas mixes, preparation of calibration curves, and compile and record data.

2.2 EQUIPMENT

The Winnebago laboratory van, instruments, and calibration gases were to be supplied by the EPA Task Order Project Officer.

2.3 PERIOD OF PERFORMANCE

Task order duration - the task order was estimated to require 11 weeks ending September 5, 1975.

2.4 REPORTS

Monthly progress reports, draft final reports, and an approved final report were to be furnished in accordance with provisions of the EPA contract 68-02-1081.

3.0 WORK PERFORMED

The work performed consisted of preparing the van for mobile calibration use, calibrating the transfer standards using National Bureau of Standards (NBS) reference materials, and performing audits on the measurement systems of independent investigators and selected RAMS stations.

During the performance of this task the Air Monitoring Center provided the services of a senior engineer and a technician to perform the required tasks. The specific calibration procedures employed and the schedule of sites and analyzers were received from the EPA Task Coordinator, S.L. Kopczynski, in the form of technical direction on a daily basis.

3.1 VAN PREPARATION

The van preparation included modifications to the physical support system and installation of the instrumentation system.

3.1.1 Physical Support System

The Winnebago van originally had a single generator for the entire vehicle when it was not using line power. In order not to overload this generator and to prevent surges in the instruments when the air conditioner started, another generator was installed by EPA. This generator was only used on the air conditioner with the other handling the instruments. The second generator was wired into the system to allow either A.C. line power or the generator to power the air conditioner (See Figure 1).

3.1.2 Instrumentation

The government supplied instruments consisted of a Bendix NO-NO_x analyzer, a Bendix Ozone analyzer, a Tracor sulfur chromatograph and a Bendix dynamic calibration system. These instruments were calibrated and installed in the van with the appropriate auxiliary gases necessary for operation. The plans for using the mobile laboratory instruments for cross-checking analyses was dropped in favor of more thorough multipoint calibration checks which utilized all budgeted manpower.

3.2 CALIBRATION METHODS

The gas sources used in calibration were directly traceable to NBS where possible. These concentrated gases were diluted to several values and a corresponding multipoint calibration performed on the analyzers.




FIGURE 1
POWER CIRCUIT MODIFICATION

3.2.1 NO-NO_x Source Calibration

The mobile lab contained a tank of approximately 100 ppm of NO, which was used as the standard for NO-NO_x calibrations. This tank was initially calibrated using a NBS NO cylinder #RSG-30-7963 which contained 93.4 ppm as the standard reference. This calibration was done in the GC lab using a Bendix NO-NO_x analyzer (which was calibrated with the NBS tank) to measure several known mixtures of the unknown tank. Efforts were made to have the same concentrations as those used in the initial calibration of the instrument.

3.2.2 SO₂ Source Calibration

Two NBS SO₂ permeation tubes were used as the standard for all audits. Since the permeation rate of the tube is highly temperature dependent, it was placed in an oven in the Bendix Dynamic Calibrator which had provisions for both heating as well as cooling in order to maintain a temperature of 25°C, \pm 0.1°C. The output of the Bendix Dynamic Calibrator was cross-checked at the RAMS station 120 with an NBS permeation tube installed in the station calibration system.

3.2.3 CO and CH₄ Source Calibration

An NBS CO tank containing 9.82 ppm CO was used in conjunction with the GC laboratory's 6800 to certify three Scott tanks, one containing a 21.2 ppm of CO (#L1722), one containing 2.01 ppm CH₄ and 5.24 ppm CO (#1749) and one containing 2.04 ppm CH₄ and 5.20 ppm CO (#2359). These tanks were either used directly to give a single point or dynamically diluted with either Matheson zero air or Scott ultrapure air to give a multipoint calibration. The zero air was also checked for residue CO and was found to be within acceptable tolerances.

3.2.4 O₃ Source Calibration

The O₃ used for the audits was generated on site through use of a uv tube which was assumed to have adequate short term stability but would vary from day to day. For this reason the O₃ output was calibrated each time the O₃ was used by means of an O₃ - NO titration immediately after calibration of an NO-NO_x analyzer. Since the instrument's calibration could be traced to a NBS NO tank, the drop in NO when the O₃ was turned on gave a calibrated O₃ value at that flow setting. Others were obtained by varying the dilution flow for a multipoint calibration.

3.3 INSTRUMENT CALIBRATION METHOD

The calibration system employed was a Bendix Dynamic Calibrator which used precision pressure regulators to maintain a given pressure differential across sets of capillaries which in turn would give a constant flow. These flows were calibrated for various pressure settings using a bubble meter at the beginning of the study. Near the end of the study, a bubble meter and mass flow meters #FM-302 and FM-300 were used to recalibrate Input C capillary #4 and check the others.


There are three pressure regulators; one controls the flow across the permeation oven which has two flow levels - a high flow and a low flow. The second controls the flow of concentrated calibrating gas to be diluted by the zero air from the third regulator. The third regulator has four capillaries associated with it for generating a wide range of dilutions. Air from one of these capillaries flows through an ozone generator and is used for NO_2 checks via gas phase titration and to calibrate the ozone analyzer (Figure 2).

The gas mixture produced by this method was introduced into a glass manifold to produce a demand system operating at atmospheric pressure. Before the arrival of the manifold, a simple Tee was used, this was found to effect the sample flow rate of some instruments at the lower concentration levels. A complex Tee arrangement removed this bias, no data was taken under the simple Tee arrangement. The RAMS stations internal calibration manifold was used when auditing stations. In all cases, the inter-connections were made with Teflon tubing and fittings.

3.4 PREPARATION OF CALIBRATION CURVES

The same procedure was used on all instruments to prepare calibration curves. This procedure consisted of generating various gas concentrations and allowing the instrument to sample them until the instrument equilibrated before a reading was taken. This reading along with the known concentration was tabulated and a least squares linear fit was used to find an intercept and a slope. All instruments were done this way including the NO-NO_x instrument used in finding the ozone concentration. This procedure assumes all of the flows are correct as well as the concentration of the tank gases and the zero air was pure. The dilution or zero air used was bottled zero air for areas other than RAMS stations with the station

Dynamic Calibration System



A1706A1

FIGURE 2

zero air used in the station except for CO and HC.

3.5 AUDIT SITES

The main thrust of the quality assurance audits were to check analyzers of RAPS field investigations and RAMS for accuracy. Several RAMS stations were checked. In particular, stations 113 and 122 were selected because unusually high ozone concentrations were consistently obtained there.

The data from each of the sites are presented in a standard format in Tables 1-22. For each case the column "data reading" contains the observed instrument response in volts; "Calibrator Value" is the known concentration (ppm) which was supplied to the instrument; "Station Value" is the concentration (ppm) calculated from the voltage response and the calibration constants in use for that instrument at that time.

A summary of results is presented in Table 23. The percent error was calculated by subtracting the calibration value from the station value (at the highest point), dividing by the calibrator value and multiplying by 100. The significance of the table is in that it demonstrates a rough confidence factor in measurements of this type.

3.5.1 RAPS Helicopters

The RAPS helicopter site was audited three times. There was only enough time to audit one of the three helicopters each visit. The first visit was not successful due to malfunction of the calibration instruments. The Bendix calibrator has several pull-to-test gas line diverters for calibration purposes, one of which had worked loose and shut off a main capillary flow. With this flow shut off, it was impossible to calibrate all instruments. The results are in Tables 1-3.

3.5.2 Meteorology Research Incorporated Plane (MRI)

The MRI plane was audited twice in the Thunderbird hangar at Spirit of St. Louis Air Field. The temperature of the permeation bath was not performing correctly during the first audit. The temperature of the Bendix permeation oven is adjusted by a multi turn pot. In the process of moving, a wire was dislodged from this pot resulting in a much higher temperature than indicated by the setting of the pot. This wire was repaired and the setting was returned to 25°C.

There was some question of this setting by MRI on the second visit since their value did not agree to the standard. The system was rechecked in station 120 and the setting was still found to be 25°C with a thermometer traceable to NBS. The results are in Tables 4-6. (NO, SO₂ data for first visit).

3.5.3 Aerosol Trailer

The aerosol trailer located west of Glasgow, Illinois was visited twice. The first time the aerosol trailer had some grounding problems on the inputs which affected many of the readings, in particular the O₃ reading gave a very poor linear fit. Since there wasn't any way to decide which values were valid, all values were reported. The results are in Tables 7-9.

3.5.4 Research Triangle Institute Van

The 6800 on the RTI was audited once for CO values. The values are in Table 10.

3.5.5 Environmental Measurements Incorporated Van (EMI)

The EMI van associated with the MRI plane was audited during the second visit to the plane. The van was located outside the Thunderbird hangar and the calibration took place outdoors since the air conditioner of the mobile laboratory was not operational. The sulfur reading was high; however, upon checking the Bendix System later in Station 120 the temperature was still found to be 25°C. During the audit of the van the indicator for heating or cooling was steady on the cooling side; this may have represented an inability of the cooling unit to keep the permtube at the proper temperatures.

The EMI calibrator was checked and found to give proper readings; however, the van was not available for further checking. The results are in Tables 11-12.

3.5.6 RAMS Stations

The stations checked were 120, 105, 113 and 122. The permeation tubes were found to be off in 105, 113 and 120. In order to check the station perm-tube with a NBS standard, the NBS tube had to be placed in the station approximately 24 hours before the audit to equilibrate. Due to the large distance and time factor for station 122, this could not be done in time to check the station tube. Station #113 and 122 were checked mainly for O₃ values. The values are

found in Tables 13-19.

3.5.7 Lou Chaney's O_3 and CO Monitors

Two portable O_3 monitors were used in a series of experiments. These were compared twice with O_3 analyzers at RAMS station 105. A CO monitor was also used in this experiment. Several bag samples as well as dynamic diluted samples were tested on this instrument. The results of the first audit are presented in the Tables. During the second audit, there appeared to be a problem with the 6800 in station 105 which gave conflicting results. This was later found to be caused by the automatic range switching program. A test in the G.C. laboratory cleared this up. The O_3 monitors during the second audit were off by as much as 30%. Questions arose as to the reliability of the NO NBS tank, Central's NBS tank and the working standard NO tank. A Bendix NO- NO_x instrument was connected to the Bendix dynamic calibrator and the various tanks were then compared by means of the NO- NO_x instrument. A difference in the working standard tank was found from the original calibration, then a difference between the two NBS tanks was found. Upon further retesting the values were not found to be repeatable since the NO- NO_x instrument was drifting. Readings taken a day apart or even morning versus afternoon gave a different value. The NO- NO_x instrument had been on previously without the vacuum but with oxygen connected. In the past this condition has been found to give erroneous results on some instruments for as long as 2-3 days after the vacuum has been reconnected. This may have been the problem; however, it was not resolved in the time frame available. This is not believed to have affected the readings of the Tables. The results are in Tables 20-22.

3.5.8 Environmental Quality Research (EQR) Theta Sensor

As a service to EQR a Theta sulfur sensor was calibrated using the stations calibration system with a NBS permeation tube. EQR inadvertently readjusted the settings and the instrument was recalibrated in a similar fashion.

3.5.9 Battelle Northwest DC-3

Battelle Northwest employed a portable sulfur gas chromatograph (AID)

(model #513) for the Metromex study. As a courtesy the gas chromatograph was calibrated for SO₂ on a visit to the Battelle plant at the Alton City Airport.

3.5.10 Gas Chromatography Laboratory

The Gas Chromatography Laboratory was not audited. However, cylinder no. 1749 (Section 3.2.3) with its certified analysis, was used as the daily calibration standard for the Beckman 6800 gas chromatograph.

3.6 RAW DATA

The raw data was placed in two laboratory notebooks with carbon copies. They were turned over to the EPA Task Coordinator.

4.0 PROBLEMS AND RECOMMENDATIONS

Most of the problems encountered during the audits were caused by lack of feedback from the calibrator. There was no flow indicator or temperature indicator. A pressure was set on the regulator and it was assumed the flow was correct.

The lack of a temperature indicator required the entire permeation oven to be taken apart in order to place a thermometer in the cavity to register the temperature. The permeation tube in the meantime would change temperature and require a settling time on the order of hours before it would register correctly. A higher setting of the temperature to 30°C would give a temperature which would be easier to maintain in a high temperature ambient environment such as was encountered in the study.

The calibrator did not have a calibration manifold attached to it, this required an additional manifold to be carried each time there was a non-RAMS station field use. The problem from a lack of a calibrator manifold was encountered early in the program when high dilution flows caused a pressure build-up in the sampling line which caused the instrument sample flow to increase.

A portable calibrator with direct reading mass flow meters, or even rotameters along with an actual temperature indicator instead of a null detector, would solve many of these problems.

TABLE 1

QUALITY ASSURANCE AUDIT

NO			RAPS Helicopter #1 NO _x			Date: July 22, 1975 CO		
DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE
1.88	.800	.94	1.86	.800	.93	.03	0 ^a	.3
1.06	.450	.53	1.05	.450	.525	-.10	Instrument	-1.0
.0002	0	0	.002	0	.001	1.633	Zero 21.2	16.33
.353	.170	.176	.345	.170	.173	.391	5.24	3.91
1.85	.800 ^a	.93	1.84	.800 ^a	.92	.89	9.82	8.9
a. Replicate at end of audit			a. Replicate at end of audit			a. Matheson zero air- uncertified purity		

Convertor Efficiency _____

A = -.0100

B = 1.1873

r² = .9993

A = -.0097

B = 1.1744

r² = .9991

A = -.0579

B = .7938

r² = .9862

DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE

Station Value, ppm = A + B (calibrator value, ppm)

TABLE 2

QUALITY ASSURANCE REPORT

RAPS Helicopter #3

Date: July 26, 1975

NO

NO_xO₃

DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE
1.578	.790	.789	1.588	.790	.794	1.737	.198	.174
1.010	.505	.505	1.003	.505	.502	1.542	.168	.154
.657	.340	.329	.661	.340	.331	.828	.094	.083
.324	.170	.162	.324	.170	.162	.587	.061	.059
.000	0	.000	.006	0	.003	.337	.031	.034
						.007	0	.001

Convertor Efficiency 98.5%A = -.0053A = -.0044A = .0039B = 1.004B = 1.0052B = .8726r² = .9997r² = .9996r² = .9979SO₂

CO

DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE
1.073V	.997	1.073	.509	5.20	5.09			
.794	.832	.794	.060	Instrument Zero	.60			
.462	.399	.462	.048	0	.48			
.115	.096	.115						
.002	0	.002						

A = .0158A = .4800A = B = 1.0186B = .8865B = r² = .9898r² = r² =

Station Value, ppm = A + B (calibrator value, ppm)

TABLE 3
QUALITY ASSURANCE REPORT
RAPS Helicopter #3

Date: August 13, 1975
NO_x

CO			NO			NO _x		
DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE
.0135	0 ^a	.135	.000	0	.000	.005	0	.003
.010	Instrument Zero	.10	.370	.193	.185	.362	.193	.181
.444	6.43	4.44	.809	.423	.405	.802	.423	.401
.399	2.97	3.99	1.140	.595	.570	1.108	.595	.554
.338	2.01	.3.38	1.529	.812	.765	1.505	.812	.753
.260	.870	2.6						
.207	.406	2.07						
.170	0 ^b	1.70						
a. At beginning of audit								
b. At end of audit								

Converter Efficiency 98.6%

A = 1.6519

B = .5322

r² = .6953

A = .0026

B = .9452

r² = .9998

A = .0042

B = .9248

r² = .9999

O ₃			SO ₂					
DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE
3.15	.390	.315	.0006	0	.000			
2.245	.251	.225	.89	.874	.890			
1.76	.203	.176	.625	.581	.625			
1.246	.131	.125	.338	.297	.338			
1.611	.176	.161	.167	.150	.167			
.017	0	.002	.075	.066	.075			
			.400	.404	.400			

A = .0133

B = .8029

r² = .9918

A = .0108

B = 1.0201

r² = .9969

Station Value, ppm = A + B (calibrator value, ppm)

TABLE 4
QUALITY ASSURANCE AUDIT
MRI Plane

Date: July 19, 1975

NO(.5ppm range)			NO(1ppm range)			NO(2ppm range)		
DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE
5	0	.005	2	0	.004	219	.840	.870
427	.390	.427	437	.840	.874	170	.670	.680
311	.280	.311	345	.670	.690	105	.390	.420
211	.190	.211	211	.390	.420	076	.280	.300
125	.110	.125	152	.280	.300	050	.190	.210
			104	.190	.208	0	0	.000
			61	.110	.120			

Convertor Efficiency 98%

A = .0055

B = 1.0840

r² = .9999

A = .0097

B = 1.0278

r² = .9996

A = .0106

B = .1020

r² = .9990

NO _x (.5 ppm range)			NO _x (1 ppm range)			NO _x (2 ppm range)		
DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE
20	0	.020	12	0	.024	8	0	.032
499	.390	.449	468	.840	.940	235	.840	.940
325	.280	.325	369	.670	.730	184	.670	.740
230	.190	.23	224	.390	.450	113	.390	.450
139	.110	.139	164	.280	.330	083	.280	.330
			111	.190	.220	056	.190	.220
			68	.110	.140			

A = .0194

B = 1.0990

r² = .9999

A = .0217

B = 1.0815

r² = .9992

A = .0253

B = 1.0802

r² = .9995

Station Value, ppm = A + B (calibrator value, ppm)

TABLE 5
QUALITY ASSURANCE AUDIT
MRI Plane

Date: July 19, 1975

O_3			NO (.2 ppm range)			NO (5 ppm range)		
DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE
260	.214	.260	312	.110	.156	177	1.71	1.77
320	.260	.320	528	.190	.264	129	1.23	1.29
124	.100	.124				85	.84	.85
91	.070	.091				66	.67	.66
0	0	.000				0	0	.00

Convertor Efficiency _____

A = .0022

A = .0075

A = -.0149

B = 1.2175

B = 1.3500

B = 1.0437

r² = .9996

r² = _____

r² = .9992

NO _x (.2 ppm range)			NO _x (5 ppm range)					
DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE
335	.110	.168	2	0	.02			
528	.190	.28	190	1.71	1.90			
			140	1.23	1.40			
			94	.84	.94			
			74	.67	.72			

A = .0140

A = .0079

B = 1.4000

B = 1.1102

r² = _____

r² = .9991

Station Value, ppm = A + B (calibrator value, ppm)

TABLE 6
QUALITY ASSURANCE AUDIT
MRI Plane

Date: July 31, 1975

NO (0.5 ppm range)			NO _x (0.5 ppm range)			SO ₂		
DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE
386	.390	.386	369	.390	.369	617	.802	.617
287	.280	.287	275	.280	.275	462	.602	.462
189	.190	.189	182	.190	.182	403	.530	.403
115	.110	.115	115	.110	.115	305	.402	.305
15	0	.015	20	0	.020	150	.199	.150
						000	0	.000

Convertor Efficiency 97.3%

A = .0119

B = .9613

r² = .9990

A = .01709

B = .9026

r² = .9989

A = -.0023

B = .7695

r² = .9999

O₃

DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE
307	.315	.307						
197	.205	.197						
147	.142	.147						
73	.068	.073						
3	0	.003						

A = .0060

B = .9551

r² = .9988

Station Value, ppm = A + B (calibrator value, ppm)

TABLE 7
QUALITY ASSURANCE AUDIT
Aerosol Trailer
SO₂

Date: July 28, 1975
CO

O₃

DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE
.8404	.091	.120	-.0001	0	.000	1.453	3.82	5.23
.4445	.056	.064	.089	.167	.157	.749	1.95	2.68
.1320	.012	.019	.055	.097	.096	.232	.45	.81
.4480	.056	.064	.042	.066	.073	7.425	21.2	26.8
.001	0	.000				2.020	5.24	7.28
.452	.030	.065 ^b				.139	0	.47
a. Using calibration value of 8/3								
b. Outlier point not used								

Converter Efficiency 100%

A = -.0005

B = 1.2537

r² = .9855

A = .0046

B = .9324

r² = .9937

A = .4212

B = 1.2475

r² = .9997

DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE

Station Value, ppm = A + B (calibrator value, ppm)

TABLE 8

QUALITY ASSURANCE AUDIT
Aerosol Trailer
NO₂

Date: July 28, 1975
NO_x

NO								
DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE
.079	0	.001				.005	0	0
4.622	.057	.046				4.642	.057	.046
2.308	.027	.023				2.259	.027	.023

Convertor Efficiency 125.8%

A = .0012

B = .7891

r² = .9997

A =

B =

r² =

A = .0004

B = .8063

r² = .9991

THC

CH₄

CO

DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE
.013	0	.06	.377	0	.69	.139	0	.47
.655	2.01	2.37	3.071	2.01	2.17	2.020	5.24	7.28

A = .0600

B = 1.1493

r² =

A = .690

B = .7363

r² =

A = .4700

B = 1.3000

r² =

Station Value, ppm = A + B (calibrator value, ppm)

TABLE 5
QUALITY ASSURANCE AUDIT
Aerosol Trailer

Date: August 13, 1975
SO₂

NO			NO _x			SO ₂		
DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE
-.3288	0	-.003	-.304	0	-.003	-.0006	0	.000
2.042	.0199	.020	2.078	.0199	.021	.093	.0914	.164
7.909	.0747	.079	8.054	.0747	.081	.198	.193	.349
10.112	.093	.101	10.384	.093	.104	.354	.427	.626
5.175	.0505	.052	5.241	.0505	.052	.085	.0914	.149
8.917	.081	.089	8.920	.081	.089	.132	.144	.233

Convertor Efficiency 98.9%

A = -.0030

B = 1.1161

r² = .9992

A = -.0030

B = 1.1352

r² = .9989

A = .0236

B = 1.4571

r² = .9870

O ₃			CH ₄			CO		
DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE
.065	.011	.009	.428	0	.48	1.794	5.20	12.74
.443	.062	.064	3.381	2.04	2.11	.092	0	.147
.695	.097	.101				1.609	4.76	11.37
1.085	.1566	.157						
.033	0	.005				.316	1.11	1.804
						.577	2.10	3.74
						.167	.55	.702
				THC				
			.073	0	-.02			
			.665	2.04	2.52			

A = .0021 A = .4800 A = -.0200 A = -.0651

B = .9963 B = .7990 B = 1.2451 B = 2.5078

r² = .9981 r² = r² = r² = .9894

Station Value, ppm = A + B (calibrator value, ppm)

TABLE 10
QUALITY ASSURANCE AUDIT
R.T.I. Trailer
CH₄

Date: August 4, 1975
THC

CO								
DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE
3.252	9.15	9.27	.798	2.04	1.52	1.035	2.04	1.91
2.217	6.23	6.30	.03	0	.04	.093	0	.10
1.426	4.06	4.03						
.923	2.46	2.58						
1.767	4.65	5.01						
3.365	9.82	9.6						
.079	0	.17						

Convertor Efficiency _____

A = .2135

B = .9751

r² = .9979

A = .0400

B = .7255

r² = _____

A = .1000

B = .8873

r² = _____

DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE

Station Value, ppm = A + B (calibrator value, ppm)

Date: July 31, 1975

$$r^2 = \underline{\underline{.9997}}$$

-23-

Date: August 7, 1975

Flow

DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE
2.2631	.312	.321*	800cc/min		794cc/min			
4.6557	.624	.662	600		590			
1.1865	.0347	.032			588			
2.3144	.0694	.064	400		367			
3.2812	.0986	.086			374			
			200		202			
			500		480			
			300		284			
* With pure air dilution								

Converter Efficiency _____

A = -.0124

$$B = 1.0774$$
$$r^2 = .9996$$
$$A = -13.2$$

B = .9997

$$r^2 = .9975$$

DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE

$$\text{Station Value, ppm} = A + B (\text{calibrator value, ppm})$$

TABLE 13

QUALITY ASSURANCE AUDIT
Station #120

Date: July 24, 1975

NO (.5 ppm range)

NO (1 ppm range)

NO_x (.5 ppm range)

DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE
2.0922	.17	.160	1.0375	.170	.159	1.7211	.170	.157
4.2846	.340	.332	2.1313	.340	.330	3.0376	.340	.337
			3.1860	.505	.495			
.8764	.070	.065	.4272	.070	.063	.7519	.070	.066
.0146	0	-.002	.0024	0	-.003	.039	0	.004

Convertor Efficiency _____

A = -.0037A = -.0028A = -.0136B = .9825B = .9810B = .9818r² = .9997r² = .9996r² = .9980NO_x (1 ppm range)O₃

DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE
.8618	.170	.157	2.7929	.180	.181			
1.7675	.340	.327						
2.6720	.505	.498	4.9633	.323	.323			
4.3041	.790	.803	1.5747	.099	.101			
.3588	.070	.063	1.0107	.061	.064			
.0468	0	.000	.0390	0	.001			

A = -.0102A = .0022B = 1.0182B = .9941r² = .9993r² = .9999

Station Value, ppm = A + B (calibrator value, ppm)

TABLE 14
QUALITY ASSURANCE AUDIT
Station #120

Date: July 24, 1975

CO			CH ₄			THC		
DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE
2.1533	5.20	4.58	.8886	2.03	1.83	.9912	2.03	1.95
2.1923	5.17	4.66	2.1801	5.23	4.60	2.2363	5.23	4.68
1.0937	2.56	2.32	1.1450	2.59	2.38	1.3134	2.59	2.66
.5078	1.19	1.079	.5737	1.20	1.15	.7958	1.20	1.52
.2832	.634	.601	.3564	.64	.688	.5859	.64	1.06
.1611	.314	.342	.2563	.318	.472	.4687	.318	.804
0.000	0	-.001	.0244	0	-.025	.1342	0	.069

Convertor Efficiency _____

A = .0345

B = .8853

r² = .9996

A = .1031

B = .8639

r² = .9980

A = .3958

B = .8305

r² = .9857

DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE

Station Value, ppm = A + B (calibrator value, ppm)

Date: July 29, 1975

 SO_2

DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE
2.3364	.0961	.121	2.5073	.0961	.126			
3.5107	.150	.182	3.8183	.150	.192			

Converter Efficiency _____

A = .0122

A = .0083

B = 1.1317

$$B = \underline{1.2245}$$
$$r^2 = \underline{\hspace{2cm}}$$
$$r^2 = \underline{\hspace{2cm}}$$

DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE

$$\text{Station Value, ppm} = A + B (\text{calibrator value, ppm})$$

TABLE 16
QUALITY ASSURANCE AUDIT
Station #113

Date: August 8, 1975

NO			NO _x			O ₃		
DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE
-.0488	0	-.008	-.0073	0	.000	2.299	.245	.226
1.6381	.193	.161	1.7480	.193	.176	3.7597	.412	.376
2.7612	.313	.275	2.8540	.313	.287	2.8002	.302	.276
3.8281	.433	.383	3.9746	.433	.399	1.5014	.158	.144
4.3505	.486	.435	4.8239	.486	.454	.00439	0	-.004

Convertor Efficiency 86.6%

A = -.0106

B = .9117

r² = .9997

A = -.0020

B = .9304

r² = .9998

A = -.0026

B = .9229

r² = .9999

DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE

Station Value, ppm = A + B (calibrator value, ppm)

Date: August 8, 1975

DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE
.0976	0	-.002						
2.1972	.082	.097						
2.0166	.075	.088						
2.7690	.103	.124						
4.1113	.155	.187						
1.6064	.060	.069						

$$r^2 = .9998$$

DATA READING	CALIBRATOR VALUE	STATION VALUE

DATA READING	CALIBRATOR VALUE	STATION VALUE

DATA READING	CALIBRATOR VALUE	STATION VALUE

-29-

TABLE 18
QUALITY ASSURANCE AUDIT
Station #120

Date: August 11, 1975

NO			NO _x			O ₃		
DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE
4.5214	.358	.361	3.8452	.358	.370	2.1212	.144	.135
2.3242	.188	.184	2.0090	.188	.191	3.914	.260	.251
3.3056	.265	.263	2.8564	.265	.274	4.807	.308	.309
						1.6223	.097	.103

Convertor Efficiency 101.7%

A = -.0122
B = 1.0416
r² = .9999

A = -.0092
B = 1.0469
r² = .9961

A = .0009
B = .9820
r² = .9943

SO ₂ (Sta. Perm tube)			SO ₂ (NBS Perm tube)			Bendix SO ₂ Calibrator (EPA)		
DATA READING	CALIBRATOR VALUE*	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE *
3.18942	.1361	.1565	2.2765	.097	.113	1.7871	.072	.076
1.52194	.0654	.0745	1.8652	.080	.092	2.6464	.110	.113
.8911	.0387	.0434	1.4550	.063	.072	4.62155	.194	.197
			4.1931	.179	.208	1.6699	.067	.072
						4.2554	.176	.180

A = -.0015
B = 1.1609
r² = 1.0000

A = .0136
B = .9762
r² = .9922

A = .0048
B = .9918
r² = .9999

Station Value, ppm = A + B (calibrator value, ppm)

*Station system calibrated with NBS perm tube.

TABLE 19
QUALITY ASSURANCE AUDIT
Station #122

Date: August 11, 1975

NO			NO _x			CO		
DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE
.0128	0	-.003	.0494	0	-.001	.1855	.331	.369
2.0422	.193	.188	1.7782	.193	.188	.8544	1.74	1.78
3.7866	.344	.353	3.4716	.381	.375	2.2265	4.74	4.48
4.6850	.437	.437	3.3129	.344	.357	-.0073	0	-.024
4.0014	.381	.373	4.0698	.437	.440			

Convertor Efficiency 99%

A = -.0035

B = 1.0079

r² = .9987

A = -.0024

B = 1.0116

r² = .9983

A = .0456

B = .9429

r² = .9988

O₃

DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE
.0267	0	.001						
2.0434	.229	.235						
3.2209	.350	.371						
3.9656	.427	.457						
1.6210	.165	.186						

A = .0020

B = 1.0589

r² = .9985

Station Value, ppm = A + B (calibrator value, ppm)

TABLE 20
 QUALITY ASSURANCE AUDIT
 Station #105 NO-NO_x
 Lou Chaney O₃ Monitors
 NO_x

Date: August 6, 1975

NO			NO _x			O ₃		
DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE	#1	CALIBRATOR VALUE	#2
3.1738	.424	.374	3.1054	.424	.389	.052	.047	.047
2.5146	.313	.291	2.4829	.313	.310	.022	.019	.021
1.4282	.193	.158	1.3769	.193	.169	.020	.0165	.018
.8276	.101	.084	.7885	.101	.094	.126	.115	.115
						.160	.142	.137
						.000	0	.000

Convertor Efficiency _____

A = <u>-.0102</u>	A = <u>-.0028</u>	A = <u>.0005</u>	A = <u>.0016</u>
B = <u>.9209</u>	B = <u>.9439</u>	B = <u>1.1104</u>	B = <u>.9670</u>
r ² = <u>.9947</u>	r ² = <u>.9910</u>	r ² = <u>.9995</u>	r ² = <u>.9992</u>

DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE

Station Value, ppm = A + B (calibrator value, ppm)

TABLE 22

QUALITY ASSURANCE AUDIT
Station #105
(O₃ Analysis for Lou Chaney)

Date: August 14, 1975O₃

DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE
.9815	.1026	.109						
.5908	.0576	.064						
.4155	.0383	.044						
.2595	.0208 ^a	.026						
a. Produced by change in sleeve setting of ozone lamp. Reproducibility of sleeve position may be questionable.								

Convertor Efficiency _____

A = .0055B = 1.0090r² = .9999

NO

DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE	DATA READING	CALIBRATOR VALUE	STATION VALUE
2.3372	.344	.285						
1.7895	.266	.215						
1.300	.188	.153						

A = -.0074B = .8462r² = .9988

Station Value, ppm = A + B (calibrator value, ppm)

TABLE 23
CALIBRATION DIFFERENCE SUMMARY
(In Percent)

SITE	NO	NO _x	O ₃	CO	CH ₄	THC	SO ₂	TS
RAPS Helicopter 1	17	16		-23				
RAPS Helicopter 3	0	0	-12	-2			7	
RAPS Helicopter 3	-6	-7	-19	-31			2	
MRI Plane 7/19 (.2)	39	47	21					
MRI Plane 7/19 (.5)	9	15						
MRI Plane 7/19 (1.0)	4	12						
MRI Plane 7/19 (2.0)	4	12						
MRI Plane 7/19 (5.0)	4	11						
MRI Plane 7/31	-1	-5	-3				-23	
Aerosol Trailer 7/28	-19	-19	+32	23	8	18	-6	
Aerosol Trailer 8/13	9	12	0	174	3	24	47	
RTI Trailer				1	-25	-6		
EMI Van		-25					53	
EMI Calibrator							6	
Station 120 (.5)	-2	-6	0	-12	-12	-11	28	21
Station 120 (1.0)	-2	2						
Station 113	-10	-7	-9				20	
Station 120 8/11			20					
Station 122	0	1	7	5				
Station 105	-12	-8	6	1				
Chaney	-17		11	-3,1				

TECHNICAL REPORT DATA <i>(Please read Instructions on the reverse before completing)</i>		
1. REPORT NO. EPA-600/4-76-032	2.	3. RECIPIENT'S ACCESSION NO.
4. TITLE AND SUBTITLE REGIONAL AIR POLLUTION STUDY: QUALITY ASSURANCE AUDITS	5. REPORT DATE June 1976	6. PERFORMING ORGANIZATION CODE
	8. PERFORMING ORGANIZATION REPORT NO.	
7. AUTHOR(S) John R. Hribar	10. PROGRAM ELEMENT NO. 1AA003	
9. PERFORMING ORGANIZATION NAME AND ADDRESS Air Monitoring Center Rockwell International 11640 Administration Drive Creve Coeur, MO 63141	11. CONTRACT/GRANT NO. 68-02-1081 Task Order 58	
	13. TYPE OF REPORT AND PERIOD COVERED Final	
12. SPONSORING AGENCY NAME AND ADDRESS Environmental Sciences Research Laboratory Office of Research and Development U.S. Environmental Protection Agency Research Triangle Park, N.C. 27711	14. SPONSORING AGENCY CODE EPA-ORD	
	15. SUPPLEMENTARY NOTES	
16. ABSTRACT <p>Aerometric data is being collected by the Regional Air Pollution Study (RAPS) for use in developing and evaluating air quality simulation models. In addition to the Regional Air Monitoring System (RAMS), data is also collected by research teams in periodic expeditions to the St. Louis study area. Data from all sources are made available for integration through the RAPS Data Bank. A quality assurance audit of instrument systems employed in the Summer 1975 RAPS Expeditionary Research Program was conducted to check the various systems for accuracy. Additionally such checks provide a basis for determining the extent to which data from different instrument systems may be integrated. This report describes the audit equipment and standards used and problems encountered. Quantitative audit results from individual instrument systems are presented. The audits included analyzers for NO, NO_x, O₃, SO₂, CO, CH₄ and THC measurements.</p>		
17. KEY WORDS AND DOCUMENT ANALYSIS		
a. DESCRIPTORS	b. IDENTIFIERS/OPEN ENDED TERMS	c. COSATI Field/Group
*Air Pollution *Measuring instruments *Quality assurance Auditing		13B 14B 14A 05A
19. DISTRIBUTION STATEMENT RELEASE TO PUBLIC	19. SECURITY CLASS (This Report) UNCLASSIFIED	21. NO. OF PAGES 44
	20. SECURITY CLASS (This page) UNCLASSIFIED	22. PRICE

INSTRUCTIONS

- REPORT NUMBER**
Insert the EPA report number as it appears on the cover of the publication.
- LEAVE BLANK**
- RECIPIENTS ACCESSION NUMBER**
Reserved for use by each report recipient.
- TITLE AND SUBTITLE**
Title should indicate clearly and briefly the subject coverage of the report, and be displayed prominently. Set subtitle, if used, in smaller type or otherwise subordinate it to main title. When a report is prepared in more than one volume, repeat the primary title, add volume number and include subtitle for the specific title.
- REPORT DATE**
Each report shall carry a date indicating at least month and year. Indicate the basis on which it was selected (*e.g., date of issue, date of approval, date of preparation, etc.*).
- PERFORMING ORGANIZATION CODE**
Leave blank.
- AUTHOR(S)**
Give name(s) in conventional order (*John R. Doe, J. Robert Doe, etc.*). List author's affiliation if it differs from the performing organization.
- PERFORMING ORGANIZATION REPORT NUMBER**
Insert if performing organization wishes to assign this number.
- PERFORMING ORGANIZATION NAME AND ADDRESS**
Give name, street, city, state, and ZIP code. List no more than two levels of an organizational hierarchy.
- PROGRAM ELEMENT NUMBER**
Use the program element number under which the report was prepared. Subordinate numbers may be included in parentheses.
- CONTRACT/GRANT NUMBER**
Insert contract or grant number under which report was prepared.
- SPONSORING AGENCY NAME AND ADDRESS**
Include ZIP code.
- TYPE OF REPORT AND PERIOD COVERED**
Indicate interim final, etc., and if applicable, dates covered.
- SPONSORING AGENCY CODE**
Leave blank.
- SUPPLEMENTARY NOTES**
Enter information not included elsewhere but useful, such as: Prepared in cooperation with, Translation of, Presented at conference of, To be published in, Supersedes, Supplements, etc.
- ABSTRACT**
Include a brief (*200 words or less*) factual summary of the most significant information contained in the report. If the report contains a significant bibliography or literature survey, mention it here.
- KEY WORDS AND DOCUMENT ANALYSIS**
 - (a) **DESCRIPTORS** - Select from the Thesaurus of Engineering and Scientific Terms the proper authorized terms that identify the major concept of the research and are sufficiently specific and precise to be used as index entries for cataloging.
 - (b) **IDENTIFIERS AND OPEN-ENDED TERMS** - Use identifiers for project names, code names, equipment designators, etc. Use open-ended terms written in descriptor form for those subjects for which no descriptor exists.
 - (c) **COSATI FIELD GROUP** - Field and group assignments are to be taken from the 1965 COSATI Subject Category List. Since the majority of documents are multidisciplinary in nature, the Primary Field/Group assignment(s) will be specific discipline, area of human endeavor, or type of physical object. The application(s) will be cross-referenced with secondary Field/Group assignments that will follow the primary posting(s).
- DISTRIBUTION STATEMENT**
Denote releasability to the public or limitation for reasons other than security for example "Release Unlimited." Cite any availability to the public, with address and price.
- 19. & 20. SECURITY CLASSIFICATION**
DO NOT submit classified reports to the National Technical Information service.
- NUMBER OF PAGES**
Insert the total number of pages, including this one and unnumbered pages, but exclude distribution list, if any.
- PRICE**
Insert the price set by the National Technical Information Service or the Government Printing Office, if known.