

United States
Environmental Protection
Agency

Environmental Monitoring
and Support Laboratory
P.O. Box 15027
Las Vegas NV 89114

EPA-600/4-78-032
June 1978

Research and Development



Environmental Monitoring Series

Environmental Radioactivity Laboratory Intercomparison Studies Program 1978-1979

LIBRARY

U. S. ENVIRONMENTAL PROTECTION AGENCY
P.O. BOX 15027

EP 680/4
73-001b
978-79

RESEARCH REPORTING SERIES

Research reports of the Office of Research and Development, U.S. Environmental Protection Agency, have been grouped into nine series. These nine broad categories were established to facilitate further development and application of environmental technology. Elimination of traditional grouping was consciously planned to foster technology transfer and a maximum interface in related fields. The nine series are:

1. Environmental Health Effects Research
2. Environmental Protection Technology
3. Ecological Research
4. Environmental Monitoring
5. Socioeconomic Environmental Studies
6. Scientific and Technical Assessment Reports (STAR)
7. Interagency Energy-Environment Research and Development
8. "Special" Reports
9. Miscellaneous Reports

This report has been assigned to the ENVIRONMENTAL MONITORING series. This series describes research conducted to develop new or improved methods and instrumentation for the identification and quantification of environmental pollutants at the lowest conceivably significant concentrations. It also includes studies to determine the ambient concentrations of pollutants in the environment and/or the variance of pollutants as a function of time or meteorological factors

EPA-600/4-78-032
June 1978

ENVIRONMENTAL RADIOACTIVITY
LABORATORY INTERCOMPARISON STUDIES PROGRAM
1978-1979

by

Quality Assurance Branch
Monitoring Systems Research and Development Division
Environmental Monitoring and Support Laboratory
Las Vegas, Nevada 89114

U.S. ENVIRONMENTAL PROTECTION AGENCY
OFFICE OF RESEARCH AND DEVELOPMENT
ENVIRONMENTAL MONITORING AND SUPPORT LABORATORY
LAS VEGAS, NEVADA 89114

LIBRARY
U.S. ENVIRONMENTAL PROTECTION AGENCY

DISCLAIMER


This report has been reviewed by the Environmental Monitoring and Support Laboratory-Las Vegas, U.S. Environmental Protection Agency, and approved for publication. Mention of trade names or commercial products does not constitute endorsement or recommendation for use.

FOREWORD

Protection of the environment requires effective regulatory actions which are based on sound technical and scientific information. This information must include the quantitative description and linking of pollutant sources, transport mechanisms, interactions, and resulting effects on man and his environment. Because of the complexities involved, assessment of specific pollutants in the environment requires a total systems approach which transcends the media of air, water, and land. The Environmental Monitoring and Support Laboratory-Las Vegas contributes to the formation and enhancement of a sound monitoring data base for exposure assessment through programs designed to:

- develop and optimize systems and strategies for monitoring pollutants and their impact on the environment
- demonstrate new monitoring systems and technologies by applying them to fulfill special monitoring needs of the Agency's operating programs

This manual has been prepared to assist laboratories involved with environmental radiation measurements in developing and maintaining a quality control program and documenting the precision and accuracy of their data. The intercomparison program described in this brochure should be of value to the EPA, other Federal agencies, State agencies, and private laboratories. For further information on matters related to the quality assurance of environmental radiation measurements contact the Quality Assurance Branch, Environmental Monitoring and Support Laboratory, Las Vegas, Nevada.



George B. Morgan

Director

Environmental Monitoring and Support Laboratory
Las Vegas

ABSTRACT

An intercomparison studies program for laboratories involved in environmental radiation measurements is conducted as part of the radiation quality assurance program of the U.S. Environmental Protection Agency.

This brochure describes the types of environmental samples distributed, the analyses required for each sample, the distribution schedule, and the statistical analysis and reporting of results. Instructions and application forms are included for laboratories desiring to participate in the program.

CONTENTS

	<u>Page</u>
Foreword	iii
Abstract	iv
The Laboratory Intercomparison Studies Program	1
Types of Environmental Samples Distributed	2
Analysis of Data	3
Participation in the Laboratory Intercomparison Studies Program	4
Statistical Calculations	14

Tables

1. Summary of Cross-Check Programs	5
2. Cross-Check Sample Distribution Schedule	6
3. Laboratory Precision: One Standard Deviation Values for Various Analyses	7

Figures

1. Example of a results report form	8
2. Example of a sample analysis and report of participant's data	10
3. Example of a control chart	12

THE LABORATORY INTERCOMPARISON STUDIES PROGRAM

Environmental measurements of radiation are made daily by many Federal, State, local, and private agencies. The data from these measurements are used for a wide variety of purposes including assessment of health effects, the establishment of standards and guides, and for enforcement activities. It is therefore imperative that the precision and accuracy of the data be assured in order that policy decisions concerning environmental quality are based on valid and comparable data of known reliability.

In order to attain this goal, an Agency-wide quality assurance program has been implemented within the U.S. Environmental Protection Agency (EPA). In the area of radiation, quality control responsibilities have been assigned to the Quality Assurance Branch at the EPA's Environmental Monitoring and Support Laboratory-Las Vegas (EMSL-LV) which carries out a program designed to encourage the development and implementation of quality control procedures for sample collection, laboratory analysis, and data handling and reporting.

A major objective of this program is to assist laboratories involved in environmental radiation measurements to develop and maintain both an intra-laboratory and an interlaboratory quality control program. In part, this is accomplished through an extensive laboratory intercomparison study ("cross-check") program involving environmental media (milk, water, air, food, soil, and gases) and a variety of radionuclides with activities at or near environmental levels.

Simulated environmental samples, containing known amounts of one or more radionuclides, are prepared and routinely distributed to laboratories upon request. These laboratories perform the required analyses and return their data to the Quality Assurance Branch for statistical analysis and comparison with known values and analytical values obtained by other participating laboratories. A report and a control chart are returned to each participant. The program thus enables each laboratory to document the precision and accuracy of its radiation data, identify instrument and procedural problems, and to compare its performance with that of other laboratories.

Each laboratory making environmental measurements for radiation should have an internal quality control program in operation to ensure that all instrumentation is calibrated and functioning, and that analytical procedures are being carried out properly. Such a program includes continual monitoring of instrumentation, the plotting of instrument control charts, frequent analysis of replicate samples to check precision, and the regular measurement of samples to which known amounts of activity have been added to check the accuracy of systems.

Participation in a laboratory intercomparison study does not automatically assure the precision and accuracy of a laboratory's data and should not be considered as a substitute for a continuous quality control program within a laboratory. Intercomparison data may be useful for documenting precision and accuracy, and helping to indicate instrument or procedural problems. Participation in intercomparison studies is useful in augmenting a laboratory's quality control program by serving as a check on its internal quality control program.

You are encouraged to have your laboratory participate or expand participation in the cross-check program of the Quality Assurance Branch of the EMSL-LV's Monitoring Systems Research and Development Division. A laboratory may participate in any one or all of the studies described below.

TYPES OF ENVIRONMENTAL SAMPLES DISTRIBUTED

The current laboratory intercomparison studies program covers the analysis of a variety of media containing various levels of radioactivity. These include:

- Water: Water containing several different mixtures of radioactive materials is included in the cross-check program.
- Four-liter samples for the analysis of gross alpha and gross beta activity are sent bimonthly to participating laboratories.
 - Four-liter samples containing chromium-51, zinc-65, cobalt-60, ruthenium-106, cesium-134, and cesium-137 are distributed bimonthly for analysis of gamma emitters.
 - Fifty-milliliter samples for tritium analysis are distributed on a bimonthly basis.
 - Four-liter samples containing plutonium-239 are shipped to laboratories twice a year.
 - Four-liter samples containing radium-226 and radium-228 are distributed four times a year.
 - Four-liter samples containing strontium-89, and strontium-90 are distributed three times a year.
 - Four-liter blind samples containing a mixture of the above radionuclides are distributed semiannually.
- Milk: Four-liter milk samples containing potassium, strontium-89, strontium-90, iodine-131, cesium-137, and barium-140 are distributed on a quarterly basis.

- Air: Two-inch diameter air filters are sent out on a quarterly basis for gross alpha, gross beta, cesium-137, and strontium-90 analyses. Three air filters are sent to each laboratory.
- Soil: Thirty-five-gram soil samples containing either plutonium-239 and plutonium-238, or thorium-228, thorium-230, and thorium-232 are each distributed once a year at 6-month intervals.
- Diet: Four-liter food slurry samples containing strontium-89, strontium-90, iodine-131, cesium-137, barium-140, and potassium are sent to participants three times a year. Two samples are sent to each laboratory.
- Urine: Fifty-milliliter urine samples containing tritium are shipped on a quarterly basis.
- Gas: Ten-liter gas samples (S.T.P.) containing krypton-85 are distributed annually in 1-liter steel cylinders.

Sample size, approximate activity levels, type of analysis, and other pertinent information concerning the cross-check samples are summarized in Table 1. The distribution schedule is outlined in Table 2.

ANALYSIS OF DATA

Each participating laboratory is expected to carry out three independent determinations for each radionuclide included in a particular cross-check study and to report its results on a form (Figure 1) provided with the sample.

Upon receipt of the reports from all participating laboratories, the data are transferred to punch cards for computer analysis. As indicated in the sample calculations, this analysis includes determination of the laboratory standard deviation, calculation of the normalized range, normalized deviation, sample standard deviation, and the grand average of all laboratories. The analytical precision values, used as a basis for judging laboratory performance for specific nuclides, are summarized in Table 3.

A report is generated containing data reported by participating laboratories, listed according to their identity code, along with the results of the analysis (Figure 2). In addition, a control chart is generated and reproduced for each radionuclide included in the sample (Figure 3). The control charts are updated each time a laboratory participates in a particular cross-check study, thus giving each laboratory a continuous record of its performance.

A letter giving the known value for each radionuclide is mailed to participants approximately 2 weeks after the report due date. This is followed by a complete report which includes a copy of the computer printout and control chart.

PARTICIPATION IN THE
LABORATORY INTERCOMPARISON STUDIES PROGRAM

Any laboratory involved in, or concerned with, environmental radiation monitoring and surveillance is eligible to participate in any one or all of the studies described. Moreover, dependent upon personnel available and their workload, a laboratory may elect to receive samples on a less frequent basis than indicated on the distribution schedule (Table 2).

To become a participant in the laboratory intercomparison studies program, complete one of the forms included at the end of this publication and return to:

U.S. Environmental Protection Agency
Environmental Monitoring and Support Laboratory
Quality Assurance Branch
P.O. Box 15027
Las Vegas, NV 89114

Should the laboratory require additional types of cross-check samples at some later date, a second form indicating the samples desired should be submitted. There is no cost to participating laboratories for this service.

TABLE 1. SUMMARY OF CROSS-CHECK PROGRAMS*

SAMPLE	ANALYSIS	ACTIVITY PER ISOTOPE	QUANTITY SUPPLIED	PRESERVATIVE	DISTRIBUTION	TIME FOR ANALYSIS & REPORT
Milk	^{89}Sr , ^{90}Sr , ^{131}I , ^{137}Cs , ^{140}Ba , K	< 200 pCi/liter	~ 4 liters	Formalin	Quarterly	6 weeks
Water						
Gross α , β *	Gross α , β	< 100 pCi/liter	~ 4 liters	0.5N HNO_3	Bimonthly	4 weeks
Gamma	^{60}Co , ^{106}Ru , ^{134}Cs , ^{137}Cs , ^{51}Cr , ^{65}Zn	< 500 pCi/liter	~ 4 liters	0.5N HNO_3	Bimonthly	4 weeks
^3H	^3H	< 3500 pCi/liter	~ 50 ml	none	Bimonthly	4 weeks
^{239}Pu *	^{239}Pu	< 10 pCi/liter	~ 4 liters	0.5N HNO_3	Semiannually	8 weeks
Radium	^{226}Ra , ^{228}Ra	< 20 pCi/liter	~ 4 liters	0.5N HNO_3	Quarterly	8 weeks
Strontium	^{89}Sr , ^{90}Sr	< 50 pCi/liter	~ 4 liters	0.5N HNO_3	Triannually	8 weeks
Blind	Any Combination of Above Radionuclides	< 200 pCi/liter	~ 4 liters	0.5N HNO_3	Semiannually	10 weeks
Air Filter	Gross α , β * ^{90}Sr , ^{137}Cs	< 200 pCi/sample	3 - 2"dia air filters	none	Quarterly	6 weeks
Soil*	^{238}Pu , ^{239}Pu ^{228}Th , ^{230}Th , ^{232}Th	< 50 pCi/sample	~ 35 grams	none	Semiannually	8 weeks
Diet	^{89}Sr , ^{90}Sr , ^{131}I , ^{137}Cs , ^{140}Ba , K	< 200 pCi/kg	2 - 4-liter samples	Formalin	Triannually	8 weeks
Urine	^3H	< 3500 pCi/liter	~ 50 ml	Formalin	Quarterly	4 weeks
Gas	^{85}Kr	< 20 pCi/ml	10 liters	none	Annually	8 weeks

* Laboratories are required to have the necessary licenses before receiving these samples.

TABLE 2. CROSS-CHECK SAMPLE DISTRIBUTION SCHEDULE
(Numbers indicate week of the month)

MONTH	WATER										MILK	FOOD	URINE	AIR FILTER	SOIL	GAS	
	Y	Gross α, β	^3H	^{226}Ra , ^{228}Ra	^{239}Pu	^{89}Sr , ^{90}Sr	Blind*	Sr, γ	Sr, γ	Sr, γ							
Oct	1		2				3			4							
Nov		3									1				$^{2^a}$		
Dec	1		2	3								2		4			
Jan		3			2	1				4							
Feb	1		2														3
Mar		3		2							1		3	4			
Apr	1		2				3			4							
May		3				1									$^{2^b}$		
Jun	1		2	3									2	4			
Jul		3			2					4	1						
Aug	1		2														
Sep		3		2		1							3	4			

- a. Thorium analysis only
- b. Plutonium analysis only
- * Performance sample for the Water Supply Laboratory Certification Program

TABLE 3. LABORATORY PRECISION: ONE STANDARD DEVIATION VALUES FOR VARIOUS ANALYSES

Analysis	Level	One Standard Deviation For Single Determination
Gamma Emitters	5 to 100 pCi/liter or kg >100 pCi/liter or kg	5 pCi/liter 5% of known value
Strontium-89	5 to 100 pCi/liter or kg >100 pCi/liter or kg	5 pCi/liter 5% of known value
Strontium-90	2 to 30 pCi/liter or kg >30 pCi/liter or kg	1.5 pCi/liter 5% of known value
Potassium	>0.1 g/liter or kg	5% of known value
Gross Alpha	<20 pCi/liter >20 pCi/liter	5 pCi/liter 25% of known value
Gross Beta	<100 pCi/liter >100 pCi/liter	5 pCi/liter 5% of known value
Tritium	<4000 pCi/liter >4000 pCi/liter	1s (pCi/liter) = 169.85 x (known) ^{0.933} 10% of known value
Radium-226, Radium-228	> 0.1 pCi/liter	15% of known value
Plutonium, Thorium Isotopes	0.1 pCi/liter, gram or sample	10% of known value
Krypton-85	5 to 100 pCi/ml (S.T.P.)	10% of known value

REPORT SHEET INSTRUCTIONS

1. Legibly type or print numbers and letters in the proper squares. Use only capital letters.
2. Record lab code in columns 1 and 2 for each isotope (one isotope per line). For a code of only one letter, use column 1.
3. Slash the letters Ø and Æ. Do not slash any numbers.
4. Do not use columns 3-7; they are for internal use only.
5. Data entry: (columns 11-40)
 - a. All numbers must be right justified (entered to far right of field).
 - b. For statistical purposes, it is important that each laboratory report three results for each radionuclide. If, for some unavoidable reason, less than three values are reported, record the data starting with the columns reserved for analysis #1 and work to the right.
 - c. Report results to the nearest whole number with the exceptions of tritium and potassium (each recorded to multiples of ten) and radium-226 (recorded to one decimal place). Decimal points are provided on the report forms.
 - d. Each line of the coding form contains the name of the isotope and its unit of activity, preprinted for your convenience. Each line should contain information pertaining to just that one isotope.
 - e. If a value is less than (LT) the minimum detectable activity (MDA), record LT, space, then the MDA value (right justified).

Figure 1 (continued). Results report form

EMSL-LV TRITIUM IN URINE CROSS-CHECK PROGRAM --- SEPTEMBER 1974

09/20/74 SAMPLE - A 3H

KNOWN VALUE = 3273 PCI/L
 EXPECTED LABORATORY PRECISION (1S, 1 DETERMINATION) = 357 PCI/L

LAB	RESULT	EXPERIMENTAL SIGMA	RNG ONLY (\bar{R} + SR)	AVERAGE	NORMALIZED DEVIATION (GRAND AVG)	(KNOWN)
AN	NO DATA PROVIDED					
CF	3269					
CF	3522					
CF	3632	186.1	.60	3474	.9	1.0
CM	3261					
CM	3373					
CM	3362	61.7	.19	3332	.2	.3
CO	NO DATA PROVIDED					
D	3060					
D	3060					
D	3240	103.9	.30	3120	-.8	-.7
J	3255					
J	3247					
J	3294	25.1	.08	3265	-.1	-.0
P	NO DATA PROVIDED					
Z	3240					
Z	3340					
Z	3190	76.4	.25	3257	-.2	-.1

EXPERIMENTAL SIGMA (ALL LABS) = 149 GRAND AVERAGE = 3290

Figure 2. Example of a sample analysis and report of participant's data

Explanation of terms in Figure 2:

- Title: Program name, sample collection date, sample code letter, analysis type, known concentration of radionuclide, expected standard deviation of analysis - single determination.
- Column 1: Laboratory identification code (A, B, C, etc.).
- Column 2: Laboratory results (0-25 results listed down column).
- Column 3: Standard deviation (1s) of the experimental results.
- Columns 4 and 5: Normalized range value in "mean range + standard error of the range" ($\bar{R} + \sigma_R$) units for comparability. (See *Statistical Techniques for Quality Control of Environmental Radioassay*, AQCS Report Stat-1, November 1964, pages 4-8.) ($S_R = \sigma_R$ for printing purposes.)
- Column 6: Average value.
- Column 7: Normalized deviation from the grand average value of all laboratories expressed in σ_m units.
- Column 8: Normalized deviation from the known value expressed in σ_m units.
- Bottom of Chart: Experimental sigma (1s) of all laboratories, and the grand average of all laboratories.

**EMSL—LV Tritium in Urine Crosscheck Program
Lab-D 3H
Normalized Deviation From Known**





Figure 3. Example of a control chart

Explanations of terms used in the control chart (Figure 3).

Title: Name of program, laboratory code letter, and type of analysis.

Upper Graph: "Normalized deviation from known versus the month of analysis." [The 95.0% ($\mu \pm 2\sigma_m$) and the 99.7% ($\mu \pm 3\sigma_m$) confidence levels were chosen as the warning levels and control limits respectively.]

Lower Graph: "Normalized range values ($\bar{R} + \sigma_R$) versus the month of analysis." [The 97.5% ($\bar{R} + 2\sigma_R$) and $\sim 100\%$ ($\bar{R} + 3\sigma_R$) confidence levels were chosen as the warning levels and control limits respectively.]

STATISTICAL CALCULATIONS

To illustrate the computations performed by the computer, example calculations are given using data for three samples analyzed at one laboratory (Laboratory D, see Figure 2).

The experimental data are listed and the mean, experimental sigma and range are computed. These statistics provide measures of the central tendency and dispersion of the data.

The normalized range is computed by first finding the mean range, \bar{R} , the control limit, CL, and the standard error of the range, σ_R . The normalized range measures the dispersion of the data (precision) in such a form that control charts may be used. Control charts allow one to readily compare past analytical performance with present performance. In the example, the normalized range equals 0.3 which is less than 3, which is the upper warning level. The precision of the results is acceptable.

The normalized deviation is calculated by computing the deviation and the standard error of the mean, σ_m . The normalized deviation allows one to readily measure central tendency (accuracy) through the use of control charts. Trends in analytical accuracy can be determined in this manner. For this example, the normalized deviation is -0.7 which falls between +2 and -2, which are the upper and lower warning levels. The accuracy of the data is acceptable.

Finally, the experimental error of all laboratories, the grand average, and the normalized deviation from the grand average are calculated in order to ascertain the performance of all the laboratories as a group. Any bias in methodology or instrumentation may be found from these results.

EXAMPLE CALCULATIONS (Laboratory D Data, see Figure 2)

Experimental data:

Known value = μ = 3273 pCi ³H/liter urine on September 24, 1974

Expected laboratory precision = σ = 357 pCi/liter

<u>Laboratory</u>	<u>Sample</u>	<u>Result</u>
D	x ₁	3060 pCi/liter
D	x ₂	3060 pCi/liter
D	x ₃	3240 pCi/liter

Mean = \bar{x}

$$\bar{x} = \frac{\sum_{i=1}^N x_i}{N} = \frac{9360}{3} = 3120 \text{ pCi/liter}$$

where N = number of results = 3

Experimental sigma = s

$$s = \sqrt{\frac{\sum_{i=1}^N (x_i)^2 - \frac{\left(\sum_{i=1}^N x_i\right)^2}{N}}{N - 1}}$$

$$s = \sqrt{\frac{(3060)^2 + (3060)^2 + (3240)^2 - \frac{(3060 + 3060 + 3240)^2}{3}}{2}}$$

$$s = 103.9 \text{ pCi/liter}$$

Range = r

$$r = | \text{maximum result} - \text{minimum result} |$$

$$r = | 3240 - 3060 | = 180 \text{ pCi/liter}$$

Range Analysis (RNG ANLY)*

$$\text{Mean range} = \bar{R}$$

$$\begin{aligned} \bar{R} &= d_2 \sigma && \text{where } d_2^{**} = 1.693 \text{ for } N = 3 \\ &= (1.693)(357) \end{aligned}$$

$$\bar{R} = 604.4 \text{ pCi/liter}$$

$$\text{Control limit} = \text{CL}$$

$$\begin{aligned} \text{CL} &= \bar{R} + 3\sigma_R \\ &= D_4 \bar{R} && \text{where } D_4^{**} = 2.575 \text{ for } N = 3 \\ &= (2.575)(604.4) \end{aligned}$$

$$\text{CL} = 1556 \text{ pCi/liter}$$

$$\text{Standard error of the range} = \sigma_R$$

$$\begin{aligned} \sigma_R &= (\bar{R} + 3\sigma_R - \bar{R}) \div 3 \\ &= (D_4 \bar{R} - \bar{R}) \div 3 \\ &= (1556 - 604.4) \div 3 \end{aligned}$$

$$\sigma_R = 317.2 \text{ pCi/liter}$$

$$\text{Let range} = r = w\bar{R} + x\sigma_R = 180 \text{ pCi/liter}$$

$$\text{Define normalized range} = w + x$$

$$\text{for } r > \bar{R}, w = 1$$

$$\text{then } r = w\bar{R} + x\sigma_R = \bar{R} + x\sigma_R$$

$$\text{or } x = \frac{r - \bar{R}}{\sigma_R}$$

$$\text{therefore } w + x = 1 + x = 1 + \frac{r - \bar{R}}{\sigma_R}$$

* Rosenstein, M., and A. S. Goldin, "Statistical Techniques for Quality Control of Environmental Radioassay," AQCS Report Stat-1, U.S. Department of Health, Education and Welfare, PHS, Nov 1964

** From table "Factors for Computing Control Limits," Handbook of Tables For Probability and Statistics, 2nd Edition, The Chemical Rubber Co., Cleveland, OH, 1968, p 454.

$$\text{for } r \leq \bar{R}, \quad x = 0$$

$$\text{then } r = w\bar{R} + x\sigma_R = w\bar{R}$$

$$\text{or } w = \frac{r}{\bar{R}}$$

$$\text{therefore } w + x = w + 0 = \frac{r}{\bar{R}}$$

$$\text{since } r < \bar{R}, \quad (180 < 604.4)$$

$$w + x = \frac{180}{604.4}$$

$$w + x = 0.30$$

Normalized deviation of the mean from the known value = ND

Deviation of mean from the known value = D

$$D = \bar{x} - \mu$$

$$= 3120 - 3273$$

$$D = -153 \text{ pCi/liter}$$

Standard error of the mean = σ_m

$$\sigma_m = \frac{\sigma}{\sqrt{N}}$$

$$= \frac{357}{\sqrt{3}}$$

$$\sigma_m = 206.1 \text{ pCi/liter}$$

$$ND = \frac{D}{\sigma_m}$$

$$= \frac{-153}{206.1}$$

$$ND = -0.7$$

Control limit = CL

$$CL = (\mu \pm 3\sigma_m)$$

Warning Limit = WL

$$WL = (\mu \pm 2\sigma_m)$$

Experimental sigma (all laboratories) = s_t (See Figure 2)

$$s_t = \sqrt{\frac{\sum_{i=1}^N (x_i)^2 - \frac{\left(\sum_{i=1}^N x_i\right)^2}{N}}{N - 1}}$$

$$= \sqrt{\frac{162639133 - \frac{(49345)^2}{15}}{14}}$$

$$s_t = 149 \text{ pCi/liter}$$

Grand average = GA

$$GA = \frac{\sum_{i=1}^N x_i}{N}$$

$$= \frac{49345}{15}$$

$$GA = 3290 \text{ pCi/liter}$$

Normalized deviation from the grand average = ND'

Deviation of the mean from the grand average = D'

$$D' = \bar{x} - GA$$

$$= 3120 - 3290$$

$$D' = -170 \text{ pCi/liter}$$

$$ND' = \frac{D'}{\sigma_m}$$

$$= \frac{-170}{206.1}$$

$$ND' = -0.8$$

SUBJECT: Participation in Intercomparison Studies Program

TO: U.S. Environmental Protection Agency
Environmental Monitoring and Support Laboratory
Quality Assurance Branch
P.O. Box 15027
Las Vegas, NV 89114

Please include our laboratory in the cross-check studies we have indicated below. All samples are to be shipped to:

Contact Person _____
Title _____
Laboratory _____
Address _____
NRC License and/or State License Type(s) _____
Number(s) _____

NOTE: When requesting participation in a study containing either nuclear byproducts or special nuclear materials, a copy of the NRC license(s) must accompany the request.

Please indicate desired frequency of participation.

	FREQUENCY DESIRED					FREQUENCY DESIRED			
	BIMONTHLY	QUARTERLY	TRIANNUALLY	SEMIANNUALLY	ANNUALLY	QUARTERLY	TRIANNUALLY	SEMIANNUALLY	ANNUALLY
Milk: Sr, γ		<input type="checkbox"/>		<input type="checkbox"/>	<input type="checkbox"/>			<input type="checkbox"/>	<input type="checkbox"/>
Diet: Sr, γ							<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Water:									
Gross α , β	<input type="checkbox"/>	<input type="checkbox"/>		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		<input type="checkbox"/>	<input type="checkbox"/>
Air Filter:									
Gross α , β	<input type="checkbox"/>	<input type="checkbox"/>		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		<input type="checkbox"/>	<input type="checkbox"/>
^{137}Cs , ^{90}Sr									
^3H	<input type="checkbox"/>	<input type="checkbox"/>		<input type="checkbox"/>	<input type="checkbox"/>				
Soil:									
^{239}Pu				<input type="checkbox"/>	<input type="checkbox"/>				<input type="checkbox"/>
Pu Isotopes				<input type="checkbox"/>	<input type="checkbox"/>				<input type="checkbox"/>
^{226}Ra , ^{228}Ra		<input type="checkbox"/>		<input type="checkbox"/>	<input type="checkbox"/>				<input type="checkbox"/>
Th Isotopes				<input type="checkbox"/>	<input type="checkbox"/>				<input type="checkbox"/>
Blind				<input type="checkbox"/>	<input type="checkbox"/>				<input type="checkbox"/>
Urine: ^3H				<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>		<input type="checkbox"/>	<input type="checkbox"/>
^{89}Sr , ^{90}Sr			<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>				<input type="checkbox"/>
Gases: ^{85}Kr									<input type="checkbox"/>

I certify this laboratory is authorized to receive the samples requested.

Signature _____
Date _____ Title _____

TECHNICAL REPORT DATA
(Please read Instructions on the reverse before completing)

1. REPORT NO. EPA-600/4-78-032		2.	3. RECIPIENT'S ACCESSION NO.	
4. TITLE AND SUBTITLE ENVIRONMENTAL RADIOACTIVITY LABORATORY INTERCOMPARISON STUDIES PROGRAM, 1978-1979			5. REPORT DATE June 1978	
			6. PERFORMING ORGANIZATION CODE	
7. AUTHOR(S) Quality Assurance Branch Monitoring Systems Research and Development Division			8. PERFORMING ORGANIZATION REPORT NO.	
9. PERFORMING ORGANIZATION NAME AND ADDRESS Environmental Monitoring and Support Laboratory Office of Research and Development U.S. Environmental Protection Agency Las Vegas, NV 89114			10. PROGRAM ELEMENT NO. 1HD621	
			11. CONTRACT/GRANT NO.	
12. SPONSORING AGENCY NAME AND ADDRESS U.S. Environmental Protection Agency-Las Vegas, NV Office of Research and Development Environmental Monitoring and Support Laboratory Las Vegas, NV 89114			13. TYPE OF REPORT AND PERIOD COVERED interim 1978-1979	
			14. SPONSORING AGENCY CODE EPA/600/07	
15. SUPPLEMENTARY NOTES				
16. ABSTRACT The U.S. Environmental Protection Agency's intercomparison studies program for laboratories involved in environmental radiation measurements is described. The types of environmental samples distributed, the analyses required for each sample, the distribution schedule, and the statistical analysis and reporting of results are discussed. Instructions and application forms are included for laboratories desiring to participate in the program. This document is not a research report. It is designed for use by laboratories participating or desiring to participate in the quality assurance program.				
17. KEY WORDS AND DOCUMENT ANALYSIS				
a. DESCRIPTORS		b. IDENTIFIERS/OPEN ENDED TERMS		c. COSATI Field/Group
quality assurance quality control quantitative analysis radioactivity statistical quality control				07 B, E 12 A 14 D
18. DISTRIBUTION STATEMENT RELEASE TO PUBLIC		19. SECURITY CLASS (This Report) UNCLASSIFIED		21. NO. OF PAGES 28
		20. SECURITY CLASS (This page) UNCLASSIFIED		22. PRICE