




Gross Alpha-Beta in Water Performance Evaluation Study

A Statistical Evaluation of the January 26, 1996 Data



Gross Alpha-Beta in Water
Performance Evaluation Study
January 26, 1996



Environmental Protection Agency
National Exposure Research Laboratory
Characterization Research Division
Las Vegas, Nevada



UNITED STATES ENVIRONMENTAL PROTECTION AGENCY
OFFICE OF RESEARCH AND DEVELOPMENT
NATIONAL EXPOSURE RESEARCH LABORATORY
CHARACTERIZATION RESEARCH DIVISION-LAS VEGAS
P.O. BOX 93478
LAS VEGAS, NEVADA 89193-3478
(702/798-2100)

Dear Participant,

Enclosed are the results of the Analytical Sciences Branch (CRD-LV) Performance Evaluation Study for *Gross Alpha-Beta in Water; January 26, 1996*.

The known value for each analysis was determined by gravimetric methods, checked by chemical analyses performed by CRD-LV's Radiochemistry Laboratory, and compared to the participating laboratories' grand average.

The expected precision, determined by the known value, was taken from "Table 3. Laboratory Precision: One Standard Deviation Values and Control Limits for Various Analyses", which is based on data accumulated over the years by the Performance Evaluation Program, and can be found in the Environmental Radioactivity Performance Evaluation Studies Program and Radioactive Standards Distribution Program information brochure.

Please take a few minutes to review this report and the analytical data your laboratory submitted to us. If there are any apparent discrepancies, please notify us immediately.

We encourage you to make use of the computer-automated data-entry system that has been in place for some time now. As the number of participants increases, and it becomes unrealistic for us to receive results by mail or FAX, the computer system will be our only avenue for accepting data.

If you have any questions or comments, please send a message via the data-entry system or contact George Dilbeck at 702/798-2104 or Patricia Honsa at 702/798-2141.

Sincerely,

A handwritten signature in black ink that reads "George Dilbeck".

George Dilbeck
Team Leader
RADQA Program

Enclosure

NOTICE

This material has been funded wholly by the U.S. Environmental Protection Agency. It has been subjected to the Agency's review, and has been approved for publication as an EPA document.

The following pages consist of separate sections for each of the nuclides in this study with four parts per section. After the first, each part is separated from the next by a new page or a thick horizontal bar. The first page of each section is a statistical summary for the nuclide and starts with a statement of the known value, the control limits, and the warning limits.

The warning limits are placed at two normalized standard deviations above and below the known value and the control limits are three normalized standard deviations above and below the known value. If you keep control charts, these values will be useful for anticipating problems with the accuracy of your analytical methods.

The coin shaped pie chart at the top of the summary page shows the fate of all the samples sent out in number and percentage terms. The pie chart starts at the top and rotates clockwise. The first sector represents those participants who submitted analytical results within both the warning and control limits. The next sector represents those who are in the warning region but not out of control. The third sector represents those who are out of control, but have passed the outlier test. The fourth sector represents those who have failed the outlier test. The last sector represents those participants who have failed to respond properly. This is the case if no analytical results were returned, or less than three determinations were reported, or if the results were received too late. The reeding on the edge of the coin is spaced at one percent intervals, and the sector shading becomes darker as the data reliability decreases. Sectors with zero width are not shown.

The table in the center shows a number of statistical quantities calculated from the submitted data based on the mean and median values in relation to the known value, both before and after outlier removal. The lower pie chart uses the same construction as the upper chart and shows the distribution of properly submitted data in terms of deviation from the known value divided into sectors representing one, two, three, and greater than three normalized standard deviations.

The second part is an alphabetical listing, in lab-code order, of submitted data and several calculated quantities. An entry that is shaded has been rejected because of one of the reasons listed above or failure of the outlier test. The fifth and sixth columns are a measure of laboratory precision. The Range analysis is a normalized value that you may use to keep precision control charts. If this value is between 2.0 and 3.0, your analytical process precision is in the warning zone; if it exceeds 3.0 it is out of control. The eighth and ninth columns are the differences from the mean of all non-outliers and from the known value, respectively. A tag symbol may appear in the last column. Each page with tags has a symbol definition summary at the bottom. If there is no tag symbol, the data is within the control limits, but it may be in the warning zone.

The third part is a three-column listing of result average, tag symbol, and lab-code in average order excluding those labs not responding properly. In this order, all outliers and out-of-control results appear at the top or bottom of the list.

The last part is two bar chart displays showing frequency distributions of responding participants. The first chart places the known value at the center and a bar at each 0.2 unit of expected precision. The second chart places the mean of the reported measurements at the center and a bar at each 0.2 unit of standard deviation. In both cases, a bar includes those results within 0.1 unit up to the maximum of six. Any results more than six units from the center value are shown cumulatively by a shaded bar one past the sixth unit. If the central tendency of the known value distribution falls away from the center, an error in accuracy is indicated. If the distribution is broad, poor precision is indicated. The mean value distribution is similar but uses the average and standard deviation of reported results as its basis.

The Range Analysis($R + SR$) is calculated from the range, mean range and standard error of the range values. The range is the difference between the maximum and minimum results for the laboratory. The mean range is calculated by multiplying the expected precision by 1.693(for three results). The standard error of the range is calculated by multiplying the mean range by 2.575(for three results), subtracting the mean range from this product, and dividing the result by 3. If the range is greater than the mean range, then the range analysis is calculated by subtracting the mean range from the range, dividing the result by the standard error of the range and adding 1. If the mean range is greater than or equal to the range, then the range analysis is calculated by dividing the range by the mean range.


The normalized deviation of the mean from the grand average is calculated from the deviation of the mean from the grand average and the standard error of the mean values. The deviation of the mean from the grand average is calculated by subtracting the grand average from the average of the laboratory's three results. The standard error of the mean is calculated by dividing the expected precision by the square root of 3(the number of results). The normalized deviation of the mean from the grand average is calculated by dividing the deviation of the mean from the grand average by the standard error of the mean.

The normalized deviation of the mean from the known value is calculated from the deviation of the mean from the known value and the standard error of the mean values. The deviation of the mean from the known value is calculated by subtracting the known value from the average of the laboratory's three results. The standard error of the mean is calculated by dividing the expected precision by the square root of 3(the number of results). The normalized deviation of the mean from the known value is calculated by dividing the deviation of the mean from the known value by the standard error of the mean.


A complete explanation of the statistical calculations involved in the report may be found in the Environmental Radioactivity Performance Evaluation Studies Program information brochure [Draft Revision of EPA-600/4-81-004], available from Patricia Honsa, CRD-LV, 702/798-2141.

Gross Alpha**Statistical Summary****172 Participants**

The known value of this nuclide is **12.1 pCi/l** with an expected precision of **5.0**; the control limits are 3.4 to 20.8; the warning regions are 3.4 to 6.3 and 17.9 to 20.8



Statistic	Respondents	Non-outliers
Mean	12.58	Grand Avg 11.90
Std. Dev.	5.08	2.52
Variance	25.79	6.36
% Coef. of Var.	40.36	21.20
% deviation of mean from known value	3.99	-1.65
Norm. dev. of mean from known value	0.09	-0.08
Median	11.63	11.58
% deviation of median from known value	-3.86	-4.27
Norm. dev. of median from known value	-0.09	-0.20



Gross Alpha

Lab	Res. 1	Res. 2	Res. 3	Exper. Sigma	Rng anal (R + SR)	Average	Normalized deviation (grand-avg) (known) Tag	
A	8.9	10.3	9.2	0.74	0.165	9.47	-0.84	-0.91
AE	11.9	13.4	10.2	1.60	0.378	11.83	-0.02	-0.09
AF	11.3	11.2	10.7	0.32	0.071	11.07	-0.29	-0.36
AH	9.9	8.4	11.0	1.31	0.307	9.77	-0.74	-0.81
AJ	10.3	9.4	11.3	0.95	0.224	10.33	-0.54	-0.61
AK	10.8	11.0	10.7	0.15	0.035	10.83	-0.37	-0.44
AL	18.9	17.9	19.2	0.68	0.154	18.67	2.34	2.27
AP	11.0	11.0	11.5	0.29	0.059	11.17	-0.25	-0.32
AR	11.5	11.1	10.1	0.72	0.165	10.90	-0.35	-0.42
AU	14.2	11.5	15.0	1.83	0.413	13.57	0.58	0.51
AW	12.4	12.7	13.8	0.74	0.165	12.97	0.37	0.30
AZ								
BA	10.3	10.4	10.5	0.10	0.024	10.40	-0.52	-0.59
BB	10.4	8.9	11.6	1.35	0.319	10.30	-0.55	-0.62
BC	12.3	12.0	11.7	0.30	0.071	12.00	0.03	-0.03
BG	11.0	11.0	11.0	0.00	0.000	11.00	-0.31	-0.38
BH	11.6	12.5	12.7	0.59	0.130	12.27	0.13	0.06
BK	13.6	14.9	14.2	0.65	0.154	14.23	0.81	0.74
BL	17.3	15.8	17.6	0.96	0.213	16.90	1.73	1.66
BM	12.2	12.8	13.8	0.81	0.189	12.93	0.36	0.29
BN	6.9	6.9	7.7	0.46	0.095	7.17	-1.64	-1.71
BO	11.4	9.9	11.4	0.87	0.177	10.90	-0.35	-0.42
BS	11.8	14.2	12.7	1.21	0.284	12.90	0.35	0.28
C	7.0	7.3	7.7	0.35	0.083	7.33	-1.58	-1.65
CA	17.7	20.3	20.4	1.53	0.319	19.47	2.62	2.55
CE	9.8	8.3	8.9	0.75	0.177	9.00	-1.00	-1.07
CG	11.9	11.6	11.2	0.35	0.083	11.57	-0.12	-0.18
CJ	20.0	18.0	19.0	1.00	0.236	19.00	2.46	2.39
CK	10.5	10.3	12.1	0.99	0.213	10.97	-0.32	-0.39
CO	11.5	10.1	8.1	1.71	0.402	9.90	-0.69	-0.76
CS	10.3	11.1	12.6	1.17	0.272	11.33	-0.20	-0.27
CX	11.9	14.5	11.7	1.56	0.331	12.70	0.28	0.21
D	12.6	13.7	13.5	0.59	0.130	13.27	0.47	0.40
DB	12.7	11.6	11.0	0.86	0.201	11.77	-0.05	-0.12
DE	9.9	9.9	10.0	0.06	0.012	9.93	-0.68	-0.75
DH	12.3	9.6	10.6	1.37	0.319	10.83	-0.37	-0.44
DI	12.5	10.3	12.2	1.19	0.260	11.67	-0.08	-0.15
DO	10.8	12.9	13.5	1.42	0.319	12.40	0.17	0.10
DS	12.5	12.0	10.3	1.15	0.260	11.60	-0.10	-0.17
DT	10.5	10.7	10.5	0.12	0.024	10.57	-0.46	-0.53
DZ	12.9	11.2	14.8	1.80	0.425	12.97	0.37	0.30
E	11.2	10.3	10.5	0.47	0.106	10.67	-0.43	-0.50
EB	10.6	10.8	13.1	1.39	0.295	11.50	-0.14	-0.21
EO	10.0	12.0	12.0	1.15	0.236	11.33	-0.20	-0.27
ER	13.5	13.3	13.1	0.20	0.047	13.30	0.48	0.42

• = No data submitted

TAG SYMBOLS

↑ = Above control limit

∅ = Insufficient data

× = Determined to be an outlier

↓ = Below control limit

Gross Alpha								
Lab	Res. 1	Res. 2	Res. 3	Exper. Sigma	Rng anal (R + SR)	Average	Normalized deviation (grand-avg) (known) Tag	
EV	11.1	8.5	12.0	1.82	0.413	10.53	-0.47	-0.54
EW								•
FF	7.5	8.6	8.4	0.59	0.130	8.17	-1.29	-1.36
FJ								•
FN	10.0	10.0	11.0	0.58	0.118	10.33	-0.54	-0.61
FZ	7.6	8.6	10.2	1.31	0.307	8.80	-1.07	-1.14
GQ	6.1	5.0	6.0	0.61	0.130	5.70	-2.15	-2.22
GT	17.5	14.9	16.6	1.32	0.307	16.33	1.54	1.47
GZ	16.5	17.7	16.8	0.62	0.142	17.00	1.77	1.70
HI	9.2	10.0	8.7	0.66	0.154	9.30	-0.90	-0.97
HK	15.2	15.1	15.4	0.15	0.035	15.23	1.15	1.09
HL	11.1	11.1	11.1	0.00	0.000	11.10	-0.28	-0.35
HP	10.0	7.0	9.3	1.57	0.354	8.77	-1.09	-1.15
I	10.7	9.4	8.6	1.06	0.248	9.57	-0.81	-0.88
ID	8.4	9.1	8.8	0.35	0.083	8.77	-1.09	-1.15
J	11.3	11.3	11.7	0.23	0.047	11.43	-0.16	-0.23
JG	12.3	12.2	12.5	0.15	0.035	12.33	0.15	0.08
JM	11.7	12.8	11.3	0.78	0.177	11.93	0.01	-0.06
JN	12.8	13.4	14.6	0.92	0.213	13.60	0.59	0.52
JP	12.8	12.6	14.3	0.93	0.201	13.23	0.46	0.39
JQ	12.9	13.2	13.2	0.17	0.035	13.10	0.42	0.35
JS	9.7	10.9	7.9	1.51	0.354	9.50	-0.83	-0.90
JY	11.6	10.8	10.4	0.61	0.142	10.93	-0.33	-0.40
K	12.9	14.0	12.5	0.78	0.177	13.13	0.43	0.36
KE	20.7	26.0	18.7	3.77	0.862	21.80	3.43	3.36 ×
KH	9.8	10.3	10.6	0.40	0.095	10.23	-0.58	-0.65
KT	8.0	8.0	10.0	1.15	0.236	8.67	-1.12	-1.19
KX	12.0	15.0	12.0	1.73	0.354	13.00	0.38	0.31
L	15.4	13.7	13.4	1.08	0.236	14.17	0.79	0.72
LF	15.0	14.0	13.0	1.00	0.236	14.00	0.73	0.66
LR	9.7	10.9	10.4	0.60	0.142	10.33	-0.54	-0.61
LT	11.3	11.4	12.3	0.55	0.118	11.67	-0.08	-0.15
M	15.0	14.0	13.0	1.00	0.236	14.00	0.73	0.66
MF	13.5	13.2	13.8	0.30	0.071	13.50	0.55	0.48
MV	9.2	13.4	13.5	2.45	0.508	12.03	0.05	-0.02
MX	53.1	55.0	60.8	4.01	0.910	56.30	15.38	15.31 ×
N	12.2	13.3	12.8	0.55	0.130	12.77	0.30	0.23
NA	36.0	36.0	34.0	1.15	0.236	35.33	8.12	8.05 ×
NB	17.4	22.0	25.2	3.92	0.921	21.53	3.34	3.27 ×
NJ	11.0	11.8	12.1	0.57	0.130	11.63	-0.09	-0.16
NK	10.3	8.1	9.5	1.11	0.260	9.30	-0.90	-0.97
NO	11.1	11.5	12.3	0.61	0.142	11.63	-0.09	-0.16
NP	11.8	10.9	10.9	0.52	0.106	11.20	-0.24	-0.31
NT	16.6	17.4	16.7	0.44	0.095	16.90	1.73	1.66
O	10.2	10.6	10.7	0.26	0.059	10.50	-0.49	-0.55

• ≡ No data submitted

TAG SYMBOLS

↑ ≡ Above control limit

∅ ≡ Insufficient data

× ≡ Determined to be an outlier

↓ ≡ Below control limit

Gross Alpha

Lab	Res. 1	Res. 2	Res. 3	Exper. Sigma	Rng anal (R + SR)	Average	Normalized deviation (grand-avg) (known)		Tag
OA	12.3	11.3	10.2	1.05	0.248	11.27	-0.22	-0.29	
OB	13.3	12.7	7.6	3.13	0.673	11.20	-0.24	-0.31	
OF	10.2	12.5	13.8	1.82	0.425	12.17	0.09	0.02	
OS	14.0	11.0	11.0	1.73	0.354	12.00	0.03	-0.03	
OX	13.0	14.5	12.2	1.17	0.272	13.23	0.46	0.39	
OY	32.4	34.2	36.4	2.00	0.473	34.33	7.77	7.70	x
P	14.7	14.4	16.7	1.25	0.272	15.27	1.17	1.10	
PB	12.2	11.8	12.9	0.56	0.130	12.30	0.14	0.07	
PG	14.5	13.9	14.3	0.31	0.071	14.23	0.81	0.74	
PM	13.4	13.4	13.4	0.00	0.000	13.40	0.52	0.45	
PQ	9.8	9.5	8.5	0.68	0.154	9.27	-0.91	-0.98	
PR	9.7	9.7	9.4	0.17	0.035	9.60	-0.80	-0.87	
PV	20.0	14.0	13.4	3.65	0.780	15.80	1.35	1.28	
Q	14.4	10.4	13.1	2.04	0.473	12.63	0.25	0.18	
QC	18.4	17.2	17.3	0.67	0.142	17.63	1.99	1.92	
QP	20.0	20.9	18.8	1.05	0.248	19.90	2.77	2.70	
QQ	13.3	10.9	12.6	1.23	0.284	12.27	0.13	0.06	
QT	11.3	10.6	9.9	0.70	0.165	10.60	-0.45	-0.52	
QU	9.0	8.0	9.0	0.58	0.118	8.67	-1.12	-1.19	
QW	9.4	13.7	13.7	2.48	0.508	12.27	0.13	0.06	
QX	10.2	9.9	9.8	0.21	0.047	9.97	-0.67	-0.74	
QZ	8.4	10.3	10.4	1.13	0.236	9.70	-0.76	-0.83	
R	11.6	13.0	12.6	0.72	0.165	12.40	0.17	0.10	
RB	17.2	17.3	18.4	0.67	0.142	17.63	1.99	1.92	
RD	14.7	15.3	14.0	0.65	0.154	14.67	0.96	0.89	
RF	9.1	10.3	11.0	0.96	0.224	10.13	-0.61	-0.68	
RG	10.2	10.0	10.1	0.10	0.024	10.10	-0.62	-0.69	
RI	6.6	11.2	8.9	2.30	0.543	8.90	-1.04	-1.11	
RR	11.0	11.0	11.0	0.00	0.000	11.00	-0.31	-0.38	
RV									•
RZ									•
S	13.8	12.9	12.7	0.59	0.130	13.13	0.43	0.36	
SC	11.7	9.9	11.7	1.04	0.213	11.10	-0.28	-0.35	
SD	11.1	11.1	10.9	0.12	0.024	11.03	-0.30	-0.37	
SF	12.9	11.2	11.0	1.04	0.224	11.70	-0.07	-0.14	
SG	17.0	8.5	12.9	4.25	1.008	12.80	0.31	0.24	
SI	12.5	12.9	13.3	0.40	0.095	12.90	0.35	0.28	
SL	13.0	11.0	14.0	1.53	0.354	12.67	0.27	0.20	
SM	9.7	13.2	10.5	1.83	0.413	11.13	-0.27	-0.33	
SN	10.3	11.6	8.9	1.35	0.319	10.27	-0.57	-0.64	
SO	10.6	9.5	10.6	0.64	0.130	10.23	-0.58	-0.65	
SR	8.4	8.8	8.1	0.35	0.083	8.43	-1.20	-1.27	
SS	11.1	11.3	10.9	0.20	0.047	11.10	-0.28	-0.35	
ST	10.4	9.5	9.9	0.45	0.106	9.93	-0.68	-0.75	
SU	5.5	6.3	6.7	0.61	0.142	6.17	-1.99	-2.06	

• = No data submitted

TAG SYMBOLS

↑ = Above control limit

∅ = Insufficient data

x = Determined to be an outlier

↓ = Below control limit

Gross Alpha

Lab	Res. 1	Res. 2	Res. 3	Exper. Sigma	Rng anal (R + SR)	Average	Normalized deviation (grand-avg) (known)	Tag
SV								•
SX	13.0	10.0	11.0	1.53	0.354	11.33	-0.20	-0.27
SZ	11.6	11.8	12.9	0.70	0.154	12.10	0.07	0.00
T	10.3	10.4	9.4	0.55	0.118	10.03	-0.65	-0.72
TD	18.0	16.5	18.9	1.21	0.284	17.80	2.04	1.97
TG	8.9	12.0	7.6	2.26	0.520	9.50	-0.83	-0.90
TH								•
TI								•
TL	17.9	11.8	10.8	3.84	0.839	13.50	0.55	0.48
TN	12.9	12.1	9.9	1.55	0.354	11.63	-0.09	-0.16
TQ	13.4	12.2	13.2	0.64	0.142	12.93	0.36	0.29
TW	12.7	11.3	9.1	1.81	0.425	11.03	-0.30	-0.37
TY	8.3	8.7	8.9	0.31	0.071	8.63	-1.13	-1.20
U	13.0	12.1	10.8	1.11	0.260	11.97	0.02	-0.05
UA	15.7	19.8	17.8	2.05	0.484	17.77	2.03	1.96
UE	10.9	12.1	11.7	0.61	0.142	11.57	-0.12	-0.18
UN	10.7	10.9	12.0	0.70	0.154	11.20	-0.24	-0.31
UP	13.2	15.1	12.1	1.52	0.354	13.47	0.54	0.47
UQ	13.3	12.0	13.7	0.89	0.201	13.00	0.38	0.31
UW	14.7	16.4	14.5	1.04	0.224	15.20	1.14	1.07
VA	12.2	8.7	10.6	1.75	0.413	10.50	-0.49	-0.55
VC								•
VH	7.8	9.4	8.8	0.81	0.189	8.67	-1.12	-1.19
VI	12.3	13.5	12.0	0.79	0.177	12.60	0.24	0.17
VJ	15.1	14.5	15.4	0.46	0.106	15.00	1.07	1.00
VO	8.3	9.7	6.3	1.71	0.402	8.10	-1.32	-1.39
VT	12.2	12.4	12.3	0.10	0.024	12.30	0.14	0.07
W	12.8	15.1	13.1	1.25	0.272	13.67	0.61	0.54
WE	9.5	9.0	8.5	0.50	0.118	9.00	-1.00	-1.07
WF	15.2	14.7	13.1	1.10	0.248	14.33	0.84	0.77
WH	10.2	12.1	12.0	1.07	0.224	11.43	-0.16	-0.23
WJ								•
WN	14.8	14.9	14.9	0.06	0.012	14.87	1.03	0.96
WO	11.9	14.9	10.7	2.16	0.496	12.50	0.21	0.14
WP								•
WQ	10.8	10.5	11.3	0.40	0.095	10.87	-0.36	-0.43
X	8.6	11.2	10.7	1.38	0.307	10.17	-0.60	-0.67

Data sorted by Laboratory Average

Average	Tag	Lab	Average	Tag	Lab	Average	Tag	Lab
5.70		GQ	8.17		FF	8.67		KT
6.17		SU	8.43		SR	8.77		ID
7.17		BN	8.63		TY	8.77		HP
7.33		C	8.67		VH	8.80		FZ
8.10		VO	8.67		QU	8.90		RI

• = No data submitted

TAG SYMBOLS

↑ = Above control limit

∅ = Insufficient data

× = Determined to be an outlier

↓ = Below control limit

Gross Alpha**Data sorted by Laboratory Average**

Average	Tag	Lab	Average	Tag	Lab	Average	Tag	Lab
9.00		WE	11.03		SD	12.60		VI
9.00		CE	11.07		AF	12.63		Q
9.27		PQ	11.10		SS	12.67		SL
9.30		NK	11.10		SC	12.70		CX
9.30		HI	11.10		HL	12.77		N
9.47		A	11.13		SM	12.80		SG
9.50		TG	11.17		AP	12.90		SI
9.50		JS	11.20		UN	12.90		BS
9.57		I	11.20		OB	12.93		TQ
9.60		PR	11.20		NP	12.93		BM
9.70		QZ	11.27		OA	12.97		DZ
9.77		AH	11.33		SX	12.97		AW
9.90		CO	11.33		EO	13.00		UQ
9.93		ST	11.33		CS	13.00		KX
9.93		DE	11.43		WH	13.10		JQ
9.97		QX	11.43		J	13.13		S
10.03		T	11.50		EB	13.13		K
10.10		RG	11.57		UE	13.23		OX
10.13		RF	11.57		CG	13.23		JP
10.17		X	11.60		DS	13.27		D
10.23		SO	11.63		TN	13.30		ER
10.23		KH	11.63		NO	13.40		PM
10.27		SN	11.63		NJ	13.47		UP
10.30		BB	11.67		LT	13.50		TL
10.33		LR	11.67		DI	13.50		MF
10.33		FN	11.70		SF	13.57		AU
10.33		AJ	11.77		DB	13.60		JN
10.40		BA	11.83		AE	13.67		W
10.50		VA	11.93		JM	14.00		M
10.50		O	11.97		U	14.00		LF
10.53		EV	12.00		OS	14.17		L
10.57		DT	12.00		BC	14.23		PG
10.60		QT	12.03		MV	14.23		BK
10.67		E	12.10		SZ	14.33		WF
10.83		DH	12.17		OF	14.67		RD
10.83		AK	12.27		QW	14.87		WN
10.87		WQ	12.27		QQ	15.00		VJ
10.90		BO	12.27		BH	15.20		UW
10.90		AR	12.30		VT	15.23		HK
10.93		JY	12.30		PB	15.27		P
10.97		CK	12.33		JG	15.80		PV
11.00		RR	12.40		R	16.33		GT
11.00		BG	12.40		DO	16.90		NT
11.03		TW	12.50		WO	16.90		BL

• = No data submitted

TAG SYMBOLS

↑ = Above control limit



∅ = Insufficient data

× = Determined to be an outlier

↓ = Below control limit

Gross Alpha**Data sorted by Laboratory Average**

Average	Tag	Lab	Average	Tag	Lab	Average	Tag	Lab
17.00		GZ	17.80		TD	21.53	×	NB
17.63		RB	18.67		AL	21.80	×	KE
17.63		QC	19.00		CJ	34.33	×	OY
17.77		UA	19.47		CA	35.33	×	NA
			19.90		QP	56.30	×	MX



• ≡ No data submitted

∅ ≡ Insufficient data

TAG SYMBOLS


× ≡ Determined to be an outlier

↑ ≡ Above control limit


↓ ≡ Below control limit

Gross Beta**Statistical Summary****172 Participants**

The known value of this nuclide is **7.0 pCi/l** with an expected precision of **5.0**; the control limits are 0.0 to 15.7; the warning regions are 0.0 to 1.2 and 12.8 to 15.7



Statistic	Respondents	Non-outliers
Mean	9.02	Grand Avg 8.50
Std. Dev.	4.09	1.94
Variance	16.70	3.75
% Coef. of Var.	45.31	22.76
% deviation of mean from known value	28.84	21.49
Norm. dev. of mean from known value	0.49	0.78
Median	8.47	8.47
% deviation of median from known value	20.95	20.95
Norm. dev. of median from known value	0.36	0.76



11 / 16 CRD-LV Performance Evaluation: Gross Alpha-Beta in Water, 26-Jan-1996

Gross Beta

Lab	Res. 1	Res. 2	Res. 3	Exper. Sigma	Rng anal (R + SR)	Average	Normalized deviation (grand-avg) (known)		Tag
A	19.4	19.2	18.9	0.25	0.059	19.17	3.69	4.21	×
AE	8.1	7.6	7.9	0.25	0.059	7.87	-0.22	0.30	
AF	6.7	7.2	5.5	0.87	0.201	6.47	-0.71	-0.18	
AH	9.1	9.1	8.0	0.64	0.130	8.73	0.08	0.60	
AJ	4.8	6.4	5.7	0.80	0.189	5.63	-0.99	-0.47	
AK	9.3	7.9	8.1	0.76	0.165	8.43	-0.02	0.50	
AL	10.6	8.8	10.6	1.04	0.213	10.00	0.52	1.04	
AP	9.0	8.3	8.1	0.47	0.106	8.47	-0.01	0.51	
AR	7.4	7.6	8.0	0.31	0.071	7.67	-0.29	0.23	
AU	4.3	5.3	4.5	0.53	0.118	4.70	-1.32	-0.80	
AW	8.8	10.1	9.9	0.70	0.154	9.60	0.38	0.90	
AZ									
BA	11.1	7.1	8.0	2.10	0.473	8.73	0.08	0.60	
BB	6.0	7.8	8.2	1.17	0.260	7.33	-0.41	0.12	
BC	9.2	9.4	9.2	0.12	0.024	9.27	0.26	0.79	
BG	7.0	7.0	7.0	0.00	0.000	7.00	-0.52	0.00	
BH	8.2	9.7	9.3	0.78	0.177	9.07	0.19	0.72	
BK	7.3	8.3	9.5	1.10	0.260	8.37	-0.05	0.47	
BL	7.4	7.7	7.8	0.21	0.047	7.63	-0.30	0.22	
BM	9.4	8.3	10.5	1.10	0.260	9.40	0.31	0.83	
BN	6.3	6.7	6.4	0.21	0.047	6.47	-0.71	-0.18	
BO	6.3	8.3	5.8	1.32	0.295	6.80	-0.59	-0.07	
BS	9.9	10.0	10.2	0.15	0.035	10.03	0.53	1.05	
C	6.6	6.0	6.0	0.35	0.071	6.20	-0.80	-0.28	
CA	7.1	8.3	8.4	0.72	0.154	7.93	-0.20	0.32	
CE	7.3	6.8	6.7	0.32	0.071	6.93	-0.54	-0.02	
CG	10.5	9.7	10.4	0.44	0.095	10.20	0.59	1.11	
CJ	7.3	6.9	7.2	0.21	0.047	7.13	-0.47	0.05	
CK	10.0	12.0	10.6	1.03	0.236	10.87	0.82	1.34	
CO	5.5	4.1	5.1	0.72	0.165	4.90	-1.25	-0.73	
CS	7.5	7.1	7.8	0.35	0.083	7.47	-0.36	0.16	
CX	12.8	15.6	11.7	2.01	0.461	13.37	1.68	2.21	
D	13.5	13.6	11.2	1.36	0.284	12.77	1.48	2.00	
DB	11.3	12.1	11.1	0.53	0.118	11.50	1.04	1.56	
DE	7.9	7.9	8.0	0.06	0.012	7.93	-0.20	0.32	
DH	10.5	9.9	8.9	0.81	0.189	9.77	0.44	0.96	
DI	5.3	5.2	5.9	0.38	0.083	5.47	-1.05	-0.53	
DO	9.0	7.7	7.3	0.89	0.201	8.00	-0.17	0.35	
DS	8.4	8.9	8.2	0.36	0.083	8.50	0.00	0.52	
DT	8.1	7.4	8.3	0.47	0.106	7.93	-0.20	0.32	
DZ	8.8	9.4	8.9	0.32	0.071	9.03	0.18	0.70	
E	9.6	8.6	8.9	0.51	0.118	9.03	0.18	0.70	
EB	9.0	10.1	11.3	1.15	0.272	10.13	0.56	1.09	
EO	9.0	11.0	16.0	3.61	0.827	12.00	1.21	1.73	
ER									

• = No data submitted

TAG SYMBOLS

↑ = Above control limit

∅ = Insufficient data

× = Determined to be an outlier

↓ = Below control limit

Gross Beta								
Lab	Res. 1	Res. 2	Res. 3	Exper. Sigma	Rng anal (R + SR)	Average	Normalized deviation (grand-avg) (known)	Tag
EV	7.6	8.3	8.6	0.51	0.118	8.17	-0.12	0.40
EW	10.3	8.5	11.9	1.70	0.402	10.23	0.60	1.12
FF	9.2	8.3	8.3	0.52	0.106	8.60	0.03	0.55
FJ								•
FN	9.0	9.0	8.0	0.58	0.118	8.67	0.06	0.58
FZ	8.4	10.4	7.6	1.44	0.331	8.80	0.10	0.62
GQ	7.8	9.0	9.8	1.01	0.236	8.87	0.13	0.65
GT	7.6	9.7	7.2	1.34	0.295	8.17	-0.12	0.40
GZ	4.0	2.3	3.3	0.85	0.201	3.20	-1.84	-1.32
HI	8.1	9.8	8.9	0.85	0.201	8.93	0.15	0.67
HK	10.3	9.5	10.4	0.49	0.106	10.07	0.54	1.06
HL	8.9	8.9	8.9	0.00	0.000	8.90	0.14	0.66
HP	6.4	6.1	8.2	1.14	0.248	6.90	-0.56	-0.03
I	6.7	4.1	6.2	1.38	0.307	5.67	-0.98	-0.46
ID	9.8	8.0	10.0	1.10	0.236	9.27	0.26	0.79
J	7.7	7.3	7.0	0.35	0.083	7.33	-0.41	0.12
JG	9.1	8.3	8.0	0.57	0.130	8.47	-0.01	0.51
JM	7.8	7.5	9.9	1.31	0.284	8.40	-0.04	0.48
JN	8.4	9.0	7.8	0.60	0.142	8.40	-0.04	0.48
JP	6.5	8.4	6.9	1.00	0.224	7.27	-0.43	0.09
JQ	8.2	5.8	4.0	2.11	0.496	6.00	-0.87	-0.35
JS	9.3	10.8	11.3	1.04	0.236	10.47	0.68	1.20
JY	6.0	6.2	7.5	0.81	0.177	6.57	-0.67	-0.15
K	10.3	11.7	10.9	0.70	0.165	10.97	0.85	1.37
KE	8.9	9.9	8.8	0.61	0.130	9.20	0.24	0.76
KH	7.6	9.1	6.3	1.40	0.331	7.67	-0.29	0.23
KT	5.0	4.0	4.0	0.58	0.118	4.33	-1.44	-0.92
KX	10.0	10.0	9.0	0.58	0.118	9.67	0.40	0.92
L	10.4	12.8	11.8	1.21	0.284	11.67	1.10	1.62
LF	7.8	7.8	9.8	1.15	0.236	8.47	-0.01	0.51
LR	10.1	9.2	11.9	1.37	0.319	10.40	0.66	1.18
LT	11.1	10.5	10.9	0.31	0.071	10.83	0.81	1.33
M	8.0	8.0	7.0	0.58	0.118	7.67	-0.29	0.23
MF	7.9	7.8	8.0	0.10	0.024	7.90	-0.21	0.31
MV								•
MX	47.8	45.0	43.3	2.27	0.532	45.37	12.77	13.29 ×
N	7.5	10.4	9.0	1.45	0.343	8.97	0.16	0.68
NA	7.0	12.0	10.0	2.52	0.591	9.67	0.40	0.92
NB	27.5	30.3	33.0	2.75	0.650	30.27	7.54	8.06 ×
NJ	13.6	13.2	12.5	0.56	0.130	13.10	1.59	2.11
NK	11.9	8.9	8.0	2.04	0.461	9.60	0.38	0.90
NO	7.6	7.8	8.3	0.36	0.083	7.90	-0.21	0.31
NP	7.7	8.9	7.8	0.67	0.142	8.13	-0.13	0.39
NT	9.1	8.4	8.7	0.35	0.083	8.73	0.08	0.60
O	6.4	6.6	6.9	0.25	0.059	6.63	-0.65	-0.13

• ≡ No data submitted

TAG SYMBOLS

↑ ≡ Above control limit

⊘ ≡ Insufficient data

× ≡ Determined to be an outlier

↓ ≡ Below control limit

13 / 16 CRD-LV Performance Evaluation: Gross Alpha-Beta in Water, 26-Jan-1996

Gross Beta

Lab	Res. 1	Res. 2	Res. 3	Exper. Sigma	Rng anal (R + SR)	Average	Normalized deviation (grand-avg) (known)		Tag
OA	11.1	12.7	15.5	2.23	0.520	13.10	1.59	2.11	
OB	6.7	8.2	9.4	1.35	0.319	8.10	-0.14	0.38	
OF	4.9	7.1	7.1	1.27	0.260	6.37	-0.74	-0.22	
OS	11.0	10.0	12.0	1.00	0.236	11.00	0.86	1.39	
OX	10.9	10.5	7.6	1.80	0.390	9.67	0.40	0.92	
OY	19.6	21.5	22.0	1.27	0.284	21.03	4.34	4.86	×
P	8.4	7.0	7.8	0.70	0.165	7.73	-0.27	0.25	
PB	8.9	8.8	9.3	0.26	0.059	9.00	0.17	0.69	
PG	9.4	13.3	8.0	2.75	0.626	10.23	0.60	1.12	
PM	11.1	10.9	8.6	1.39	0.295	10.20	0.59	1.11	
PQ	8.6	8.7	8.4	0.15	0.035	8.57	0.02	0.54	
PR	7.8	7.3	6.3	0.76	0.177	7.13	-0.47	0.05	
PV	11.9	12.1	7.6	2.54	0.532	10.53	0.70	1.22	
Q	8.6	10.8	9.9	1.11	0.260	9.77	0.44	0.96	
QC	8.9	10.4	9.8	0.75	0.177	9.70	0.41	0.94	
QP	3.2	4.1	5.8	1.32	0.307	4.37	-1.43	-0.91	
QQ	5.6	6.8	7.9	1.15	0.272	6.77	-0.60	-0.08	
QT	7.7	12.0	7.3	2.61	0.555	9.00	0.17	0.69	
QU	5.9	5.5	6.0	0.26	0.059	5.80	-0.94	-0.42	
QW	6.6	7.4	6.6	0.46	0.095	6.87	-0.57	-0.05	
QX	8.5	8.5	8.4	0.06	0.012	8.47	-0.01	0.51	
QZ	7.3	7.3	7.8	0.29	0.059	7.47	-0.36	0.16	
R	10.6	10.4	9.7	0.47	0.106	10.23	0.60	1.12	
RB	9.8	8.9	10.4	0.75	0.177	9.70	0.41	0.94	
RD	7.5	9.6	9.1	1.10	0.248	8.73	0.08	0.60	
RF	5.3	7.3	6.4	1.00	0.236	6.33	-0.75	-0.23	
RG	9.5	8.2	8.6	0.67	0.154	8.77	0.09	0.61	
RI	8.0	5.6	6.8	1.20	0.284	6.80	-0.59	-0.07	
RR	11.3	10.9	10.7	0.31	0.071	10.97	0.85	1.37	
RV									•
RZ									•
S	8.1	8.2	8.5	0.21	0.047	8.27	-0.08	0.44	
SC	10.2	7.3	7.4	1.65	0.343	8.30	-0.07	0.45	
SD	11.7	8.4	10.8	1.71	0.390	10.30	0.62	1.14	
SF	7.5	7.1	7.8	0.35	0.083	7.47	-0.36	0.16	
SG	6.9	5.7	6.1	0.61	0.142	6.23	-0.79	-0.27	
SI	6.2	5.7	6.8	0.55	0.130	6.23	-0.79	-0.27	
SL	8.8	13.0	8.5	2.52	0.532	10.10	0.55	1.07	
SM	7.8	8.5	8.0	0.36	0.083	8.10	-0.14	0.38	
SN	8.9	11.2	9.3	1.23	0.272	9.80	0.45	0.97	
SO	5.6	5.9	5.7	0.15	0.035	5.73	-0.96	-0.44	
SR	6.5	7.6	5.4	1.10	0.260	6.50	-0.69	-0.17	
SS	9.4	9.3	9.3	0.06	0.012	9.33	0.29	0.81	
ST	7.1	6.8	5.5	0.85	0.189	6.47	-0.71	-0.18	
SU	9.0	6.2	6.8	1.47	0.331	7.33	-0.41	0.12	

• = No data submitted

TAG SYMBOLS

↑ = Above control limit

∅ = Insufficient data

× = Determined to be an outlier

↓ = Below control limit

Gross Beta

Lab	Res. 1	Res. 2	Res. 3	Exper. Sigma	Rng anal (R + SR)	Average	Normalized deviation (grand-avg)	(known) Tag
SV								•
SX	11.1	10.2	9.6	0.75	0.177	10.30	0.62	1.14
SZ	8.4	7.4	8.3	0.55	0.118	8.03	-0.16	0.36
T	6.3	6.4	6.8	0.26	0.059	6.50	-0.69	-0.17
TD	14.6	11.1	8.0	3.30	0.780	11.23	0.95	1.47
TG	5.9	6.9	6.7	0.53	0.118	6.50	-0.69	-0.17
TH								•
TI								•
TL	6.4	7.0	7.2	0.42	0.095	6.87	-0.57	-0.05
TN	10.4	11.4	9.9	0.76	0.177	10.57	0.71	1.24
TQ	5.6	6.4	6.6	0.53	0.118	6.20	-0.80	-0.28
TW	4.7	11.3	8.9	3.34	0.780	8.30	-0.07	0.45
TY	6.8	5.9	6.1	0.47	0.106	6.27	-0.78	-0.25
U	9.9	8.8	9.2	0.56	0.130	9.30	0.28	0.80
UA	9.1	9.5	10.3	0.61	0.142	9.63	0.39	0.91
UE	11.7	12.2	10.2	1.04	0.236	11.37	0.99	1.51
UN	9.2	11.2	11.7	1.32	0.295	10.70	0.76	1.28
UP	7.7	8.3	7.2	0.55	0.130	7.73	-0.27	0.25
UQ	8.2	8.9	9.1	0.47	0.106	8.73	0.08	0.60
UW	16.3	14.7	13.2	1.55	0.366	14.73	2.16	2.68
VA	8.4	8.8	6.9	1.00	0.224	8.03	-0.16	0.36
VC								•
VH	7.3	9.8	7.5	1.39	0.295	8.20	-0.11	0.42
VI	5.1	4.9	4.4	0.36	0.083	4.80	-1.28	-0.76
VJ	7.0	6.5	6.6	0.26	0.059	6.70	-0.62	-0.10
VO	6.5	8.0	9.7	1.60	0.378	8.07	-0.15	0.37
VT	10.2	10.4	10.8	0.31	0.071	10.47	0.68	1.20
W	7.7	6.3	7.2	0.71	0.165	7.07	-0.50	0.02
WE	9.7	9.6	9.3	0.21	0.047	9.53	0.36	0.88
WF	11.5	11.2	10.7	0.40	0.095	11.13	0.91	1.43
WH	4.5	5.9	6.7	1.11	0.260	5.70	-0.97	-0.45
WJ								•
WN								•
WO	14.2	11.4	12.1	1.46	0.331	12.57	1.41	1.93
WP								•
WQ	10.0	9.0	9.5	0.50	0.118	9.50	0.34	0.87
X	7.1	9.9	7.9	1.44	0.331	8.30	-0.07	0.45

Data sorted by Laboratory Average

Average	Tag	Lab	Average	Tag	Lab	Average	Tag	Lab
3.20		GZ	4.90		CO	5.73		SO
4.33		KT	5.47		DI	5.80		QU
4.37		QP	5.63		AJ	6.00		JQ
4.70		AU	5.67		I	6.20		TQ
4.80		VI	5.70		WH	6.20		C

• = No data submitted

TAG SYMBOLS

↑ = Above control limit

∅ = Insufficient data

× = Determined to be an outlier

↓ = Below control limit

15 / 16 CRD-LV Performance Evaluation: Gross Alpha-Beta in Water, 26-Jan-1996
Gross Beta
Data sorted by Laboratory Average

Average	Tag	Lab	Average	Tag	Lab	Average	Tag	Lab
6.23		SI	8.00		DO	9.27		ID
6.23		SG	8.03		VA	9.30		U
6.27		TY	8.03		SZ	9.33		SS
6.33		RF	8.07		VO	9.40		BM
6.37		OF	8.10		OB	9.50		WQ
6.47		ST	8.10		SM	9.53		WE
6.47		BN	8.13		NP	9.60		NK
6.47		AF	8.17		GT	9.60		AW
6.50		TG	8.17		EV	9.63		UA
6.50		T	8.20		VH	9.67		OX
6.50		SR	8.27		S	9.67		NA
6.57		JY	8.30		X	9.67		KX
6.63		O	8.30		TW	9.70		RB
6.70		VJ	8.30		SC	9.70		QC
6.77		QQ	8.37		BK	9.77		DH
6.80		RI	8.40		JN	9.77		Q
6.80		BO	8.40		JM	9.80		SN
6.87		TL	8.43		AK	10.00		AL
6.87		QW	8.47		QX	10.03		BS
6.90		HP	8.47		LF	10.07		HK
6.93		CE	8.47		JG	10.10		SL
7.00		BG	8.47		AP	10.13		EB
7.07		W	8.50		DS	10.20		PM
7.13		CJ	8.57		PQ	10.20		CG
7.13		PR	8.60		FF	10.23		R
7.27		JP	8.67		FN	10.23		PG
7.33		SU	8.73		UQ	10.23		EW
7.33		J	8.73		RD	10.30		SX
7.33		BB	8.73		NT	10.30		SD
7.47		SF	8.73		BA	10.40		LR
7.47		CS	8.73		AH	10.47		VT
7.47		QZ	8.77		RG	10.47		JS
7.63		BL	8.80		FZ	10.53		PV
7.67		M	8.87		GQ	10.57		TN
7.67		KH	8.90		HL	10.70		UN
7.67		AR	8.93		HI	10.83		LT
7.73		UP	8.97		N	10.87		CK
7.73		P	9.00		QT	10.97		RR
7.87		AE	9.00		PB	10.97		K
7.90		NO	9.03		E	11.00		OS
7.90		MF	9.03		DZ	11.13		WF
7.93		CA	9.07		BH	11.23		TD
7.93		DT	9.20		KE	11.37		UE
7.93		DE	9.27		BC	11.50		DB

• = No data submitted

TAG SYMBOLS

↑ = Above control limit

∅ = Insufficient data

× = Determined to be an outlier


↓ = Below control limit

Gross Beta**Data sorted by Laboratory Average**

Average	Tag	Lab	Average	Tag	Lab	Average	Tag	Lab
11.67		L	13.10		OA	19.17	×	A
12.00		EO	13.10		NJ	21.03	×	OY
12.57		WO	13.37		CX	30.27	×	NB
12.77		D	14.73		UW	45.37	×	MX


% responding labs

Frequency distribution



% responding labs

Frequency distribution



• ≡ No data submitted

TAG SYMBOLS

↑ ≡ Above control limit

∅ ≡ Insufficient data

× ≡ Determined to be an outlier

↓ ≡ Below control limit