




Gross Alpha-Beta in Water Performance Evaluation Study

A Statistical Evaluation of the October 31, 1997 Data

006ECB98COV-009

A decorative graphic at the bottom of the page consisting of two sets of parallel black and white chevrons pointing to the right.

Gross Alpha-Beta in Water
Performance Evaluation Study
October 31, 1997



Environmental Protection Agency
National Exposure Research Laboratory
Environmental Sciences Division
Las Vegas, Nevada



UNITED STATES ENVIRONMENTAL PROTECTION AGENCY
OFFICE OF RESEARCH AND DEVELOPMENT
NATIONAL EXPOSURE RESEARCH LABORATORY
ENVIRONMENTAL SCIENCES DIVISION-LAS VEGAS
P.O. BOX 93478
LAS VEGAS, NEVADA 89193-3478
(702/798-2100)

Dear Participant,

Enclosed are the results of the Environmental Sciences Division (ESD-LV) Performance Evaluation Study for ***Gross Alpha-Beta in Water; October 31, 1997.***

The known value for each analysis was determined by gravimetric methods, checked by chemical analyses performed by ESD-LV's Radiochemistry Laboratory, and compared to the participating laboratories' grand average.

The expected precision, determined by the known value, was taken from "Table 3. Laboratory Precision: One Standard Deviation Values and Control Limits for Various Analyses", which is based on data accumulated over the years by the Performance Evaluation Program, and can be found in the Environmental Radioactivity Performance Evaluation Studies Program and Radioactive Standards Distribution Program information brochure.

Please take a few minutes to review this report and the analytical data your laboratory submitted to us. If there are any apparent discrepancies, please notify us immediately.

We encourage you to make use of the computer-automated data-entry system that has been in place for some time now. As the number of participants increases, and it becomes unrealistic for us to receive results by mail or FAX, the computer system will be our only avenue for accepting data.

If you have any questions or comments, please send a message via the data-entry system or contact Stephen Pia at 702/798-2102 or Patricia Honsa at 702/798-2141.

Sincerely,

A handwritten signature in black ink, appearing to read "Stephen Pia".

Stephen Pia
Team Leader
RADQA Program

Enclosure

NOTICE

This material has been funded wholly by the U.S. Environmental Protection Agency. It has been subjected to the Agency's review, and has been approved for publication as an EPA document.

The following pages consist of separate sections for each of the nuclides in this study with four parts per section. After the first, each part is separated from the next by a new page or a thick horizontal bar. The first page of each section is a statistical summary for the nuclide and starts with a statement of the known value, the control limits, and the warning limits.

The warning limits are placed at two normalized standard deviations above and below the known value and the control limits are three normalized standard deviations above and below the known value. If you keep control charts, these values will be useful for anticipating problems with the accuracy of your analytical methods.

The coin shaped pie chart at the top of the summary page shows the fate of all the samples sent out in number and percentage terms. The pie chart starts at the top and rotates clockwise. The first sector represents those participants who submitted analytical results within both the warning and control limits. The next sector represents those who are in the warning region but not out of control. The third sector represents those who are out of control, but have passed the outlier test. The fourth sector represents those who have failed the outlier test. The last sector represents those participants who have failed to respond properly. This is the case if no analytical results were returned, or less than three determinations were reported, or if the results were received too late. The reeding on the edge of the coin is spaced at one percent intervals, and the sector shading becomes darker as the data reliability decreases. Sectors with zero width are not shown.

The table in the center shows a number of statistical quantities calculated from the submitted data based on the mean and median values in relation to the known value, both before and after outlier removal. The lower pie chart uses the same construction as the upper chart and shows the distribution of properly submitted data in terms of deviation from the known value divided into sectors representing one, two, three, and greater than three normalized standard deviations.

The second part is an alphabetical listing, in lab-code order, of submitted data and several calculated quantities. An entry that is shaded has been rejected because of one of the reasons listed above or failure of the outlier test. The fifth and sixth columns are a measure of laboratory precision. The Range analysis is a normalized value that you may use to keep precision control charts. The eighth and ninth columns are the differences from the mean of all non-outliers and from the known value, respectively. If this value is between 2.0 and 3.0, your analytical process precision is in the warning zone; if it exceeds 3.0 it is out of control. A tag symbol may appear in the last column. Each page with tags has a symbol definition summary at the bottom. If there is no tag symbol, the data is within the control limits, but it may be in the warning zone.

The third part is a three-column listing of result average, tag symbol, and lab-code in average order excluding those labs not responding properly. In this order, all outliers and out-of-control results appear at the top or bottom of the list.

The last part is two bar chart displays showing frequency distributions of responding participants. The first chart places the known value at the center and a bar at each 0.2 unit of expected precision. The second chart places the mean of the reported measurements at the center and a bar at each 0.2 unit of standard deviation. In both cases, a bar includes those results within 0.1 unit up to the maximum of six. Any results more than six units from the center value are shown cumulatively by a shaded bar one past the sixth unit. If the central tendency of the known value distribution falls away from the center, an error in accuracy is indicated. If the distribution is broad, poor precision is indicated. The mean value distribution is similar but uses the average and standard deviation of reported results as its basis.

The Range Analysis($R + SR$) is calculated from the range, mean range and standard error of the range values. The range is the difference between the maximum and minimum results for the laboratory. The mean range is calculated by multiplying the expected precision by 1.693(for three results). The standard error of the range is calculated by multiplying the mean range by 2.575(for three results), subtracting the mean range from this product, and dividing the result by 3. If the range is greater than the mean range, then the range analysis is calculated by subtracting the mean range from the range, dividing the result by the standard error of the range and adding 1. If the mean range is greater than or equal to the range, then the range analysis is calculated by dividing the range by the mean range.

The normalized deviation of the mean from the grand average is calculated from the deviation of the mean from the grand average and the standard error of the mean values. The deviation of the mean from the grand average is calculated by subtracting the grand average from the average of the laboratory's three results. The standard error of the mean is calculated by dividing the expected precision by the square root of 3(the number of results). The normalized deviation of the mean from the grand average is calculated by dividing the deviation of the mean from the grand average by the standard error of the mean.

The normalized deviation of the mean from the known value is calculated from the deviation of the mean from the known value and the standard error of the mean values. The deviation of the mean from the known value is calculated by subtracting the known value from the average of the laboratory's three results. The standard error of the mean is calculated by dividing the expected precision by the square root of 3(the number of results). The normalized deviation of the mean from the known value is calculated by dividing the deviation of the mean from the known value by the standard error of the mean.


A complete explanation of the statistical calculations involved in the report may be found in the Environmental Radioactivity Performance Evaluation Studies Program information brochure [Draft Revision of EPA-600/4-81-004], available from Patricia Honsa, ESD-LV, 702/798-2141.

Gross Alpha


Statistical Summary

188 Participants

The known value of this nuclide is **14.7 pCi/l** with an expected precision of **5.0**; the control limits are 6.0 to 23.4; the warning regions are 6.0 to 8.9 and 20.5 to 23.4



Statistic	Respondents	Non-outliers
Mean	12.90	Grand Avg 12.26
Std. Dev.	5.70	3.92
Variance	32.54	15.33
% Coef. of Var.	44.21	31.93
% deviation of mean from known value	-12.23	-16.60
Norm. dev. of mean from known value	-0.32	-0.62
Median	12.20	12.17
% deviation of median from known value	-17.01	-17.23
Norm. dev. of median from known value	-0.44	-0.65



Gross Alpha

Lab	Res. 1	Res. 2	Res. 3	Exper. Sigma	Rng anal (R + SR)	Average	Normalized deviation (grand-avg) (known)		Tag
A	9.7	9.5	11.0	0.81	0.177	10.07	-0.76	-1.61	
AE	17.1	18.2	16.5	0.86	0.201	17.27	-1.73	0.89	
AF	9.4	7.9	8.5	0.75	0.177	8.60	-1.27	-2.11	
AH	11.0	10.6	9.2	0.95	0.213	10.27	-0.69	-1.54	
AJ	5.3	5.6	6.9	0.85	0.189	5.93	-2.19	-3.04	↓
AK	11.6	11.9	12.5	0.46	0.106	12.00	-0.09	-0.94	
AL	14.3	14.5	16.5	1.22	0.260	15.10	0.98	0.14	
AP	13.7	13.5	13.2	0.25	0.059	13.47	0.42	-0.43	
AR	14.0	14.5	10.0	2.47	0.532	12.83	0.20	-0.65	
AU	7.5	7.0	7.7	0.36	0.083	7.40	-1.68	-2.53	
AW	10.4	11.5	13.7	1.68	0.390	11.87	-0.14	-0.98	
AZ	15.0	14.5	12.8	1.15	0.260	14.10	0.64	-0.21	
BA	13.2	12.0	13.0	0.64	0.142	12.73	0.16	-0.68	
BB	5.1	5.1	5.1	0.00	0.000	5.10	-2.48	-3.33	↓
BC	15.1	16.2	14.7	0.78	0.177	15.33	1.06	0.22	
BG									•
BH	13.9	16.9	15.7	1.51	0.354	15.50	1.12	0.28	
BK	12.9	12.7	13.0	0.15	0.035	12.87	0.21	-0.64	
BL	18.1	20.7	19.8	1.32	0.307	19.53	2.52	1.67	
BM	17.5	18.4	18.4	0.52	0.106	18.10	2.02	1.18	
BN	28.3	24.1	27.8	2.29	0.496	26.73	5.01	4.17	×
BO	10.3	9.4	8.8	0.75	0.177	9.50	-0.96	-1.80	
BS	26.6	21.7	19.0	3.85	0.898	22.43	3.52	2.68	
C	11.8	12.2	12.6	0.40	0.095	12.20	-0.02	-0.87	
CA	5.1	5.0	4.9	0.10	0.024	5.00	-2.51	-3.36	↓
CC									•
CE	12.7	14.2	12.7	0.87	0.177	13.20	0.33	-0.52	
CJ	18.0	20.0	21.0	1.53	0.354	19.67	2.57	1.72	
CO	45.2	52.3	46.6	3.76	0.839	48.03	12.39	11.55	×
CP	9.2	13.5	9.3	2.45	0.508	10.67	-0.55	-1.40	
CS	12.8	12.8	11.7	0.64	0.130	12.43	0.06	-0.79	
CX	8.3	9.8	8.8	0.76	0.177	8.97	-1.14	-1.99	
D	17.2	16.2	22.0	3.10	0.685	18.47	2.15	1.30	
DB	13.4	13.1	12.3	0.57	0.130	12.93	0.23	-0.61	
DD	14.9	12.8	14.4	1.10	0.248	14.03	0.61	-0.23	
DE	12.3	12.6	12.9	0.30	0.071	12.60	0.12	-0.73	
DH	13.1	11.4	13.1	0.98	0.201	12.53	0.09	-0.75	
DO	13.4	14.8	15.1	0.91	0.201	14.43	0.75	-0.09	
DR	3.2	2.4	2.6	0.42	0.095	2.73	-3.30	-4.15	↓
DT	12.3	12.4	13.5	0.67	0.142	12.73	0.16	-0.68	
DZ	14.2	14.3	15.1	0.49	0.106	14.53	0.79	-0.06	
E	11.6	12.7	12.2	0.55	0.130	12.17	-0.03	-0.88	
EB	23.1	24.4	21.2	1.61	0.378	22.90	3.69	2.84	
EL	9.9	8.3	8.2	0.95	0.201	8.80	-1.20	-2.04	
EO	8.0	11.0	12.0	2.08	0.473	10.33	-0.67	-1.51	

• = No data submitted

TAG SYMBOLS

↑ = Above control limit

Ø = Insufficient data

× = Determined to be an outlier

↓ = Below control limit

Gross Alpha

Lab	Res. 1	Res. 2	Res. 3	Exper. Sigma	Rng anal (R + SR)	Average	Normalized deviation (grand-avg) (known) Tag	
ER	12.1	12.1	11.9	0.12	0.024	12.03	-0.08	-0.92
EV	6.7	8.1	7.7	0.72	0.165	7.50	-1.65	-2.49
EW								•
EX	8.9	10.4	7.6	1.40	0.331	8.97	-1.14	-1.99
FE	14.1	15.1	14.4	0.51	0.118	14.53	0.79	-0.06
FF	12.5	11.8	11.1	0.70	0.165	11.80	-0.16	-1.00
FJ								•
FL	13.3	18.0	13.7	2.61	0.555	15.00	0.95	0.10
FN	15.0	13.0	15.0	1.15	0.236	14.33	0.72	-0.13
GQ	12.5	12.5	11.9	0.35	0.071	12.30	0.01	-0.83
GZ	12.8	13.3	13.7	0.45	0.106	13.27	0.35	-0.50
HE	10.6	10.4	11.2	0.42	0.095	10.73	-0.53	-1.37
HI	9.3	9.5	8.0	0.81	0.177	8.93	-1.15	-2.00
HK	9.7	9.1	9.3	0.31	0.071	9.37	-1.00	-1.85
HL	7.9	6.9	8.4	0.76	0.177	7.73	-1.57	-2.41
HP								•
I	8.8	8.1	10.0	0.96	0.224	8.97	-1.14	-1.99
ID	18.3	11.0	15.5	3.68	0.862	14.93	0.93	0.08
IU								•
J								•
JE	8.2	8.6	7.7	0.45	0.106	8.17	-1.42	-2.26
JG	9.6	9.6	9.5	0.06	0.012	9.57	-0.93	-1.78
JM	13.5	14.2	12.5	0.85	0.201	13.40	0.40	-0.45
JN	7.1	9.1	12.9	2.95	0.685	9.70	-0.89	-1.73
JP								•
JQ	13.6	9.9	13.2	2.03	0.437	12.23	-0.01	-0.85
JS	11.5	12.5	10.1	1.21	0.284	11.37	-0.31	-1.15
JY	9.7	10.5	13.5	2.00	0.449	11.23	-0.36	-1.20
K	11.7	9.4	12.6	1.65	0.378	11.23	-0.36	-1.20
KE	11.8	11.5	11.4	0.21	0.047	11.57	-0.24	-1.09
KH	8.7	8.9	8.5	0.20	0.047	8.70	-1.23	-2.08
KT	14.0	15.0	16.0	1.00	0.236	15.00	0.95	0.10
KX	16.0	18.0	19.0	1.53	0.354	17.67	1.87	1.03
L	9.4	9.5	10.2	0.44	0.095	9.70	-0.89	-1.73
LE	6.9	9.0	5.8	1.63	0.378	7.23	-1.74	-2.59
LF	13.0	12.0	18.0	3.21	0.709	14.33	0.72	-0.13
LL	21.8	20.4	20.7	0.74	0.165	20.97	3.02	2.17
LR	6.8	7.1	8.2	0.74	0.165	7.37	-1.69	-2.54
LT	13.5	14.9	8.9	3.14	0.709	12.43	0.06	-0.79
M								•
MF	12.1	12.3	12.9	0.42	0.095	12.43	0.06	-0.79
MV	13.9	11.2	12.9	1.37	0.319	12.67	0.14	-0.70
MX	6.8	6.0	6.3	0.40	0.095	6.37	-2.04	-2.89
N								•
NA	17.6	17.8	14.3	1.97	0.413	16.57	1.49	0.65

• = No data submitted

TAG SYMBOLS

↑ = Above control limit

Ø = Insufficient data

× = Determined to be an outlier

↓ = Below control limit

Gross Alpha

Lab	Res. 1	Res. 2	Res. 3	Exper. Sigma	Rng anal (R + SR)	Average	Normalized deviation (grand-avg) (known)		Tag
NB	39.2	43.0	46.3	3.55	0.839	42.83	10.59	9.75	×
NH	13.0	12.0	10.0	1.53	0.354	11.67	-0.21	-1.05	
NJ	6.8	7.9	7.5	0.56	0.130	7.40	-1.68	-2.53	
NK	10.1	12.3	13.4	1.68	0.390	11.93	-0.11	-0.96	
NO	14.0	16.0	18.0	2.00	0.473	16.00	1.30	0.45	
NP	11.7	9.0	8.5	1.72	0.378	9.73	-0.88	-1.72	
O	15.1	15.5	14.8	0.35	0.083	15.13	1.00	0.15	
OA									•
OB	5.1	7.4	7.3	1.30	0.272	6.60	-1.96	-2.81	
OF	24.3	13.0	7.9	8.39	2.786	15.07	0.97	0.13	
OS	12.2	13.7	10.8	1.45	0.343	12.23	-0.01	-0.85	
OT									•
OX									•
OY									•
P	20.4	21.4	24.7	2.25	0.508	22.17	3.43	2.59	
PA	5.8	6.0	5.9	0.10	0.024	5.90	-2.20	-3.05	↓
PB	14.4	13.4	12.5	0.95	0.224	13.43	0.41	-0.44	
PD									•
PG	16.3	14.0	12.7	1.82	0.425	14.33	0.72	-0.13	
PM	16.1	17.7	14.6	1.55	0.366	16.13	1.34	0.50	
PQ									•
Q	9.5	9.7	10.8	0.70	0.154	10.00	-0.78	-1.63	
QM	8.9	12.6	9.0	2.11	0.437	10.17	-0.73	-1.57	
QP	14.0	14.8	12.1	1.39	0.319	13.63	0.48	-0.37	
QQ	14.3	12.0	12.9	1.16	0.272	13.07	0.28	-0.57	
QT	10.7	18.8	11.1	4.57	0.957	13.53	0.44	-0.40	
QU	9.1	10.5	8.6	0.98	0.224	9.40	-0.99	-1.84	
QW	13.2	9.3	11.4	1.95	0.461	11.30	-0.33	-1.18	
QX	8.6	8.9	9.3	0.35	0.083	8.93	-1.15	-2.00	
QZ	12.6	14.0	13.1	0.71	0.165	13.23	0.34	-0.51	
R	7.2	8.2	6.7	0.76	0.177	7.37	-1.69	-2.54	
RB	16.0	18.6	17.1	1.31	0.307	17.23	1.72	0.88	
RD	25.5	13.4	14.1	6.79	1.818	17.67	1.87	1.03	
RG	10.5	10.3	10.3	0.12	0.024	10.37	-0.66	-1.50	
RI	14.1	10.6	11.0	1.92	0.413	11.90	-0.12	-0.97	
RK									•
RR	18.0	17.0	18.0	0.58	0.118	17.67	1.87	1.03	
RZ	12.7	11.3	11.6	0.74	0.165	11.87	-0.14	-0.98	
S	12.6	11.7	12.2	0.45	0.106	12.17	-0.03	-0.88	
SC	9.0	13.3	14.3	2.82	0.626	12.20	-0.02	-0.87	
SD	15.4	14.2	14.8	0.60	0.142	14.80	0.88	0.03	
SF	9.6	12.0	11.0	1.21	0.284	10.87	-0.48	-1.33	
SG	9.3	8.7	8.8	0.32	0.071	8.93	-1.15	-2.00	
SI	8.4	8.5	8.5	0.06	0.012	8.47	-1.31	-2.16	
SL									•

• = No data submitted

TAG SYMBOLS

↑ = Above control limit

∅ = Insufficient data

× = Determined to be an outlier

↓ = Below control limit

Gross Alpha

Lab	Res. 1	Res. 2	Res. 3	Exper. Sigma	Rng anal (R + SR)	Average	Normalized deviation (grand-avg) (known)		Tag
SM	12.6	15.9	16.9	2.25	0.508	15.13	1.00	0.15	
SN	5.7	5.4	5.5	0.15	0.035	5.53	-2.33	-3.18	↓
SO	9.4	10.4	11.2	0.90	0.213	10.33	-0.67	-1.51	
SR	21.6	15.7	18.6	2.95	0.697	18.63	2.21	1.36	
SS	10.4	10.0	9.7	0.35	0.083	10.03	-0.77	-1.62	
SU	9.6	8.5	7.2	1.20	0.284	8.43	-1.33	-2.17	
SV	15.1	12.3	16.6	2.18	0.508	14.67	0.83	-0.01	
SX	20.8	20.5	23.5	1.65	0.354	21.60	3.24	2.39	
SZ	10.2	10.9	10.1	0.44	0.095	10.40	-0.64	-1.49	
T	13.5	11.7	13.0	0.93	0.213	12.73	0.16	-0.68	
TD	6.6	9.2	12.4	2.91	0.685	9.40	-0.99	-1.84	
TL									•
TN	7.6	6.1	5.8	0.96	0.213	6.50	-2.00	-2.84	
TO	10.2	7.9	9.8	1.23	0.272	9.30	-1.03	-1.87	
TQ	8.5	7.9	7.6	0.46	0.106	8.00	-1.48	-2.32	
TW	12.2	17.0	10.1	3.54	0.815	13.10	0.29	-0.55	
TX									•
TY	8.6	8.3	8.2	0.21	0.047	8.37	-1.35	-2.19	
U	15.2	18.0	18.6	1.81	0.402	17.27	1.73	0.89	
UA	30.0	33.5	29.5	2.18	0.473	31.00	6.49	5.65	×
UP	10.8	11.0	11.3	0.25	0.059	11.03	-0.42	-1.27	
UQ	15.3	13.5	13.7	0.99	0.213	14.17	0.66	-0.18	
UY									•
VA									•
VC									•
VH	11.3	10.5	10.6	0.44	0.095	10.80	-0.51	-1.35	
VI	7.8	7.8	8.0	0.12	0.024	7.87	-1.52	-2.37	
VJ	6.0	6.5	8.1	1.10	0.248	6.87	-1.87	-2.71	
VO	12.1	11.3	12.5	0.61	0.142	11.97	-0.10	-0.95	
VT									•
W	10.0	13.0	12.0	1.53	0.354	11.67	-0.21	-1.05	
WE	10.7	9.4	11.0	0.85	0.189	10.37	-0.66	-1.50	
WH	7.5	7.7	9.0	0.81	0.177	8.07	-1.45	-2.30	
WJ									•
WN	9.6	9.5	11.0	0.84	0.177	10.03	-0.77	-1.62	
WO									•
WP	5.7	11.4	13.0	3.84	0.862	10.03	-0.77	-1.62	
WR	17.7	17.7	19.1	0.81	0.165	18.17	2.05	1.20	
WS									•
WU									•
WW	10.2	9.9	10.0	0.15	0.035	10.03	-0.77	-1.62	
WX									•
X	15.6	13.0	12.8	1.56	0.331	13.80	0.53	-0.31	
XB									•
XC									•

• = No data submitted

TAG SYMBOLS

↑ = Above control limit

∅ = Insufficient data

× = Determined to be an outlier

↓ = Below control limit

Gross Alpha

Lab	Res. 1	Res. 2	Res. 3	Exper. Sigma	Rng anal (R + SR)	Average	Normalized deviation (grand-avg) (known)		Tag
XD									•
XF	19.0	22.5	13.9	4.32	1.030	18.47	2.15	1.30	
XI									•
XJ	13.2	11.9	13.1	0.72	0.154	12.73	0.16	-0.68	
XK	14.2	14.0	19.7	3.23	0.673	15.97	1.28	0.44	
XL	15.0	15.8	20.2	2.80	0.614	17.00	1.64	0.80	
XM	14.5	14.2	11.9	1.42	0.307	13.53	0.44	-0.40	
XN	28.0	22.5	24.7	2.77	0.650	25.07	4.44	3.59	↑

Data sorted by Laboratory Average

Average	Tag	Lab	Average	Tag	Lab	Average	Tag	Lab
2.73	↓	DR	9.37		HK	11.87		AW
5.00	↓	CA	9.40		TD	11.87		RZ
5.10	↓	BB	9.40		QU	11.90		RI
5.53	↓	SN	9.50		BO	11.93		NK
5.90	↓	PA	9.57		JG	11.97		VO
5.93	↓	AJ	9.70		L	12.00		AK
6.37		MX	9.70		JN	12.03		ER
6.50		TN	9.73		NP	12.17		S
6.60		OB	10.00		Q	12.17		E
6.87		VJ	10.03		WW	12.20		SC
7.23		LE	10.03		WP	12.20		C
7.37		R	10.03		WN	12.23		OS
7.37		LR	10.03		SS	12.23		JQ
7.40		NJ	10.07		A	12.30		GQ
7.40		AU	10.17		QM	12.43		MF
7.50		EV	10.27		AH	12.43		LT
7.73		HL	10.33		SO	12.43		CS
7.87		VI	10.33		EO	12.53		DH
8.00		TQ	10.37		WE	12.60		DE
8.07		WH	10.37		RG	12.67		MV
8.17		JE	10.40		SZ	12.73		XJ
8.37		TY	10.67		CP	12.73		T
8.43		SU	10.73		HE	12.73		DT
8.47		SI	10.80		VH	12.73		BA
8.60		AF	10.87		SF	12.83		AR
8.70		KH	11.03		UP	12.87		BK
8.80		EL	11.23		K	12.93		DB
8.93		SG	11.23		JY	13.07		QQ
8.93		QX	11.30		QW	13.10		TW
8.93		HI	11.37		JS	13.20		CE
8.97		EX	11.57		KE	13.23		QZ
8.97		I	11.67		W	13.27		GZ
8.97		CX	11.67		NH	13.40		JM
9.30		TO	11.80		FF	13.43		PB

• = No data submitted

TAG SYMBOLS

↑ = Above control limit



∅ = Insufficient data

× = Determined to be an outlier

↓ = Below control limit

Gross Alpha**Data sorted by Laboratory Average**

Average	Tag	Lab	Average	Tag	Lab	Average	Tag	Lab
13.47		AP	15.00		KT	17.67		KX
13.53		XM	15.00		FL	18.10		BM
13.53		QT	15.07		OF	18.17		WR
13.63		QP	15.10		AL	18.47		XF
13.80		X	15.13		SM	18.47		D
14.03		DD	15.13		O	18.63		SR
14.10		AZ	15.33		BC	19.53		BL
14.17		UQ	15.50		BH	19.67		CJ
14.33		PG	15.97		XK	20.97		LL
14.33		LF	16.00		NO	21.60		SX
14.33		FN	16.13		PM	22.17		P
14.43		DO	16.57		NA	22.43		BS
14.53		FE	17.00		XL	22.90		EB
14.53		DZ	17.23		RB	25.07	↑↑	XN
14.67		SV	17.27		U	26.73	×	BN
14.80		SD	17.27		AE	31.00	×	UA
14.93		ID	17.67		RR	42.83	×	NB
			17.67		RD	48.03	×	CO



• ≡ No data submitted

TAG SYMBOLS

↑↑ ≡ Above control limit


∅ ≡ Insufficient data

× ≡ Determined to be an outlier


↓↓ ≡ Below control limit

Gross Beta**Statistical Summary****188 Participants**

The known value of this nuclide is **48.9 pCi/l** with an expected precision of **5.0**; the control limits are 40.2 to 57.6; the warning regions are 40.2 to 43.1 and 54.7 to 57.6



Statistic	Respondents	Non-outliers
Mean	48.67	Grand Avg 48.88
Std. Dev.	8.55	5.88
Variance	73.15	34.61
% Coef. of Var.	17.57	12.04
% deviation of mean from known value	-0.47	-0.03
Norm. dev. of mean from known value	-0.03	0.00
Median	49.07	49.07
% deviation of median from known value	0.34	0.34
Norm. dev. of median from known value	0.02	0.03



11 / 16 ESD-LV Performance Evaluation: Gross Alpha-Beta in Water, 31-Oct-1997
Gross Beta

Lab	Res. 1	Res. 2	Res. 3	Exper. Sigma	Rng anal (R + SR)	Average	Normalized deviation (grand-avg) (known)		Tag
A	50.7	48.4	51.4	1.57	0.354	50.17	0.44	0.44	
AE	43.8	43.4	45.9	1.34	0.295	44.37	-1.57	-1.57	
AF	49.4	48.2	50.9	1.35	0.319	49.50	0.21	0.21	
AH	53.3	57.1	52.5	2.46	0.543	54.30	1.88	1.87	
AJ	47.0	46.9	45.8	0.67	0.142	46.57	-0.80	-0.81	
AK	53.8	54.9	54.7	0.59	0.130	54.47	1.93	1.93	
AL	44.0	46.2	46.1	1.24	0.260	45.43	-1.20	-1.20	
AP	47.8	50.4	51.0	1.70	0.378	49.73	0.29	0.29	
AR	54.0	52.3	54.1	1.01	0.213	53.47	1.59	1.58	
AU	45.4	46.3	46.7	0.67	0.154	46.13	-0.95	-0.96	
AW	49.4	49.7	54.8	3.03	0.638	51.30	0.84	0.83	
AZ	51.7	47.6	48.0	2.26	0.484	49.10	0.07	0.07	
BA	53.9	50.2	51.7	1.86	0.437	51.93	1.06	1.05	
BB	51.3	45.4	48.4	2.95	0.697	48.37	-0.18	-0.18	
BC	47.0	46.9	46.4	0.32	0.071	46.77	-0.73	-0.74	
BG									•
BH	45.0	46.3	45.5	0.66	0.154	45.60	-1.14	-1.14	
BK	50.2	48.9	51.8	1.45	0.343	50.30	0.49	0.48	
BL	47.9	47.1	48.4	0.66	0.154	47.80	-0.38	-0.38	
BM	51.6	48.3	48.0	2.00	0.425	49.30	0.14	0.14	
BN	69.9	65.4	83.2	9.26	3.101	72.83	8.30	8.29	×
BO	46.2	47.3	47.7	0.78	0.177	47.07	-0.63	-0.64	
BS	51.9	50.2	47.4	2.27	0.532	49.83	0.33	0.32	
C	49.9	50.3	46.7	1.97	0.425	48.97	0.03	0.02	
CA	20.1	19.2	20.7	0.75	0.177	20.00	-10.01	-10.01	×
CC									•
CE	46.7	49.6	48.1	1.45	0.343	48.13	-0.26	-0.27	
CJ	51.0	54.0	47.0	3.51	0.827	50.67	0.62	0.61	
CO	33.5	36.6	33.6	1.76	0.366	34.57	-4.96	-4.97	↓
CP	58.2	63.1	59.9	2.49	0.579	60.40	3.99	3.98	↑
CS	48.8	41.3	42.3	4.07	0.886	44.13	-1.65	-1.65	
CX	37.4	38.7	34.8	1.99	0.461	36.97	-4.13	-4.13	↓
D	52.5	53.1	54.7	1.14	0.260	53.43	1.58	1.57	
DB	50.0	49.9	48.6	0.78	0.165	49.50	0.21	0.21	
DD	47.6	48.9	49.1	0.81	0.177	48.53	-0.12	-0.13	
DE	52.0	52.5	53.3	0.66	0.154	52.60	1.29	1.28	
DH	44.8	48.5	51.7	3.45	0.815	48.33	-0.19	-0.20	
DO	55.0	53.0	51.0	2.00	0.473	53.00	1.43	1.42	
DR	50.0	49.3	54.1	2.59	0.567	51.13	0.78	0.77	
DT	48.2	48.5	46.7	0.96	0.213	47.80	-0.38	-0.38	
DZ	39.7	38.7	41.2	1.26	0.295	39.87	-3.12	-3.13	↓
E	49.3	48.0	49.3	0.75	0.154	48.87	-0.01	-0.01	
EB	53.3	52.9	53.6	0.35	0.083	53.27	1.52	1.51	
EL	48.6	45.4	47.0	1.60	0.378	47.00	-0.65	-0.66	
EO	57.0	58.0	60.0	1.53	0.354	58.33	3.27	3.27	↑

• = No data submitted

TAG SYMBOLS

↑ = Above control limit

∅ = Insufficient data

× = Determined to be an outlier

↓ = Below control limit

Gross Beta

Lab	Res. 1	Res. 2	Res. 3	Exper. Sigma	Rng anal (R + SR)	Average	Normalized deviation (grand-avg) (known)		Tag
ER									•
EV	54.8	60.6	54.0	3.60	0.780	56.47	2.63	2.62	
EW	50.1	50.6	49.4	0.60	0.142	50.03	0.40	0.39	
EX	45.1	47.8	47.5	1.48	0.319	46.80	-0.72	-0.73	
FE	41.4	41.3	40.4	0.55	0.118	41.03	-2.72	-2.73	
FF	52.2	52.1	47.5	2.69	0.555	50.60	0.59	0.59	
FJ									•
FL	50.7	37.5	55.2	9.20	3.078	47.80	-0.38	-0.38	
FN	49.0	53.0	52.0	2.08	0.473	51.33	0.85	0.84	
GQ	43.0	43.0	43.0	0.00	0.000	43.00	-2.04	-2.04	
GZ	60.1	61.0	59.2	0.90	0.213	60.10	3.89	3.88	↑
HE	41.0	39.8	39.1	0.96	0.224	39.97	-3.09	-3.09	↓
HI	48.9	50.3	50.3	0.81	0.165	49.83	0.33	0.32	
HK	55.5	57.5	54.2	1.66	0.390	55.73	2.37	2.37	
HL	49.0	50.4	49.0	0.81	0.165	49.47	0.20	0.20	
HP									•
I	45.1	50.3	48.7	2.66	0.614	48.03	-0.29	-0.30	
ID	58.0	44.7	54.4	6.88	2.088	52.37	1.21	1.20	
IU	52.0	48.0	46.0	3.06	0.709	48.67	-0.08	-0.08	
J									•
JE	45.1	44.6	40.5	2.52	0.543	43.40	-1.90	-1.91	
JG	48.8	48.7	49.7	0.55	0.118	49.07	0.06	0.06	
JM	50.8	49.2	46.8	2.01	0.473	48.93	0.02	0.01	
JN	48.4	47.9	53.9	3.33	0.709	50.07	0.41	0.40	
JP									•
JQ	33.7	44.9	46.3	6.91	1.930	41.63	-2.51	-2.52	
JS	51.3	54.2	47.9	3.15	0.744	51.13	0.78	0.77	
JY	43.7	44.8	45.3	0.82	0.189	44.60	-1.48	-1.49	
K	42.9	43.2	43.0	0.15	0.035	43.03	-2.03	-2.03	
KE	39.2	41.9	35.6	3.16	0.744	38.90	-3.46	-3.46	↓
KH	43.7	47.7	50.5	3.42	0.803	47.30	-0.55	-0.55	
KT	60.0	75.0	70.0	7.64	2.470	68.33	6.74	6.73	↑
KX	54.0	48.0	52.0	3.06	0.709	51.33	0.85	0.84	
L	53.1	48.3	48.1	2.83	0.591	49.83	0.33	0.32	
LE	42.7	46.3	47.7	2.58	0.591	45.57	-1.15	-1.15	
LF	43.0	44.0	56.0	7.23	2.020	47.67	-0.42	-0.43	
LL	44.8	43.3	41.8	1.50	0.354	43.30	-1.93	-1.94	
LR	59.0	59.0	59.3	0.17	0.035	59.10	3.54	3.53	↑
LT	55.8	50.7	48.4	3.79	0.874	51.63	0.95	0.95	
M									•
MF	56.0	57.4	61.1	2.63	0.602	58.17	3.22	3.21	↑
MV	38.4	49.3	46.9	5.73	1.548	44.87	-1.39	-1.40	
MX	36.7	42.6	43.2	3.59	0.768	40.83	-2.79	-2.79	
N									•
NA	53.8	52.5	54.3	0.93	0.213	53.53	1.61	1.61	

• = No data submitted

∅ = Insufficient data

TAG SYMBOLS

× = Determined to be an outlier

↑ = Above control limit

↓ = Below control limit

13 / 16 ESD-LV Performance Evaluation: Gross Alpha-Beta in Water, 31-Oct-1997

Gross Beta

Lab	Res. 1	Res. 2	Res. 3	Exper. Sigma	Rng anal (R + SR)	Average	Normalized deviation (grand-avg) (known)		Tag
NB	26.5	30.1	33.6	3.55	0.839	30.07	-6.52	-6.52	↓
NH	54.0	53.0	52.0	1.00	0.236	53.00	1.43	1.42	
NJ	49.4	46.7	50.5	1.96	0.449	48.87	-0.01	-0.01	
NK	41.5	50.9	39.8	5.98	1.593	44.07	-1.67	-1.67	
NO	47.0	49.0	51.0	2.00	0.473	49.00	0.04	0.03	
NP	44.4	41.4	49.9	4.31	1.008	45.23	-1.26	-1.27	
O	51.0	50.6	51.3	0.35	0.083	50.97	0.72	0.72	
OA									•
OB	48.5	50.5	50.8	1.25	0.272	49.93	0.36	0.36	
OF	48.5	45.8	44.4	2.08	0.484	46.23	-0.92	-0.92	
OS	50.8	50.0	50.0	0.46	0.095	50.27	0.48	0.47	
OT	53.5	50.9	51.7	1.33	0.307	52.03	1.09	1.09	
OX									•
OY									•
P	53.3	48.3	51.3	2.52	0.591	50.97	0.72	0.72	
PA	65.9	62.0	61.4	2.44	0.532	63.10	4.92	4.92	↑
PB	48.5	45.2	46.6	1.66	0.390	46.77	-0.73	-0.74	
PD	76.5	101.6	101.2	14.38	4.743	93.10	15.32	15.31	×
PG	54.4	51.1	52.0	1.71	0.390	52.50	1.25	1.25	
PM	53.8	52.6	49.6	2.16	0.496	52.00	1.08	1.07	
PQ									•
Q	49.1	50.4	52.7	1.82	0.425	50.73	0.64	0.64	
QM	40.8	52.3	45.4	5.79	1.683	46.17	-0.94	-0.95	
QP	50.4	68.1	54.6	9.25	3.078	57.70	3.05	3.05	↑
QQ	57.1	51.8	56.4	2.88	0.626	55.10	2.15	2.15	
QT	49.3	53.8	53.1	2.42	0.532	52.07	1.10	1.10	
QU	46.7	40.0	43.8	3.36	0.791	43.50	-1.87	-1.87	
QW	37.3	44.0	45.3	4.29	0.945	42.20	-2.32	-2.32	
QX	51.4	51.0	53.1	1.12	0.248	51.83	1.02	1.02	
QZ	41.8	42.1	40.7	0.74	0.165	41.53	-2.55	-2.55	
R	51.6	54.0	54.6	1.59	0.354	53.40	1.56	1.56	
RB	57.8	53.5	54.1	2.33	0.508	55.13	2.16	2.16	
RD	61.5	43.2	44.9	10.11	3.213	49.87	0.34	0.33	
RG	55.8	55.8	55.8	0.02	0.000	55.80	2.40	2.39	
RI	46.5	53.1	52.6	3.67	0.780	50.73	0.64	0.64	
RK									•
RR	56.0	55.0	57.0	1.00	0.236	56.00	2.46	2.46	
RZ	47.6	47.6	50.0	1.39	0.284	48.40	-0.17	-0.17	
S	56.8	52.6	53.9	2.15	0.496	54.43	1.92	1.92	
SC	41.7	44.7	49.8	4.10	0.957	45.40	-1.21	-1.21	
SD	46.7	54.8	50.8	4.05	0.957	50.77	0.65	0.65	
SF	52.0	49.0	55.0	3.00	0.709	52.00	1.08	1.07	
SG	41.8	49.4	46.0	3.81	0.898	45.73	-1.09	-1.10	
SI	45.1	45.4	45.9	0.40	0.095	45.47	-1.18	-1.19	
SL									•

• = No data submitted

TAG SYMBOLS

↑ = Above control limit

∅ = Insufficient data

× = Determined to be an outlier

↓ = Below control limit

Gross Beta

Lab	Res. 1	Res. 2	Res. 3	Exper. Sigma	Rng anal (R + SR)	Average	Normalized deviation (grand-avg) (known)		Tag
SM	46.0	47.4	46.0	0.81	0.165	46.47	-0.84	-0.84	
SN	48.1	48.0	46.0	1.18	0.248	47.37	-0.53	-0.53	
SO	47.6	48.4	46.6	0.90	0.213	47.53	-0.47	-0.47	
SR	55.1	52.9	52.4	1.44	0.319	53.47	1.59	1.58	
SS	43.7	49.3	49.2	3.20	0.662	47.40	-0.51	-0.52	
SU	46.4	51.7	48.4	2.68	0.626	48.83	-0.02	-0.02	
SV	46.8	37.5	40.3	4.77	1.188	41.53	-2.55	-2.55	
SX	61.4	66.7	66.1	2.90	0.626	64.73	5.49	5.48	↑
SZ	50.1	53.5	48.5	2.55	0.591	50.70	0.63	0.62	
T	42.5	45.1	49.8	3.70	0.862	45.80	-1.07	-1.07	
TD	44.0	42.3	38.9	2.60	0.602	41.73	-2.48	-2.48	
TL									•
TN	37.4	38.8	35.8	1.50	0.354	37.33	-4.00	-4.01	↓
TO	52.3	49.9	39.6	6.75	1.953	47.27	-0.56	-0.57	
TQ	52.7	47.8	46.7	3.19	0.709	49.07	0.06	0.06	
TW	23.1	30.0	35.8	6.36	1.953	29.63	-6.67	-6.67	↓
TX									•
TY	50.4	50.3	50.6	0.15	0.035	50.43	0.54	0.53	
U	49.0	49.8	47.7	1.06	0.248	48.83	-0.02	-0.02	
UA	59.0	61.6	60.0	1.31	0.307	60.20	3.92	3.91	↑
UP	51.2	51.2	53.4	1.27	0.260	51.93	1.06	1.05	
UQ	51.4	52.2	50.8	0.70	0.165	51.47	0.89	0.89	
UY									•
VA									•
VC									•
VH	56.1	51.2	57.3	3.23	0.721	54.87	2.07	2.07	
VI	3.2	3.3	3.6	0.21	0.047	3.37	-15.77	-15.77	×
VJ	47.5	46.2	50.6	2.26	0.520	48.10	-0.27	-0.28	
VO	50.8	51.3	46.9	2.41	0.520	49.67	0.27	0.27	
VT									•
W	44.0	46.0	48.0	2.00	0.473	46.00	-1.00	-1.00	
WE	58.5	53.0	52.3	3.40	0.732	54.60	1.98	1.97	
WH	48.0	45.6	49.2	1.83	0.425	47.60	-0.44	-0.45	
WJ									•
WN	28.4	32.7	32.1	2.33	0.508	31.07	-6.17	-6.18	↓
WO									•
WP									•
WR	46.7	48.1	48.3	0.87	0.189	47.70	-0.41	-0.42	
WS									•
WU									•
WW	44.6	44.5	57.8	7.65	2.088	48.97	0.03	0.02	
WX									•
X	49.4	43.4	43.7	3.38	0.709	45.50	-1.17	-1.18	
XB									•
XC									•

• ≡ No data submitted

TAG SYMBOLS

↑ ≡ Above control limit

∅ ≡ Insufficient data

× ≡ Determined to be an outlier

↓ ≡ Below control limit

15 / 16 ESD-LV Performance Evaluation: Gross Alpha-Beta in Water, 31-Oct-1997

Gross Beta

Lab	Res. 1	Res. 2	Res. 3	Exper. Sigma	Rng anal (R + SR)	Average	Normalized deviation (grand-avg) (known)		Tag
XD									•
XF	52.8	53.0	50.8	1.22	0.260	52.20	1.15	1.14	
XI									•
XJ	19.5	27.7	16.9	5.64	1.525	21.37	-9.53	-9.54	×
XK	50.2	54.8	52.1	2.31	0.543	52.37	1.21	1.20	
XL	50.4	47.4	51.2	2.00	0.449	49.67	0.27	0.27	
XM	34.3	33.9	34.0	0.21	0.047	34.07	-5.13	-5.14	↓
XN	48.2	46.6	45.9	1.18	0.272	46.90	-0.69	-0.69	

Data sorted by Laboratory Average

Average	Tag	Lab	Average	Tag	Lab	Average	Tag	Lab
3.37	×	VI	45.50		X	48.53		DD
20.00	×	CA	45.57		LE	48.67		IU
21.37	×	XJ	45.60		BH	48.83		U
29.63	↓	TW	45.73		SG	48.83		SU
30.07	↓	NB	45.80		T	48.87		NJ
31.07	↓	WN	46.00		W	48.87		E
34.07	↓	XM	46.13		AU	48.93		JM
34.57	↓	CO	46.17		QM	48.97		WW
36.97	↓	CX	46.23		OF	48.97		C
37.33	↓	TN	46.47		SM	49.00		NO
38.90	↓	KE	46.57		AJ	49.07		TQ
39.87	↓	DZ	46.77		PB	49.07		JG
39.97	↓	HE	46.77		BC	49.10		AZ
40.83		MX	46.80		EX	49.30		BM
41.03		FE	46.90		XN	49.47		HL
41.53		SV	47.00		EL	49.50		DB
41.53		QZ	47.07		BO	49.50		AF
41.63		JQ	47.27		TO	49.67		XL
41.73		TD	47.30		KH	49.67		VO
42.20		QW	47.37		SN	49.73		AP
43.00		GQ	47.40		SS	49.83		L
43.03		K	47.53		SO	49.83		HI
43.30		LL	47.60		WH	49.83		BS
43.40		JE	47.67		LF	49.87		RD
43.50		QU	47.70		WR	49.93		OB
44.07		NK	47.80		FL	50.03		EW
44.13		CS	47.80		DT	50.07		JN
44.37		AE	47.80		BL	50.17		A
44.60		JY	48.03		I	50.27		OS
44.87		MV	48.10		VJ	50.30		BK
45.23		NP	48.13		CE	50.43		TY
45.40		SC	48.33		DH	50.60		FF
45.43		AL	48.37		BB	50.67		CJ
45.47		SI	48.40		RZ	50.70		SZ

• = No data submitted

TAG SYMBOLS

↑ = Above control limit

∅ = Insufficient data



× = Determined to be an outlier

↓ = Below control limit

Gross Beta

Data sorted by Laboratory Average

Average	Tag	Lab	Average	Tag	Lab	Average	Tag	Lab
50.73		RI	52.07		QT	54.87		VH
50.73		Q	52.20		XF	55.10		QQ
50.77		SD	52.37		XK	55.13		RB
50.97		P	52.37		ID	55.73		HK
50.97		O	52.50		PG	55.80		RG
51.13		DR	52.60		DE	56.00		RR
51.13		JS	53.00		NH	56.47		EV
51.30		AW	53.00		DO	57.70	↑	QP
51.33		KX	53.27		EB	58.17	↑	MF
51.33		FN	53.40		R	58.33	↑	EO
51.47		UQ	53.43		D	59.10	↑	LR
51.63		LT	53.47		AR	60.10	↑	GZ
51.83		QX	53.47		SR	60.20	↑	UA
51.93		UP	53.53		NA	60.40	↑	CP
51.93		BA	54.30		AH	63.10	↑	PA
52.00		SF	54.43		S	64.73	↑	SX
52.00		PM	54.47		AK	68.33	↑	KT
52.03		OT	54.60		WE	72.83	×	BN
						93.10	×	PD



• ≡ No data submitted

∅ ≡ Insufficient data

TAG SYMBOLS

× ≡ Determined to be an outlier

↑ ≡ Above control limit

↓ ≡ Below control limit