




Gross Alpha-Beta in Water Performance Evaluation Study

A Statistical Evaluation
of the July 24, 1998 Data



Gross Alpha-Beta in Water
Performance Evaluation Study
July 24, 1998



Environmental Protection Agency
National Exposure Research Laboratory
Environmental Sciences Division
Las Vegas, Nevada



UNITED STATES ENVIRONMENTAL PROTECTION AGENCY
OFFICE OF RESEARCH AND DEVELOPMENT
NATIONAL EXPOSURE RESEARCH LABORATORY
ENVIRONMENTAL SCIENCES DIVISION-LAS VEGAS
P.O. BOX 93478
LAS VEGAS, NEVADA 89193-3478
(702/798-2100)

Dear Participant,

Enclosed are the results of the Environmental Sciences Division (ESD-LV) Performance Evaluation Study for *Gross Alpha-Beta in Water; July 24, 1998*.

The known value for each analysis was determined by gravimetric methods, checked by chemical analyses performed by ESD-LV's Radiochemistry Laboratory, and compared to the participating laboratories' grand average.

The expected precision, determined by the known value, was taken from "Table 3. Laboratory Precision: One Standard Deviation Values and Control Limits for Various Analyses", which is based on data accumulated over the years by the Performance Evaluation Program, and can be found in the Environmental Radioactivity Performance Evaluation Studies Program and Radioactive Standards Distribution Program information brochure.

Please take a few minutes to review this report and the analytical data your laboratory submitted to us. If there are any apparent discrepancies, please notify us immediately.

We encourage you to make use of the computer-automated data-entry system that has been in place for some time now. As the number of participants increases, and it becomes unrealistic for us to receive results by mail or FAX, the computer system will be our only avenue for accepting data.

If you have any questions or comments, please send a message via the data-entry system or contact Stephen Pia at 702/798-2102 or Patricia Honsa at 702/798-2141.

Sincerely,

A handwritten signature in cursive script that reads "Stephen Pia".

Stephen Pia
Team Leader
RADQA Program

Enclosure

NOTICE

**This material has been funded wholly by
the U.S. Environmental Protection Agency.
It has been subjected to the Agency's review,
and has been approved for publication as
an EPA document.**

The following pages consist of separate sections for each of the nuclides in this study with four parts per section. After the first, each part is separated from the next by a new page or a thick horizontal bar. The first page of each section is a statistical summary for the nuclide and starts with a statement of the known value, the control limits, and the warning limits.

The warning limits are placed at two normalized standard deviations above and below the known value and the control limits are three normalized standard deviations above and below the known value. If you keep control charts, these values will be useful for anticipating problems with the accuracy of your analytical methods.

The coin shaped pie chart at the top of the summary page shows the fate of all the samples sent out in number and percentage terms. The pie chart starts at the top and rotates clockwise. The first sector represents those participants who submitted analytical results within both the warning and control limits. The next sector represents those who are in the warning region but not out of control. The third sector represents those who are out of control, but have passed the outlier test. The fourth sector represents those who have failed the outlier test. The last sector represents those participants who have failed to respond properly. This is the case if no analytical results were returned, or less than three determinations were reported, or if the results were received too late. The reeding on the edge of the coin is spaced at one percent intervals, and the sector shading becomes darker as the data reliability decreases. Sectors with zero width are not shown.

The table in the center shows a number of statistical quantities calculated from the submitted data based on the mean and median values in relation to the known value, both before and after outlier removal. The lower pie chart uses the same construction as the upper chart and shows the distribution of properly submitted data in terms of deviation from the known value divided into sectors representing one, two, three, and greater than three normalized standard deviations.

The second part is an alphabetical listing, in lab-code order, of submitted data and several calculated quantities. An entry that is shaded has been rejected because of one of the reasons listed above or failure of the outlier test. The fifth and sixth columns are a measure of laboratory precision. The Range analysis is a normalized value that you may use to keep precision control charts. The eighth and ninth columns are the differences from the mean of all non-outliers and from the known value, respectively. If this value is between 2.0 and 3.0, your analytical process precision is in the warning zone; if it exceeds 3.0 it is out of control. A tag symbol may appear in the last column. Each page with tags has a symbol definition summary at the bottom. If there is no tag symbol, the data is within the control limits, but it may be in the warning zone.

The third part is a three-column listing of result average, tag symbol, and lab-code in average order excluding those labs not responding properly. In this order, all outliers and out-of-control results appear at the top or bottom of the list.

The last part is two bar chart displays showing frequency distributions of responding participants. The first chart places the known value at the center and a bar at each 0.2 unit of expected precision. The second chart places the mean of the reported measurements at the center and a bar at each 0.2 unit of standard deviation. In both cases, a bar includes those results within 0.1 unit up to the maximum of six. Any results more than six units from the center value are shown cumulatively by a shaded bar one past the sixth unit. If the central tendency of the known value distribution falls away from the center, an error in accuracy is indicated. If the distribution is broad, poor precision is indicated. The mean value distribution is similar but uses the average and standard deviation of reported results as its basis.

The Range Analysis(R + SR) is calculated from the range, mean range and standard error of the range values. The range is the difference between the maximum and minimum results for the laboratory. The mean range is calculated by multiplying the expected precision by 1.693(for three results). The standard error of the range is calculated by multiplying the mean range by 2.575(for three results), subtracting the mean range from this product, and dividing the result by 3. If the range is greater than the mean range, then the range analysis is calculated by subtracting the mean range from the range, dividing the result by the standard error of the range and adding 1. If the mean range is greater than or equal to the range, then the range analysis is calculated by dividing the range by the mean range.

The normalized deviation of the mean from the grand average is calculated from the deviation of the mean from the grand average and the standard error of the mean values. The deviation of the mean from the grand average is calculated by subtracting the grand average from the average of the laboratory's three results. The standard error of the mean is calculated by dividing the expected precision by the square root of 3(the number of results). The normalized deviation of the mean from the grand average is calculated by dividing the deviation of the mean from the grand average by the standard error of the mean.


The normalized deviation of the mean from the known value is calculated from the deviation of the mean from the known value and the standard error of the mean values. The deviation of the mean from the known value is calculated by subtracting the known value from the average of the laboratory's three results. The standard error of the mean is calculated by dividing the expected precision by the square root of 3(the number of results). The normalized deviation of the mean from the known value is calculated by dividing the deviation of the mean from the known value by the standard error of the mean.

A complete explanation of the statistical calculations involved in the report may be found in the Environmental Radioactivity Performance Evaluation Studies Program information brochure [Draft Revision of EPA-600/4-81-004], available from Patricia Honsa, ESD-LV, 702/798-2141.


Gross Alpha**Statistical Summary**

186 Participants

The known value of this nuclide is **7.2 pCi/l** with an expected precision of **5.0**; the control limits are 0.0 to 15.9; the warning regions are 0.0 to 1.4 and 13.0 to 15.9



Statistic	Respondents	Non-outliers
Mean	7.67	Grand Avg 7.27
Std. Dev.	3.92	1.98
Variance	15.33	3.94
% Coef. of Var.	51.08	27.31
% deviation of mean from known value	6.47	0.93
Norm. dev. of mean from known value	0.12	0.03
Median	7.10	7.03
% deviation of median from known value	-1.39	-2.31
Norm. dev. of median from known value	-0.03	-0.08



Gross Alpha

Lab	Res. 1	Res. 2	Res. 3	Exper. Sigma	Rng anal (R + SR)	Average	Normalized deviation (grand-avg) (known) Tag	
A	8.0	8.1	7.8	0.15	0.035	7.97	0.24	0.27
AE	10.7	8.7	7.7	1.53	0.354	9.03	0.61	0.64
AF	6.1	8.5	6.9	1.22	0.284	7.17	-0.03	-0.01
AH	8.2	7.9	7.5	0.35	0.083	7.87	0.21	0.23
AJ	8.9	8.4	8.8	0.26	0.059	8.70	0.50	0.52
AK	9.9	9.6	8.9	0.51	0.118	9.47	0.76	0.79
AL	7.1	7.3	6.0	0.70	0.154	6.80	-0.16	-0.14
AP								•
AR	8.3	8.8	8.2	0.32	0.071	8.43	0.40	0.43
AU	8.3	8.1	9.5	0.76	0.165	8.63	0.47	0.50
AW	10.9	10.9	12.8	1.10	0.224	11.53	1.48	1.50
AZ	8.9	9.4	7.7	0.87	0.201	8.67	0.48	0.51
BA	6.8	6.2	6.9	0.38	0.083	6.63	-0.22	-0.20
BB	6.6	6.5	6.2	0.21	0.047	6.43	-0.29	-0.27
BC	7.8	7.9	8.6	0.44	0.095	8.10	0.29	0.31
BG								•
BH	7.5	7.1	7.1	0.23	0.047	7.23	-0.01	0.01
BK	7.5	8.4	7.5	0.52	0.106	7.80	0.18	0.21
BL	15.5	16.3	17.1	0.80	0.189	16.30	3.13	3.15 ×
BM	9.1	8.7	7.3	0.95	0.213	8.37	0.38	0.40
BN	6.6	5.1	5.6	0.76	0.177	5.77	-0.52	-0.50
BO	3.4	3.2	4.0	0.42	0.095	3.53	-1.29	-1.27
BS	9.4	10.4	6.5	2.03	0.461	8.77	0.52	0.54
C	7.5	6.7	7.3	0.42	0.095	7.17	-0.03	-0.01
CA	5.3	5.9	6.1	0.42	0.095	5.77	-0.52	-0.50
CC								•
CE	5.6	4.8	7.3	1.28	0.295	5.90	-0.47	-0.45
CJ	5.7	5.9	4.7	0.64	0.142	5.43	-0.64	-0.61
CO	3.2	2.4	5.4	1.55	0.354	3.67	-1.25	-1.22
CP	11.9	8.7	11.2	1.68	0.378	10.60	1.15	1.18
CS								•
CX	46.1	47.8	42.9	2.49	0.579	45.60	13.28	13.30 ×
D	9.9	8.1	11.2	1.56	0.366	9.73	0.85	0.88
DB	7.1	7.7	6.6	0.55	0.130	7.13	-0.05	-0.02
DD								•
DE	8.8	9.5	9.9	0.56	0.130	9.40	0.74	0.76
DH	6.3	6.8	4.3	1.32	0.295	5.80	-0.51	-0.48
DO	10.7	11.2	9.0	1.15	0.260	10.30	1.05	1.07
DR								•
DT	8.9	6.9	7.1	1.10	0.236	7.63	0.13	0.15
DZ	11.6	12.8	10.8	1.01	0.236	11.73	1.55	1.57
E	9.4	9.8	8.3	0.78	0.177	9.17	0.66	0.68
EB	7.0	5.8	6.6	0.61	0.142	6.47	-0.28	-0.25
EL	6.2	4.0	6.1	1.24	0.260	5.43	-0.64	-0.61
EO	7.0	7.0	5.0	1.15	0.236	6.33	-0.32	-0.30

• = No data submitted

TAG SYMBOLS

↑ = Above control limit

Ø = Insufficient data

× = Determined to be an outlier

↓ = Below control limit

Gross Alpha								
Lab	Res. 1	Res. 2	Res. 3	Exper. Sigma	Rng anal (R + SR)	Average	Normalized deviation (grand-avg) (known) Tag	
ER	6.6	6.9	7.2	0.30	0.071	6.90	-0.13	-0.10
EV	5.2	4.2	3.9	0.68	0.154	4.43	-0.98	-0.96
EW								
EX	4.1	4.8	4.4	0.35	0.083	4.43	-0.98	-0.96
FE	8.2	8.4	10.0	0.99	0.213	8.87	0.55	0.58
FF	5.1	5.3	5.5	0.20	0.047	5.30	-0.68	-0.66
FJ	4.6	5.2	5.5	0.46	0.106	5.10	-0.75	-0.73
FL	7.2	4.3	5.1	1.50	0.343	5.53	-0.60	-0.58
FN	8.0	7.0	9.0	1.00	0.236	8.00	0.25	0.28
GQ	7.6	6.6	6.3	0.68	0.154	6.83	-0.15	-0.13
GZ	7.6	7.4	8.4	0.53	0.118	7.80	0.18	0.21
HE	11.0	12.4	12.4	0.81	0.165	11.93	1.62	1.64
HI	9.3	7.6	6.8	1.28	0.295	7.90	0.22	0.24
HK	6.1	6.1	6.1	0.00	0.000	6.10	-0.40	-0.38
HL	4.0	4.9	5.3	0.67	0.154	4.73	-0.88	-0.85
HP	4.8	5.4	6.1	0.65	0.154	5.43	-0.64	-0.61
I	7.6	3.4	4.9	2.13	0.496	5.30	-0.68	-0.66
ID	6.0	8.0	4.8	1.62	0.378	6.27	-0.35	-0.32
J								
JE	7.4	6.6	7.4	0.46	0.095	7.13	-0.05	-0.02
JG								
JM	5.5	5.9	6.0	0.26	0.059	5.80	-0.51	-0.48
JN	5.1	6.1	5.4	0.51	0.118	5.53	-0.60	-0.58
JP	7.2	7.8	8.5	0.65	0.154	7.83	0.20	0.22
JQ	6.0	7.0	5.7	0.68	0.154	6.23	-0.36	-0.33
JS	8.5	7.7	7.7	0.46	0.095	7.97	0.24	0.27
JY	8.6	7.6	11.2	1.86	0.425	9.13	0.65	0.67
K								
KE	7.7	9.2	8.3	0.75	0.177	8.40	0.39	0.42
KH	7.9	7.9	7.6	0.17	0.035	7.80	0.18	0.21
KT	8.0	9.5	11.0	1.50	0.354	9.50	0.77	0.80
KX	9.0	9.0	8.5	0.29	0.059	8.83	0.54	0.57
L	4.9	4.8	5.5	0.38	0.083	5.07	-0.76	-0.74
LE								
LF	6.3	6.5	6.5	0.12	0.024	6.43	-0.29	-0.27
LL								
LR	5.5	5.3	4.6	0.47	0.106	5.13	-0.74	-0.72
M								
MF	7.7	7.6	9.0	0.78	0.165	8.10	0.29	0.31
MT								
MV								
MX	4.6	3.3	4.0	0.65	0.154	3.97	-1.14	-1.12
N	5.0	5.4	5.2	0.20	0.047	5.20	-0.72	-0.69
NA	8.6	7.1	8.1	0.76	0.177	7.93	0.23	0.25
NB	4.1	3.3	5.2	0.95	0.224	4.20	-1.06	-1.04

• = No data submitted

TAG SYMBOLS

↑ = Above control limit

Ø = Insufficient data

× = Determined to be an outlier

↓ = Below control limit

Gross Alpha

Lab	Res. 1	Res. 2	Res. 3	Exper. Sigma	Rng anal (R + SR)	Average	Normalized deviation (grand-avg) (known) Tag	
NH	6.0	5.0	5.0	0.58	0.118	5.33	-0.67	-0.65
NJ	6.2	6.5	6.3	0.15	0.035	6.33	-0.32	-0.30
NK								•
NO								•
NP	6.9	5.8	7.0	0.67	0.142	6.57	-0.24	-0.22
O	9.9	11.5	12.0	1.10	0.248	11.13	1.34	1.36
OA	7.4	6.9	5.1	1.21	0.272	6.47	-0.28	-0.25
OB	9.4	10.4	10.1	0.51	0.118	9.97	0.94	0.96
OF	9.6	8.2	9.2	0.72	0.165	9.00	0.60	0.62
OS	6.0	6.0	6.0	0.00	0.000	6.00	-0.44	-0.42
OT								•
OY								•
P	5.0	5.8	8.7	1.95	0.437	6.50	-0.27	-0.24
PB								•
PD								•
PG	0.1	0.3	0.2	0.10	0.024	0.20	-2.45	-2.42
PM	7.4	6.1	6.7	0.65	0.154	6.73	-0.18	-0.16
Q	8.6	6.3	8.6	1.33	0.272	7.83	0.20	0.22
QM	5.3	4.2	5.7	0.78	0.177	5.07	-0.76	-0.74
QP	10.6	10.6	10.4	0.12	0.024	10.53	1.13	1.15
QQ	6.5	8.3	5.2	1.56	0.366	6.67	-0.21	-0.18
QT	5.9	7.4	11.6	2.95	0.673	8.30	0.36	0.38
QU	11.6	10.5	9.3	1.15	0.272	10.47	1.11	1.13
QW	10.3	9.7	10.6	0.46	0.106	10.20	1.02	1.04
QX	9.3	8.8	8.7	0.32	0.071	8.93	0.58	0.60
QZ	5.9	5.7	5.6	0.15	0.035	5.73	-0.53	-0.51
R	5.9	5.6	5.8	0.15	0.035	5.77	-0.52	-0.50
RB	8.6	7.4	8.6	0.69	0.142	8.20	0.32	0.35
RD	6.8	5.3	3.3	1.76	0.413	5.13	-0.74	-0.72
RG	10.3	9.5	9.5	0.46	0.095	9.77	0.87	0.89
RI	7.2	6.2	5.3	0.95	0.224	6.23	-0.36	-0.33
RK	3.4	4.8	3.6	0.76	0.165	3.93	-1.15	-1.13
RR								•
RZ	6.8	7.2	7.6	0.40	0.095	7.20	-0.02	0.00
S	8.0	9.4	10.7	1.35	0.319	9.37	0.73	0.75
SD	9.2	9.1	8.9	0.15	0.035	9.07	0.62	0.65
SF								•
SG	13.6	10.3	10.9	1.76	0.390	11.60	1.50	1.52
SI	6.8	6.3	6.7	0.26	0.059	6.60	-0.23	-0.21
SL	7.2	8.0	7.8	0.42	0.095	7.67	0.14	0.16
SM	6.0	5.7	6.8	0.57	0.130	6.17	-0.38	-0.36
SN	7.4	7.7	6.0	0.91	0.201	7.03	-0.08	-0.06
SO	6.4	6.3	5.8	0.32	0.071	6.17	-0.38	-0.36
SR								•
SS	6.5	7.6	7.3	0.57	0.130	7.13	-0.05	-0.02

• = No data submitted

TAG SYMBOLS

↑ = Above control limit

Ø = Insufficient data

× = Determined to be an outlier

↓ = Below control limit

Gross Alpha

Lab	Res. 1	Res. 2	Res. 3	Exper. Sigma	Rng anal (R + SR)	Average	Normalized deviation (grand-avg) (known)		Tag
SU	10.5	21.9	16.0	5.70	1.660	16.13	3.07	3.09	x
SV	11.7	10.4	10.3	0.78	0.165	10.80	1.22	1.25	
SX	8.0	8.7	10.8	1.46	0.331	9.17	0.66	0.68	
SZ	4.8	5.8	6.0	0.64	0.142	5.53	-0.60	-0.58	
T	6.2	5.2	5.1	0.61	0.130	5.50	-0.61	-0.59	
TD	11.5	7.6	10.6	2.04	0.461	9.90	0.91	0.94	
TL									•
TN	12.1	11.1	8.1	2.08	0.473	10.43	1.10	1.12	
TO									•
TQ	4.7	5.3	5.3	0.35	0.071	5.10	-0.75	-0.73	
TW	7.0	5.6	5.7	0.78	0.165	6.10	-0.40	-0.38	
TX									•
TY	5.4	5.8	5.8	0.23	0.047	5.67	-0.55	-0.53	
U	7.3	6.9	6.9	0.23	0.047	7.03	-0.08	-0.06	
UA									•
UM									•
UP	7.0	6.5	6.8	0.25	0.059	6.77	-0.17	-0.15	
UQ									•
UY									•
VA	7.0	6.3	7.1	0.44	0.095	6.80	-0.16	-0.14	
VC									•
VH									•
VI	8.1	7.6	7.4	0.36	0.083	7.70	0.15	0.17	
VJ	9.0	9.0	9.5	0.29	0.059	9.17	0.66	0.68	
VO	5.4	8.3	5.1	1.77	0.378	6.27	-0.35	-0.32	
VT									•
W	7.7	7.0	6.6	0.56	0.130	7.10	-0.06	-0.03	
WE	6.3	5.5	5.2	0.57	0.130	5.67	-0.55	-0.53	
WH	8.2	8.9	7.9	0.51	0.118	8.33	0.37	0.39	
WJ									•
WN	10.6	10.3	11.5	0.62	0.142	10.80	1.22	1.25	
WO	5.0	4.2	5.2	0.53	0.118	4.80	-0.85	-0.83	
WP									•
WR	6.2	7.3	11.4	2.74	0.614	8.30	0.36	0.38	
WS									•
WW	6.5	6.2	4.7	0.96	0.213	5.80	-0.51	-0.48	
WX									•
X	5.8	6.2	5.2	0.50	0.118	5.73	-0.53	-0.51	
XB									•
XC									•
XD									•
XF	8.0	7.3	7.8	0.36	0.083	7.70	0.15	0.17	
XI	7.7	6.3	6.2	0.84	0.177	6.73	-0.18	-0.16	
XJ									•
XK									•

• = No data submitted

TAG SYMBOLS

↑ = Above control limit

∅ = Insufficient data

x = Determined to be an outlier

↓ = Below control limit

Gross Alpha

Lab	Res. 1	Res. 2	Res. 3	Exper. Sigma	Rng anal (R + SR)	Average	Normalized deviation (grand-avg) (known) Tag	
XL	5.6	6.2	5.6	0.35	0.071	5.80	-0.51	-0.48
XM	6.8	6.8	6.8	0.00	0.000	6.80	-0.16	-0.14
XN	12.0	8.9	10.9	1.57	0.366	10.60	1.15	1.18
XO								•
XQ	5.6	4.0	8.8	2.44	0.567	6.13	-0.39	-0.37
XR								•

Data sorted by Laboratory Average

Average	Tag	Lab	Average	Tag	Lab	Average	Tag	Lab
0.20		PG	5.80		JM	7.13		JE
3.53		BO	5.80		DH	7.13		DB
3.67		CO	5.90		CE	7.17		C
3.93		RK	6.00		OS	7.17		AF
3.97		MX	6.10		TW	7.20		RZ
4.20		NB	6.10		HK	7.23		BH
4.43		EX	6.13		XQ	7.63		DT
4.43		EV	6.17		SO	7.67		SL
4.73		HL	6.17		SM	7.70		XF
4.80		WO	6.23		RI	7.70		VI
5.07		QM	6.23		JQ	7.80		KH
5.07		L	6.27		VO	7.80		GZ
5.10		TQ	6.27		ID	7.80		BK
5.10		FJ	6.33		NJ	7.83		JP
5.13		RD	6.33		EO	7.83		Q
5.13		LR	6.43		BB	7.87		AH
5.20		N	6.43		LF	7.90		HI
5.30		I	6.47		OA	7.93		NA
5.30		FF	6.47		EB	7.97		JS
5.33		NH	6.50		P	7.97		A
5.43		EL	6.57		NP	8.00		FN
5.43		CJ	6.60		SI	8.10		MF
5.43		HP	6.63		BA	8.10		BC
5.50		T	6.67		QQ	8.20		RB
5.53		SZ	6.73		XI	8.30		WR
5.53		JN	6.73		PM	8.30		QT
5.53		FL	6.77		UP	8.33		WH
5.67		WE	6.80		VA	8.37		BM
5.67		TY	6.80		AL	8.40		KE
5.73		X	6.80		XM	8.43		AR
5.73		QZ	6.83		GQ	8.63		AU
5.77		BN	6.90		ER	8.67		AZ
5.77		R	7.03		U	8.70		AJ
5.77		CA	7.03		SN	8.77		BS
5.80		XL	7.10		W	8.83		KX
5.80		WW	7.13		SS	8.87		FE

• = No data submitted

TAG SYMBOLS

↑ = Above control limit

Ø = Insufficient data

× = Determined to be an outlier

↓ = Below control limit


Gross Alpha

Data sorted by Laboratory Average

Average	Tag	Lab	Average	Tag	Lab	Average	Tag	Lab
8.93		QX	9.50		KT	10.60		CP
9.00		OF	9.73		D	10.80		WN
9.03		AE	9.77		RG	10.80		SV
9.07		SD	9.90		TD	11.13		O
9.13		JY	9.97		OB	11.53		AW
9.17		VJ	10.20		QW	11.60		SG
9.17		SX	10.30		DO	11.73		DZ
9.17		E	10.43		TN	11.93		HE
9.37		S	10.47		QU	16.13	x	SU
9.40		DE	10.53		QP	16.30	x	BL
9.47		AK	10.60		XN	45.60	x	CX


% responding labs

Frequency distribution



% responding labs

Frequency distribution



• ≡ No data submitted

∅ ≡ Insufficient data

TAG SYMBOLS

x ≡ Determined to be an outlier


↑ ≡ Above control limit

↓ ≡ Below control limit


Gross Beta**Statistical Summary**

186 Participants

The known value of this nuclide is **12.8 pCi/l** with an expected precision of **5.0**; the control limits are 4.1 to 21.5; the warning regions are 4.1 to 7.0 and 18.6 to 21.5



Statistic	Respondents	Non-outliers
Mean	13.43	Grand Avg 13.23
Std. Dev.	3.93	2.84
Variance	15.44	8.07
% Coef. of Var.	29.26	21.48
% deviation of mean from known value	4.89	3.33
Norm. dev. of mean from known value	0.16	0.15
Median	13.23	13.23
% deviation of median from known value	3.39	3.39
Norm. dev. of median from known value	0.11	0.15



11 / 16 ESD-LV Performance Evaluation: Gross Alpha-Beta in Water, 24-Jul-1998

Gross Beta

Lab	Res. 1	Res. 2	Res. 3	Exper. Sigma	Rng anal (R + SR)	Average	Normalized deviation (grand-avg) (known) Tag	
A	12.5	12.5	11.5	0.58	0.118	12.17	-0.37	-0.22
AE	12.1	10.5	11.3	0.80	0.189	11.30	-0.67	-0.52
AF	13.2	12.6	11.1	1.08	0.248	12.30	-0.32	-0.17
AH	16.6	16.6	15.9	0.40	0.083	16.37	1.09	1.24
AJ	9.2	10.3	9.0	0.70	0.154	9.50	-1.29	-1.14
AK	14.4	13.9	14.9	0.50	0.118	14.40	0.41	0.55
AL	12.4	12.6	12.0	0.31	0.071	12.33	-0.31	-0.16
AP								
AR	15.0	16.0	17.4	1.21	0.284	16.13	1.01	1.15
AU	13.0	12.8	12.5	0.25	0.059	12.77	-0.16	-0.01
AW	15.7	16.7	17.1	0.72	0.165	16.50	1.13	1.28
AZ	11.7	9.0	10.6	1.36	0.319	10.43	-0.97	-0.82
BA	12.1	13.0	12.1	0.52	0.106	12.40	-0.29	-0.14
BB	13.1	13.9	13.3	0.42	0.095	13.43	0.07	0.22
BC	8.2	9.5	8.1	0.78	0.165	8.60	-1.60	-1.45
BG								
BH	11.3	12.1	12.1	0.46	0.095	11.83	-0.48	-0.33
BK	13.0	13.8	15.5	1.28	0.295	14.10	0.30	0.45
BL	15.5	16.8	15.1	0.89	0.201	15.80	0.89	1.04
BM	12.3	14.5	12.8	1.15	0.260	13.20	-0.01	0.14
BN	5.5	7.1	7.5	1.06	0.236	6.70	-2.26	-2.11
BO	7.8	7.0	6.0	0.90	0.213	6.93	-2.18	-2.03
BS	14.6	14.6	12.2	1.39	0.284	13.80	0.20	0.35
C	13.0	12.6	12.5	0.26	0.059	12.70	-0.18	-0.03
CA	12.2	12.8	12.2	0.35	0.071	12.40	-0.29	-0.14
CC								
CE	12.2	11.7	16.3	2.52	0.543	13.40	0.06	0.21
CJ	17.0	14.0	13.0	2.08	0.473	14.67	0.50	0.65
CO	8.2	6.5	8.3	1.01	0.213	7.67	-1.93	-1.78
CP	40.6	33.1	45.9	6.43	1.975	39.87	9.23	9.38
CS								
CX	13.2	11.2	10.5	1.40	0.319	11.63	-0.55	-0.40
D	17.8	17.5	17.7	0.15	0.035	17.67	1.54	1.69
DB	15.7	16.2	16.6	0.45	0.106	16.17	1.02	1.17
DD								
DE	13.3	13.6	13.6	0.17	0.035	13.50	0.09	0.24
DH	12.2	14.6	15.7	1.79	0.413	14.17	0.33	0.47
DO	14.0	11.0	13.0	1.53	0.354	12.67	-0.19	-0.05
DR								
DT	12.3	12.5	13.1	0.42	0.095	12.63	-0.21	-0.06
DZ	17.3	16.9	20.0	1.69	0.366	18.07	1.68	1.82
E	14.1	12.1	13.1	1.00	0.236	13.10	-0.04	0.10
EB	12.4	11.3	12.0	0.56	0.130	11.90	-0.46	-0.31
EL	12.4	12.9	11.9	0.50	0.118	12.40	-0.29	-0.14
EO	18.0	13.0	12.0	3.21	0.709	14.33	0.38	0.53

• = No data submitted

TAG SYMBOLS

↑ = Above control limit

∅ = Insufficient data

× = Determined to be an outlier

↓ = Below control limit

Gross Beta

Lab	Res. 1	Res. 2	Res. 3	Exper. Sigma	Rng anal (R + SR)	Average	Normalized deviation (grand-avg) (known) Tag	
ER								•
EV	14.6	13.3	12.7	0.97	0.224	13.53	0.11	0.25
EW	19.3	19.0	19.3	0.17	0.035	19.20	2.07	2.22
EX	13.6	12.9	13.2	0.35	0.083	13.23	0.00	0.15
FE	11.9	11.9	11.2	0.40	0.083	11.67	-0.54	-0.39
FF	12.4	13.4	12.8	0.50	0.118	12.87	-0.12	0.02
FJ								•
FL	15.6	14.2	9.9	2.97	0.673	13.23	0.00	0.15
FN	7.0	8.0	7.0	0.58	0.118	7.33	-2.04	-1.89
GQ	13.6	14.6	11.7	1.47	0.343	13.30	0.03	0.17
GZ	11.8	12.4	12.7	0.46	0.106	12.30	-0.32	-0.17
HE	13.6	15.2	14.9	0.85	0.189	14.57	0.46	0.61
HI	15.7	13.0	13.2	1.50	0.319	13.97	0.26	0.40
HK	10.7	14.4	15.4	2.48	0.555	13.50	0.09	0.24
HL	13.6	12.6	11.9	0.85	0.201	12.70	-0.18	-0.03
HP	9.1	11.4	10.8	1.19	0.272	10.43	-0.97	-0.82
I	15.1	14.8	13.4	0.91	0.201	14.43	0.42	0.57
ID	14.1	12.3	10.0	2.06	0.484	12.13	-0.38	-0.23
J								•
JE	11.2	13.1	9.3	1.90	0.449	11.20	-0.70	-0.55
JG								•
JM	14.6	14.3	13.5	0.57	0.130	14.13	0.31	0.46
JN	9.3	9.6	9.8	0.25	0.059	9.57	-1.27	-1.12
JP	13.8	14.7	13.9	0.49	0.106	14.13	0.31	0.46
JQ	8.6	10.7	11.4	1.46	0.331	10.23	-1.04	-0.89
JS	15.2	13.0	13.8	1.11	0.260	14.00	0.27	0.42
JY	17.2	14.3	13.4	1.99	0.449	14.97	0.60	0.75
K								•
KE	15.0	14.4	14.1	0.46	0.106	14.50	0.44	0.59
KH	14.9	12.6	13.7	1.15	0.272	13.73	0.18	0.32
KT	6.5	7.0	8.0	0.76	0.177	7.17	-2.10	-1.95
KX	12.2	14.3	14.2	1.18	0.248	13.57	0.12	0.27
L	14.7	12.1	15.0	1.59	0.343	13.93	0.25	0.39
LE								•
LF	14.0	14.0	14.0	0.00	0.000	14.00	0.27	0.42
LL								•
LR	16.2	16.5	17.0	0.40	0.095	16.57	1.16	1.30
M								•
MF	13.7	15.0	14.5	0.66	0.154	14.40	0.41	0.55
MT								•
MV								•
MX	11.3	12.1	11.7	0.40	0.095	11.70	-0.53	-0.38
N	7.6	10.1	8.9	1.25	0.295	8.87	-1.51	-1.36
NA	10.7	12.0	11.0	0.68	0.154	11.23	-0.69	-0.54
NB	17.0	13.6	15.2	1.70	0.402	15.27	0.71	0.85

• = No data submitted

TAG SYMBOLS

↑ = Above control limit

∅ = Insufficient data

× = Determined to be an outlier

↓ = Below control limit

13 / 16 ESD-LV Performance Evaluation: Gross Alpha-Beta in Water, 24-Jul-1998

Gross Beta

Lab	Res. 1	Res. 2	Res. 3	Exper. Sigma	Rng anal (R + SR)	Average	Normalized deviation (grand-avg) (known) Tag	
NH	12.0	12.0	12.0	0.00	0.000	12.00	-0.42	-0.28
NJ	13.5	13.5	12.3	0.69	0.142	13.10	-0.04	0.10
NK								•
NO								•
NP	9.4	8.4	9.0	0.50	0.118	8.93	-1.49	-1.34
O	14.1	14.7	15.2	0.55	0.130	14.67	0.50	0.65
OA	29.0	26.0	28.0	1.53	0.354	27.67	5.00	5.15 ×
OB	13.0	10.8	15.4	2.30	0.543	13.07	-0.06	0.09
OF	10.6	11.2	11.5	0.46	0.106	11.10	-0.74	-0.59
OS	13.0	13.0	14.0	0.58	0.118	13.33	0.04	0.18
OT	16.3	17.2	15.2	1.00	0.236	16.23	1.04	1.19
OY								•
P	13.0	18.5	17.0	2.84	0.650	16.17	1.02	1.17
PB								•
PD	15.8	14.0	13.7	1.14	0.248	14.50	0.44	0.59
PG	0.7	0.6	0.7	0.06	0.012	0.67	-4.35	-4.20 ×
PM	15.2	14.9	15.4	0.25	0.059	15.17	0.67	0.82
Q	15.8	14.0	16.7	1.37	0.319	15.50	0.79	0.94
QM	7.0	6.8	7.0	0.12	0.024	6.93	-2.18	-2.03
QP	8.0	7.6	8.7	0.56	0.130	8.10	-1.78	-1.63
QQ	14.5	14.0	14.5	0.29	0.059	14.33	0.38	0.53
QT	17.4	30.9	13.2	9.25	3.078	20.50	2.52	2.67
QU	15.1	15.3	16.4	0.70	0.154	15.60	0.82	0.97
QW	16.8	17.3	17.5	0.36	0.083	17.20	1.38	1.52
QX	19.3	20.0	20.5	0.60	0.142	19.93	2.32	2.47
QZ	11.3	12.5	11.8	0.60	0.142	11.87	-0.47	-0.32
R	13.4	14.0	15.4	1.03	0.236	14.27	0.36	0.51
RB	12.0	11.5	12.0	0.29	0.059	11.83	-0.48	-0.33
RD	11.6	17.1	19.4	4.01	0.921	16.03	0.97	1.12
RG	17.9	18.9	18.0	0.55	0.118	18.27	1.75	1.89
RI	12.3	11.5	15.0	1.83	0.413	12.93	-0.10	0.05
RK	12.4	13.7	11.7	1.01	0.236	12.60	-0.22	-0.07
RR								•
RZ	13.0	12.9	14.3	0.78	0.165	13.40	0.06	0.21
S	17.6	17.5	17.3	0.15	0.035	17.47	1.47	1.62
SD	10.0	10.0	10.1	0.06	0.012	10.03	-1.11	-0.96
SF								•
SG	17.9	15.6	13.5	2.20	0.520	15.67	0.85	0.99
SI	15.1	11.2	13.8	1.99	0.461	13.37	0.05	0.20
SL	11.9	12.1	11.0	0.59	0.130	11.67	-0.54	-0.39
SM	10.1	9.8	11.0	0.62	0.142	10.30	-1.01	-0.87
SN	14.4	14.0	13.2	0.61	0.142	13.87	0.22	0.37
SO	12.1	12.0	11.1	0.55	0.118	11.73	-0.52	-0.37
SR								•
SS	12.2	13.1	12.6	0.45	0.106	12.63	-0.21	-0.06

• = No data submitted

TAG SYMBOLS

↑ = Above control limit

∅ = Insufficient data

× = Determined to be an outlier

↓ = Below control limit

Gross Beta

Lab	Res. 1	Res. 2	Res. 3	Exper. Sigma	Rng anal (R + SR)	Average	Normalized deviation (grand-avg) (known)		Tag
SU	26.1	19.9	21.0	3.31	0.732	22.33	3.15	3.30	↑
SV	14.2	13.7	14.1	0.26	0.059	14.00	0.27	0.42	
SX	16.2	20.6	14.5	3.15	0.721	17.10	1.34	1.49	
SZ	11.7	12.2	11.4	0.40	0.095	11.77	-0.51	-0.36	
T	13.1	11.5	12.2	0.80	0.189	12.27	-0.33	-0.18	
TD	15.0	17.7	16.1	1.36	0.319	16.27	1.05	1.20	
TL									•
TN	13.4	13.2	12.0	0.76	0.165	12.87	-0.12	0.02	
TO									•
TQ	12.4	12.9	12.5	0.26	0.059	12.60	-0.22	-0.07	
TW	12.3	9.9	10.0	1.36	0.284	10.73	-0.86	-0.72	
TX									•
TY	12.1	11.7	11.9	0.20	0.047	11.90	-0.46	-0.31	
U	12.3	11.8	12.1	0.25	0.059	12.07	-0.40	-0.25	
UA									•
UM									•
UP	14.5	11.2	12.4	1.67	0.390	12.70	-0.18	-0.03	
UQ									•
UY	10.3	11.3	11.1	0.53	0.118	10.90	-0.81	-0.66	
VA	9.2	8.7	7.2	1.04	0.236	8.37	-1.68	-1.54	
VC									•
VH									•
VI	15.6	14.2	14.1	0.84	0.177	14.63	0.49	0.64	
VJ	17.5	13.5	11.8	2.93	0.673	14.27	0.36	0.51	
VO	14.1	19.4	9.8	4.81	1.255	14.43	0.42	0.57	
VT									•
W	13.7	13.8	12.2	0.90	0.189	13.23	0.00	0.15	
WE	16.1	13.9	15.2	1.11	0.260	15.07	0.64	0.79	
WH	17.6	13.7	14.7	2.03	0.461	15.33	0.73	0.88	
WJ									•
WN	4.5	4.9	5.2	0.35	0.083	4.87	-2.90	-2.75	
WO	17.0	11.5	13.0	2.84	0.650	13.83	0.21	0.36	
WP									•
WR	11.2	11.7	18.1	3.85	0.815	13.67	0.15	0.30	
WS									•
WW	12.6	11.0	10.8	0.99	0.213	11.47	-0.61	-0.46	
WX									•
X	13.0	11.2	13.4	1.17	0.260	12.53	-0.24	-0.09	
XB									•
XC									•
XD									•
XF	14.6	13.8	13.3	0.66	0.154	13.90	0.23	0.38	
XI	12.0	11.0	12.0	0.58	0.118	11.67	-0.54	-0.39	
XJ									•
XK									•

• = No data submitted

TAG SYMBOLS

↑ = Above control limit

∅ = Insufficient data

× = Determined to be an outlier

↓ = Below control limit

15 / 16 ESD-LV Performance Evaluation: Gross Alpha-Beta in Water, 24-Jul-1998

Gross Beta

Lab	Res. 1	Res. 2	Res. 3	Exper. Sigma	Rng anal (R + SR)	Average	Normalized deviation (grand-avg) (known)		Tag
XL	13.7	16.4	13.7	1.56	0.319	14.60	0.48	0.62	
XM	16.4	11.6	8.9	3.80	0.886	12.30	-0.32	-0.17	
XN	12.9	12.1	10.2	1.39	0.319	11.73	-0.52	-0.37	
XO									•
XQ	2.8	10.2	53.9	27.62	10.594	22.30	3.14	3.29	↑
XR									•

Data sorted by Laboratory Average

Average	Tag	Lab	Average	Tag	Lab	Average	Tag	Lab
0.67	×	PG	11.83		BH	13.30		GQ
4.87		WN	11.87		QZ	13.33		OS
6.70		BN	11.90		TY	13.37		SI
6.93		QM	11.90		EB	13.40		RZ
6.93		BO	12.00		NH	13.40		CE
7.17		KT	12.07		U	13.43		BB
7.33		FN	12.13		ID	13.50		HK
7.67		CO	12.17		A	13.50		DE
8.10		QP	12.27		T	13.53		EV
8.37		VA	12.30		XM	13.57		KX
8.60		BC	12.30		GZ	13.67		WR
8.87		N	12.30		AF	13.73		KH
8.93		NP	12.33		AL	13.80		BS
9.50		AJ	12.40		EL	13.83		WO
9.57		JN	12.40		CA	13.87		SN
10.03		SD	12.40		BA	13.90		XF
10.23		JQ	12.53		X	13.93		L
10.30		SM	12.60		TQ	13.97		HI
10.43		HP	12.60		RK	14.00		SV
10.43		AZ	12.63		SS	14.00		LF
10.73		TW	12.63		DT	14.00		JS
10.90		UY	12.67		DO	14.10		BK
11.10		OF	12.70		UP	14.13		JP
11.20		JE	12.70		HL	14.13		JM
11.23		NA	12.70		C	14.17		DH
11.30		AE	12.77		AU	14.27		VJ
11.47		WW	12.87		TN	14.27		R
11.63		CX	12.87		FF	14.33		QQ
11.67		XI	12.93		RI	14.33		EO
11.67		SL	13.07		OB	14.40		MF
11.67		FE	13.10		NJ	14.40		AK
11.70		MX	13.10		E	14.43		VO
11.73		XN	13.20		BM	14.43		I
11.73		SO	13.23		W	14.50		PD
11.77		SZ	13.23		FL	14.50		KE
11.83		RB	13.23		EX	14.57		HE

• = No data submitted

TAG SYMBOLS

↑ = Above control limit

Ø = Insufficient data

× = Determined to be an outlier


↓ = Below control limit

Gross Beta**Data sorted by Laboratory Average**

Average	Tag	Lab	Average	Tag	Lab	Average	Tag	Lab
14.60		XL	15.67		SG	17.20		QW
14.63		VI	15.80		BL	17.47		S
14.67		O	16.03		RD	17.67		D
14.67		CJ	16.13		AR	18.07		DZ
14.97		JY	16.17		P	18.27		RG
15.07		WE	16.17		DB	19.20		EW
15.17		PM	16.23		OT	19.93		QX
15.27		NB	16.27		TD	20.50		QT
15.33		WH	16.37		AH	22.30	↑	XQ
15.50		Q	16.50		AW	22.33	↑	SU
15.60		QU	16.57		LR	27.67	×	OA
			17.10		SX	39.87	×	CP


% responding labs

Frequency distribution



% responding labs

Frequency distribution



• ≡ No data submitted

∅ ≡ Insufficient data

TAG SYMBOLS

× ≡ Determined to be an outlier

↑ ≡ Above control limit

↓ ≡ Below control limit